

LB 86612

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

3/29/2017 12:58

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.914	0.0	0.107	
ICB1	0.007	0.0	0.009	
CCV1	0.983	0.0	0.114	
CCB1	0.003	0.0	0.008	
PB97890BL	0.004	0.0	0.008	
PB97890BS	0.911	0.0	0.107	
I2422-01	0.014	0.0	0.009	
I2422-01DUP	0.017	0.0	0.010	
I2422-01MS	1.027	0.0	0.119	
I2422-01MSD	1.014	0.0	0.118	
CCV2	0.976	0.0	0.114	
CCB2	0.005	0.0	0.008	

N	12
Mean	0.490
SD	0.5037
CV%	102.87

LB-86612

Aquakem v. 7.2AQ1

Results from time period:

AK

Wed Mar 29 11:59:52 2017

Wed Mar 29 12:34:40 2017

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0538	mg/l	3/29/2017 11:59:52	
0.1PPM	A	Phenolics-	P	0.1015	mg/l	3/29/2017 11:59:53	
0.25PPM	A	Phenolics-	P	0.2556	mg/l	3/29/2017 11:59:54	
0.50PPM	A	Phenolics-	P	0.4974	mg/l	3/29/2017 11:59:55	
1.0PPM	A	Phenolics-	P	0.9838	mg/l	3/29/2017 11:59:56	
2.0PPM	A	Phenolics-	P	2.0079	mg/l	3/29/2017 11:59:57	
ICV1	S	Phenolics-	P	0.914	mg/l	3/29/2017 12:30:41	
ICB1	S	Phenolics-	P	0.0069	mg/l	3/29/2017 12:30:42	
CCV1	S	Phenolics-	P	0.9832	mg/l	3/29/2017 12:30:43	
CCB1	S	Phenolics-	P	0.0034	mg/l	3/29/2017 12:30:44	
PB97890BL	S	Phenolics-	P	0.0036	mg/l	3/29/2017 12:30:45	
PB97890BS	S	Phenolics-	P	0.9106	mg/l	3/29/2017 12:30:46	
I2422-01	S	Phenolics-	P	0.0141	mg/l	3/29/2017 12:30:47	
I2422-01DUP	S	Phenolics-	P	0.0175	mg/l	3/29/2017 12:30:48	
I2422-01MS	S	Phenolics-	P	1.0274	mg/l	3/29/2017 12:30:49	
I2422-01MSD	S	Phenolics-	P	1.014	mg/l	3/29/2017 12:30:50	
CCV2	S	Phenolics-	P	0.976	mg/l	3/29/2017 12:30:51	
CCB2	S	Phenolics-	P	0.0052	mg/l	3/29/2017 12:30:52	

C886612

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

3/29/2017 12:02

Test Phenolics-


Accepted 3/29/2017 12:02

Factor 9.221

Bias 0.008

Coeff. of det. 0.999863

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.014	0.0538	0.0500	
2	Phenol-2	0.019	0.1015	0.1000	
3	Phenol-2	0.035	0.2556	0.2500	
4	Phenol-2	0.062	0.4974	0.5000	
5	Phenol-2	0.114	0.9838	1.0000	
6	Phenol-2	0.226	2.0079	2.0000	