

Daily Analysis Runlog For Sequence/QC Batch ID # LB86623

Review By		Review On				
SubDirectory		LB86623		Test : Cyanide		
STD. NAME		STD REF.#				
ICAL Standard		WP53861,WP53862,WP53863,WP53864,WP53865,WP53866,WP53867				
ICV Standard		W1540				
CCV Standard		WP53869				
ICSA Standard						
CRI Standard						
LCS Standard		WP51099				
Chk Standard		WP53868,WP51626,WP51276				

Sr#	SampleId	ClientID	QcType	Date	Comment	Status
1	0.0PPBCN	0.0PPBCN	CAL1	03/29/17 14:30		
2	5.0PPBCN	5.0PPBCN	CAL2	03/29/17 14:30		
3	10PPBCN	10PPBCN	CAL3	03/29/17 14:30		
4	50PPBCN	50PPBCN	CAL4	03/29/17 14:30		
5	100PPBCN	100PPBCN	CAL5	03/29/17 14:30		
6	250PPBCN	250PPBCN	CAL6	03/29/17 14:30		
7	500PPBCN	500PPBCN	CAL7	03/29/17 14:30		
8	ICV1	ICV1	ICV	03/29/17 16:12		
9	ICB1	ICB1	ICB	03/29/17 16:12		
10	CCV1	CCV1	CCV	03/29/17 16:12		
11	CCB1	CCB1	CCB	03/29/17 16:12		
12	PB97838BL	PB97838BL	MB	03/29/17 16:12		
13	PB97838BS	PB97838BS	LCS	03/29/17 16:12		
14	LOWPB97838	LOWPB97838	SAM	03/29/17 16:12		
15	HIGHPB97838	HIGHPB97838	SAM	03/29/17 16:12		
16	I2399-02	MH-6	SAM	03/29/17 16:12		
17	I2399-02DUP	MH-6DUP	DUP	03/29/17 16:12		
18	I2399-02MS	MH-6MS	MS	03/29/17 16:12		
19	I2399-02MSD	MH-6MSD	MSD	03/29/17 16:20		
20	I2421-01	HHD-GW1	SAM	03/29/17 16:20		
21	I2421-03	HHD-GW2	SAM	03/29/17 16:20		

284 Sheffield Street, Mountainside, New Jersey 07092 Phone : 908 789 8900 Fax : 908 789 8922

Daily Analysis Runlog For Sequence/QC Batch ID # LB86623

Review By			Review On			
SubDirectory		LB86623		Test : Cyanide		
STD. NAME		STD REF.#				
ICAL Standard		WP53861,WP53862,WP53863,WP53864,WP53865,WP53866,WP53867				
ICV Standard		W1540				
CCV Standard		WP53869				
ICSA Standard						
CRI Standard						
LCS Standard		WP51099				
Chk Standard		WP53868,WP51626,WP51276				
22	CCV2	CCV2	CCV	03/29/17 16:20		
23	CCB2	CCB2	CCB	03/29/17 16:20		
24	PB97887BL	PB97887BL	MB	03/29/17 16:20		
25	PB97887BS	PB97887BS	LCS	03/29/17 16:20		
26	I2422-01	HHD-S1	SAM	03/29/17 16:20		
27	I2422-01DUP	HHD-S1DUP	DUP	03/29/17 16:20		
28	I2422-01MS	HHD-S1MS	MS	03/29/17 16:20		
29	PB97920BL	PB97920BL	MB	03/29/17 16:27		
30	PB97920BS	PB97920BS	LCS	03/29/17 16:27		
31	LOWPB97920	LOWPB97920	SAM	03/29/17 16:27		
32	HIGHPB97920	HIGHPB97920	SAM	03/29/17 16:27		
33	CCV3	CCV3	CCV	03/29/17 16:27		
34	CCB3	CCB3	CCB	03/29/17 16:27		
35	I2440-01	ROLL-OFF-15-50	SAM	03/29/17 16:27		
36	I2440-01DUP	ROLL-OFF-15-50DU	DUP	03/29/17 16:27		
37	I2440-01MS	ROLL-OFF-15-50MS	MS	03/29/17 16:27		
38	I2440-01MSD	ROLL-OFF-15-50MS	MSD	03/29/17 16:27		
39	I2440-04	ROLL-OFF-20-41	SAM	03/29/17 16:27		
40	I2422-01MSD	HHD-S1MSD	MSD	03/29/17 17:03		
41	CCV4	CCV4	CCV	03/29/17 17:03		
42	CCB4	CCB4	CCB	03/29/17 17:03		



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB86623	Date Received:	
Client Sample ID:	HIGHPB97838	SDG No.:	LB86623
Lab Sample ID:	HIGHPB97838	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.49		1	0.003		0.005	mg/L		03/29/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB86623	Date Received:	
Client Sample ID:	HIGHPB97920	SDG No.:	LB86623
Lab Sample ID:	HIGHPB97920	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.48		1	0.003		0.005	mg/L		03/29/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/23/2017 12:00:00 AM
Project:	LB86623	Date Received:	3/24/2017 12:00:00 AM
Client Sample ID:	MH-6	SDG No.:	LB86623
Lab Sample ID:	I2399-02	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.00076	U	1	0.003		0.005	mg/L	03/28/2017	03/29/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/24/2017 12:00:00 AM
Project:	LB86623	Date Received:	3/24/2017 12:00:00 AM
Client Sample ID:	HHD-GW1	SDG No.:	LB86623
Lab Sample ID:	I2421-01	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.0017	U	1	0.003		0.005	mg/L	03/28/2017	03/29/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/24/2017 12:00:00 AM
Project:	LB86623	Date Received:	3/24/2017 12:00:00 AM
Client Sample ID:	HHD-GW2	SDG No.:	LB86623
Lab Sample ID:	I2421-03	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.00066	U	1	0.003		0.005	mg/L	03/28/2017	03/29/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/23/2017 12:00:00 AM
Project:	LB86623	Date Received:	3/24/2017 12:00:00 AM
Client Sample ID:	HHD-S1	SDG No.:	LB86623
Lab Sample ID:	I2422-01	Matrix:	Solid
Level (low/med):	low	% Solid:	63.8

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.57		1	0.05		0.38	mg/Kg	03/28/2017	03/29/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/28/2017 12:00:00 AM
Project:	LB86623	Date Received:	3/28/2017 12:00:00 AM
Client Sample ID:	ROLL-OFF-15-50	SDG No.:	LB86623
Lab Sample ID:	I2440-01	Matrix:	Solid
Level (low/med):	low	% Solid:	81.3

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.36		1	0.039		0.3	mg/Kg	03/29/2017	03/29/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	3/28/2017 12:00:00 AM
Project:	LB86623	Date Received:	3/28/2017 12:00:00 AM
Client Sample ID:	ROLL-OFF-20-41	SDG No.:	LB86623
Lab Sample ID:	I2440-04	Matrix:	Solid
Level (low/med):	low	% Solid:	88.4

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.0001	U	1	0.036		0.27	mg/Kg	03/29/2017	03/29/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB86623	Date Received:	
Client Sample ID:	LOWPB97838	SDG No.:	LB86623
Lab Sample ID:	LOWPB97838	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.01		1	0.003		0.005	mg/L		03/29/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB86623	Date Received:	
Client Sample ID:	LOWPB97920	SDG No.:	LB86623
Lab Sample ID:	LOWPB97920	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.01		1	0.003		0.005	mg/L		03/29/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB86623	Date Received:	
Client Sample ID:	LOWPB97920	SDG No.:	LB86623
Lab Sample ID:	LOWPB97920	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
-----	-----------	-------	------	----	-----	-----	------------	-------	-----------	-----------	----------

Color Before:	Clarity Before:	Texture:
Color After:	Clarity After:	Artifacts:
Comments:		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits

Sample ID	Analyte	Result	True Value	% Recovery	% RSD	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
		mg/L								
ICV1	Cyanide	0.095	0.099	96	0	85 - 115		03/29/2017	16:12	LB86623



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB86623

SDG No.: LB86623

Contract: _____ Lab Code: CHEM

Case No.: LB86623

SAS No.: LB86623

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	% RSD	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV1	Cyanide	0.24	0.25	96	0	90 - 110		03/29/2017	16:12	LB86623
CCV2	Cyanide	0.24	0.25	96	0	90 - 110		03/29/2017	16:20	LB86623
CCV3	Cyanide	0.24	0.25	96	0	90 - 110		03/29/2017	16:27	LB86623
CCV4	Cyanide	0.24	0.25	96	0	90 - 110		03/29/2017	17:03	LB86623



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB86623 **SDG No.:** LB86623
Contract: **Lab Code:** CHEM **Case No.:** LB86623 **SAS No.:** LB86623

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	Cyanide	-0.00054	+/-0.005	U		0.005		03/29/2017	16:12	LB86623

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	Cyanide	-0.00014	+/-0.005	U		0.005		03/29/2017	16:12	LB86623
CCB2	Cyanide	-0.00087	+/-0.005	U		0.005		03/29/2017	16:20	LB86623
CCB3	Cyanide	-0.00015	+/-0.005	U		0.005		03/29/2017	16:27	LB86623
CCB4	Cyanide	0.0003	+/-0.005	U		0.005		03/29/2017	17:03	LB86623



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB86623

SDG No.: LB86623

Contract: Lab Code: CHEM

Case No.: LB86623

SAS No.: LB86623

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	-----	------	---	------------------	------------------	---------------



GENCHEM
- 3b -
PREPARATION BLANK SUMMARY

Client: LB86623

SDG No.: LB86623

Instrument: Konelab 20

Sample ID	Analyte	Result (mg/L)	Acceptance Limit	Conc Qual	LOD mg/L	CRQL mg/L	M	Analysis Date	Analysis Time	Run
PB97838BL	WATER		Batch Number:		PB97838		Prep Date:		03/28/2017	
	Cyanide	-0.00026	<0.005	U		0.005		03/29/2017	16:12	LB86623
Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	LOD mg/Kg	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB97887BL	SOLID		Batch Number:		PB97887		Prep Date:		03/28/2017	
	Cyanide	-0.027	<0.25	U		0.25		03/29/2017	16:20	LB86623
Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	LOD mg/Kg	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB97920BL	SOLID		Batch Number:		PB97920		Prep Date:		03/29/2017	
	Cyanide	-0.034	<0.25	U		0.25		03/29/2017	16:27	LB86623



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB86623 level: low sdg no.: LB86623
contract: lab code: CHEM case no.: LB86623 sas no.: LB86623
matrix: WATER sample id: I2399-02 client id: MH-6MS
Percent Solids for Sample: 0 Spiked ID: I2399-02MS Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/L	75 - 125	0.032		0.003	U	0.04	80		



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB86623 level: low sdg no.: LB86623
contract: lab code: CHEM case no.: LB86623 sas no.: LB86623
matrix: WATER sample id: I2399-02 client id: MH-6MSD
Percent Solids for Sample: 0 Spiked ID: I2399-02MSD Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/L	75 - 125	0.031		0.003	U	0.04	77.5		



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB86623 level: low sdg no.: LB86623
contract: lab code: CHEM case no.: LB86623 sas no.: LB86623
matrix: Solid sample id: I2422-01 client id: HHD-S1MS
Percent Solids for Sample: 63.8 Spiked ID: I2422-01MS Percent Solids for Spike Sample: 63.8

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	3.2		0.57		3.1	85		



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB86623 level: low sdg no.: LB86623
contract: lab code: CHEM case no.: LB86623 sas no.: LB86623
matrix: Solid sample id: I2422-01 client id: HHD-S1MSD
Percent Solids for Sample: 63.8 Spiked ID: I2422-01MSD Percent Solids for Spike Sample: 63.8

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	3.1		0.57		3	84		



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB86623 level: low sdg no.: LB86623
contract: lab code: CHEM case no.: LB86623 sas no.: LB86623
matrix: Solid sample id: I2440-01 client id: ROLL-OFF-15-50MS
Percent Solids for Sample: 81.3 Spiked ID: I2440-01MS Percent Solids for Spike Sample: 81.3

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	2.1		0.36		2.4	72	*	



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB86623 level: low sdg no.: LB86623
contract: lab code: CHEM case no.: LB86623 sas no.: LB86623
matrix: Solid sample id: I2440-01 client id: ROLL-OFF-15-50MSD
Percent Solids for Sample: 81.3 Spiked ID: I2440-01MSD Percent Solids for Spike Sample: 81.3

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125		2	0.36		2.3	71	*	

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB86623 **Level:** LOW **SDG No.:** LB86623
Contract: **Lab Code:** CHEM **Case No.:** LB86623 **SAS No.:** LB86623
Matrix: WATER **Sample ID:** I2399-02 **Client ID:** MH-6DUP
Percent Solids for Sample: 0 **Duplicate ID** I2399-02DUP **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/L	20	0.003	U		0	U		



GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB86623 **Level:** LOW **SDG No.:** LB86623
Contract: **Lab Code:** CHEM **Case No.:** LB86623 **SAS No.:** LB86623
Matrix: WATER **Sample ID:** I2399-02MS **Client ID:** MH-6MSD
Percent Solids for Sample: 0 **Duplicate ID** I2399-02MSD **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/L	20	0.032		0.031		3		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB86623 **Level:** LOW **SDG No.:** LB86623
Contract: **Lab Code:** CHEM **Case No.:** LB86623 **SAS No.:** LB86623
Matrix: Solid **Sample ID:** I2422-01 **Client ID:** HHD-S1DUP
Percent Solids for Sample: 63.8 **Duplicate ID** I2422-01DUP **Percent Solids for Spike Sample:** 63.8

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	0.57		0.58		2		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB86623 **Level:** LOW **SDG No.:** LB86623
Contract: **Lab Code:** CHEM **Case No.:** LB86623 **SAS No.:** LB86623
Matrix: Solid **Sample ID:** I2422-01MS **Client ID:** HHD-S1MSD
Percent Solids for Sample: 63.8 **Duplicate ID** I2422-01MSD **Percent Solids for Spike Sample:** 63.8

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	3.2		3.1		3		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB86623 **Level:** LOW **SDG No.:** LB86623
Contract: **Lab Code:** CHEM **Case No.:** LB86623 **SAS No.:** LB86623
Matrix: Solid **Sample ID:** I2440-01 **Client ID:** ROLL-OFF-15-50DUP
Percent Solids for Sample: 81.3 **Duplicate ID** I2440-01DUP **Percent Solids for Spike Sample:** 81.3

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	0.36		0.39		8		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB86623 **Level:** LOW **SDG No.:** LB86623
Contract: **Lab Code:** CHEM **Case No.:** LB86623 **SAS No.:** LB86623
Matrix: Solid **Sample ID:** I2440-01MS **Client ID:** ROLL-OFF-15-50MSD
Percent Solids for Sample: 81.3 **Duplicate ID** I2440-01MSD **Percent Solids for Spike Sample:** 81.3

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	2.1			2	5		



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB86623

SDG No.: LB86623

Contract:

Lab Code: CHEM

Case No.: LB86623

SAS No.: LB86623

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB97838BS Cyanide	mg/L	0.1	0.098		98	85 - 115	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB86623

SDG No.: LB86623

Contract:

Lab Code: CHEM

Case No.: LB86623

SAS No.: LB86623

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB97887BS Cyanide	mg/Kg	5	5		100	85 - 115	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB86623 SDG No.: LB86623
Contract: Lab Code: CHEM Case No.: LB86623 SAS No.: LB86623

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB97920BS Cyanide	mg/Kg	5	4.9		98	85 - 115	
