

LB 26692

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

4/4/2017 11:36

Test: Total CN

4/4/17

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	93.018	0.0	0.092	
ICB1	0.218	0.0	0.005	
CCV1	236.083	0.0	0.226	
CCB1	0.915	0.0	0.005	
PB97999BL	0.321	0.0	0.005	
I2488-02	-0.160	0.0	0.004	
I2488-02DUP	-0.226	0.0	0.004	
I2493-17	-0.005	0.0	0.005	
I2493-19	-0.612	0.0	0.004	
I2493-25	-0.119	0.0	0.005	
I2493-25	-0.159	0.0	0.004	
I2493-27	0.184	0.0	0.005	
I2495-11	-0.218	0.0	0.004	
I2495-12	0.432	0.0	0.005	
I2495-13	0.434	0.0	0.005	
CCV2	239.883	0.0	0.229	
CCB2	0.430	0.0	0.005	
I2495-14	0.160	0.0	0.005	
I2500-03	-0.104	0.0	0.005	
I2508-02	-0.381	0.0	0.004	
I2508-04	-0.036	0.0	0.005	
CCV3	242.007	0.0	0.231	
CCB3	0.208	0.0	0.005	

N	23
Mean	35.316
SD	83.0702
CV%	235.22

LB86692

Aquakem v. 7.2AQ1

Results from time period:

Tue Apr 04 11:24:19 2017

A1c

Tue Apr 04 11:34:42 2017

Sample Id	Sam/Ctr/cf	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	-0.3866	µg/l	4/4/2017 9:41:10	
5.0PPBCN	A	Reactive CN	P	4.6039	µg/l	4/4/2017 9:41:11	
10PPBCN	A	Reactive CN	P	9.3502	µg/l	4/4/2017 9:41:12	
50PPBCN	A	Reactive CN	P	48.4244	µg/l	4/4/2017 9:41:13	
100PPBCN	A	Reactive CN	P	101.8653	µg/l	4/4/2017 9:41:14	
250PPBCN	A	Reactive CN	P	252.6828	µg/l	4/4/2017 9:41:15	
500PPBCN	A	Reactive CN	P	498.46	µg/l	4/4/2017 9:41:16	
ICV1	S	Reactive CN	P	93.0184	µg/l	4/4/2017 11:24:19	
ICB1	S	Reactive CN	P	0.2184	µg/l	4/4/2017 11:24:20	
CCV1	S	Reactive CN	P	236.0826	µg/l	4/4/2017 11:24:21	
CCB1	S	Reactive CN	P	0.9153	µg/l	4/4/2017 11:24:22	
PB97999BL	S	Reactive CN	P	0.3208	µg/l	4/4/2017 11:24:23	
I2488-02	S	Reactive CN	P	-0.1598	µg/l	4/4/2017 11:24:24	
I2488-02DUP	S	Reactive CN	P	-0.2258	µg/l	4/4/2017 11:24:25	
I2493-17	S	Reactive CN	P	-0.0051	µg/l	4/4/2017 11:24:26	
I2493-19	S	Reactive CN	P	-0.612	µg/l	4/4/2017 11:24:27	
I2493-25	S	Reactive CN	P	-0.1188	µg/l	4/4/2017 11:24:28	
I2493-27	S	Reactive CN	P	0.1836	µg/l	4/4/2017 11:31:54	
I2495-11	S	Reactive CN	P	-0.2183	µg/l	4/4/2017 11:31:55	
I2495-12	S	Reactive CN	P	0.432	µg/l	4/4/2017 11:31:56	
I2495-13	S	Reactive CN	P	0.4337	µg/l	4/4/2017 11:31:57	
CCV2	S	Reactive CN	P	239.8828	µg/l	4/4/2017 11:31:58	
CCB2	S	Reactive CN	P	0.4296	µg/l	4/4/2017 11:31:59	
I2495-14	S	Reactive CN	P	0.1596	µg/l	4/4/2017 11:32:00	
I2500-03	S	Reactive CN	P	-0.1042	µg/l	4/4/2017 11:32:01	
I2508-02	S	Reactive CN	P	-0.3815	µg/l	4/4/2017 11:32:02	
I2508-04	S	Reactive CN	P	-0.0361	µg/l	4/4/2017 11:32:03	
CCV3	S	Reactive CN	P	242.0065	µg/l	4/4/2017 11:32:04	
CCB3	S	Reactive CN	P	0.2082	µg/l	4/4/2017 11:34:42	

C386692

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

4/4/2017 9:43


Test Total CN

Accepted 4/4/2017 9:43

Factor 1067
Bias 0.005

Coeff. of det. 0.999921

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.3866	0.0000	
2	5.0PPBCN	0.009	4.6039	5.0000	
3	10PPBCN	0.013	9.3502	10.0000	
4	50PPBCN	0.050	48.4244	50.0000	
5	100PPBCN	0.100	101.8653	100.0000	
6	250PPBCN	0.241	252.6828	250.0000	
7	500PPBCN	0.472	498.4600	500.0000	