


7470A

LB 86697
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35952



A= 0.0000e+000

B= 2.7593e-004

C= 2.5805e-002

Rho= 0.9998861

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.031	0.031	18	0.000	18				
0.2	0.200	0.199	-0.001	628	0.0 %	628				
2.5	2.500	2.411	-0.089	8644	0.0 %	8644				
5.0	5.000	5.042	0.042	18179	0.0 %	18179				
7.5	7.500	7.571	0.071	27345	0.0 %	27345				
10.0	10.000	9.946	-0.054	35952	0.0 %	35952				

LB86697 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 Apr 2017 16:28:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	18	-	0.00007470A	PPB		04 Apr 2017 16:30:25	S	Std
0.2 - 1	50.2	628	-	0.20007470A	PPB		04 Apr 2017 16:32:33	S	Std
2.5 - 1	52.5	8644	-	2.50007470A	PPB		04 Apr 2017 16:34:42	S	Std
5.0 - 1	55	18179	-	5.00007470A	PPB		04 Apr 2017 16:36:50	S	Std
7.5 - 1	57.5	27345	-	7.50007470A	PPB		04 Apr 2017 16:38:58	S	Std
10.0 - 1	510	35952	-	10.00007470A	PPB		04 Apr 2017 16:41:05	S	Std
ICV04 - 1	ICV04	13871	3.8532	-7470A	PPB		04 Apr 2017 16:43:39	U	SMPL
ICB04 - 1	ICB04	-109	-0.0043	-7470A	PPB		04 Apr 2017 16:45:45	U	SMPL
CCV44 - 1	CCV44	18248	5.0610	-7470A	PPB		04 Apr 2017 16:47:52	U	SMPL
CCB44 - 1	CCB44	-118	-0.0068	-7470A	PPB		04 Apr 2017 16:49:58	U	SMPL
CRA - 1	CRA	610	0.1941	-7470A	PPB		04 Apr 2017 16:52:05	U	SMPL
HighStd - 1	HighStd	35942	9.9433	-7470A	PPB		04 Apr 2017 16:54:11	U	SMPL
ChkStd - 1	ChkStd	25383	7.0298	-7470A	PPB		04 Apr 2017 16:56:18	U	SMPL
PB98032BL - 1	PBW	-157	-0.0175	-7470A	PPB		04 Apr 2017 16:58:27	U	SMPL
PB98032BS - 1	LCSW	13696	3.8050	-7470A	PPB		04 Apr 2017 17:00:35	U	SMPL
PB98032TB - 1	PB97973TB	-96	-0.0007	-7470A	PPB		04 Apr 2017 17:02:42	U	SMPL
I2488-02 - 1	WC-COMP-13	160	0.0700	-7470A	PPB		04 Apr 2017 17:04:50	U	SMPL
I2493-17 - 1	LSS-SP05-COMP	24	0.0324	-7470A	PPB		04 Apr 2017 17:06:58	U	SMPL
I2493-19 - 1	LSS-SP06-COMP	-24	0.0192	-7470A	PPB		04 Apr 2017 17:09:07	U	SMPL
I2493-21 - 1	LSS-SP07-COMP	12	0.0291	-7470A	PPB		04 Apr 2017 17:11:13	U	SMPL
CCV45 - 1	CCV45	17810	4.9401	-7470A	PPB		04 Apr 2017 17:13:20	U	SMPL
CCB45 - 1	CCB45	-121	-0.0076	-7470A	PPB		04 Apr 2017 17:15:27	U	SMPL
I2493-25 - 1	LSS-SP08-COMP	14	0.0297	-7470A	PPB		04 Apr 2017 17:17:33	U	SMPL
I2493-26 - 1	LSS-SP09-COMP	53	0.0404	-7470A	PPB		04 Apr 2017 17:19:40	U	SMPL
I2493-27 - 1	LSS-SP10-COMP	51	0.0399	-7470A	PPB		04 Apr 2017 17:21:47	U	SMPL
I2495-11 - 1	LSS-SP01-COMP	37	0.0360	-7470A	PPB		04 Apr 2017 17:23:54	U	SMPL
I2495-12 - 1	LSS-SP02-COMP	38	0.0363	-7470A	PPB		04 Apr 2017 17:26:01	U	SMPL
I2495-13 - 1	LSS-SP03-COMP	53	0.0404	-7470A	PPB		04 Apr 2017 17:28:09	U	SMPL
I2495-14 - 1	LSS-SP04-COMP	15	0.0299	-7470A	PPB		04 Apr 2017 17:30:17	U	SMPL
I2508-02 - 1	WC-B58-B59-001	17	0.0305	-7470A	PPB		04 Apr 2017 17:32:25	U	SMPL
I2508-04 - 1	WC-B53-B57-001	628	0.1991	-7470A	PPB		04 Apr 2017 17:39:18	U	SMPL
I2508-04DUP - 1	WC-B53-B57-001DUP	601	0.1916	-7470A	PPB		04 Apr 2017 17:41:24	U	SMPL
CCV46 - 1	CCV46	18059	5.0088	-7470A	PPB		04 Apr 2017 17:43:31	U	SMPL
CCB46 - 1	CCB46	-79	0.0040	-7470A	PPB		04 Apr 2017 17:45:38	U	SMPL
I2508-04MS - 1	WC-B53-B57-001MS	14485	4.0227	-7470A	PPB		04 Apr 2017 17:47:45	U	SMPL
I2508-04MSD - 1	WC-B53-B57-001MSD	14266	3.9622	-7470A	PPB		04 Apr 2017 17:49:52	U	SMPL
PB98033BL - 1	PBW	-70	0.0065	-7470A	PPB		04 Apr 2017 17:51:59	U	SMPL
PB98033BS - 1	LCSW	12613	3.5061	-7470A	PPB		04 Apr 2017 17:54:06	U	SMPL
I2493-29 - 1	LSS-EB	-58	0.0098	-7470A	PPB		04 Apr 2017 17:56:14	U	SMPL
I2519-01 - 1	WC040317-LIQUID	91	0.0509	-7470A	PPB		04 Apr 2017 17:58:21	U	SMPL
I2519-01DUP - 1	WC040317-LIQUIDDUP	64	0.0435	-7470A	PPB		04 Apr 2017 18:00:29	U	SMPL
I2519-01MS - 1	WC040317-LIQUIDMS	14777	4.1032	-7470A	PPB		04 Apr 2017 18:02:38	U	SMPL
I2519-01MSD - 1	WC040317-LIQUIDMSD	16008	4.4429	-7470A	PPB		04 Apr 2017 18:04:46	U	SMPL
CCV47 - 1	CCV47	18305	5.0767	-7470A	PPB		04 Apr 2017 18:06:53	U	SMPL
CCB47 - 1	CCB47	-96	-0.0007	-7470A	PPB		04 Apr 2017 18:09:00	U	SMPL
I2508-04LX5 - 1		94	0.0517	-7470A	PPB		04 Apr 2017 18:11:08	U	SMPL
I2508-04A - 1		15363	4.2649	-7470A	PPB		04 Apr 2017 18:13:15	U	SMPL
I2519-01LX5 - 1		-41	0.0145	-7470A	PPB		04 Apr 2017 18:15:22	U	SMPL
I2519-01A - 1		15889	4.4101	-7470A	PPB		04 Apr 2017 18:17:29	U	SMPL
CCV48 - 1	CCV48	18006	4.9942	-7470A	PPB		04 Apr 2017 18:19:36	U	SMPL
CCB48 - 1	CCB48	-76	0.0048	-7470A	PPB		04 Apr 2017 18:21:44	U	SMPL