

LB86793

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

4/10/2017 14:29

Test: Positive
Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.057	0.0	0.097	
ICB1	-0.642	0.0	0.005	
CCV1	248.843	0.0	0.241	
CCB1	-0.206	0.0	0.006	
PB98133BL	-0.634	0.0	0.005	
I2582-02	-0.556	0.0	0.005	
I2582-02DUP	-0.706	0.0	0.005	
I2582-05	-0.624	0.0	0.005	
CCV2	248.466	0.0	0.241	
CCB2	-0.323	0.0	0.005	

N	10
Mean	58.967
SD	104.4112
CV%	177.07

UB86793

AK

Aquakem v. 7.2AQ1

Results from time period:

Mon Apr 10 14:28:39 2017

Mon Apr 10 14:28:48 2017

Sample Id	Sam/Ctr/c/	Test short nam	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	-1.4343	µg/l	4/10/2017 10:35:11	
5.0PPBCN	A	Reactive CN	P	3.4615	µg/l	4/10/2017 10:35:12	
10PPBCN	A	Reactive CN	P	9.0311	µg/l	4/10/2017 10:35:13	
50PPBCN	A	Reactive CN	P	49.732	µg/l	4/10/2017 10:35:14	
100PPBCN	A	Reactive CN	P	100.5041	µg/l	4/10/2017 10:35:15	
250PPBCN	A	Reactive CN	P	257.4898	µg/l	4/10/2017 10:35:16	
500PPBCN	A	Reactive CN	P	496.2159	µg/l	4/10/2017 10:35:17	
ICV1	S	Reactive CN	P	96.0566	µg/l	4/10/2017 14:28:39	
ICB1	S	Reactive CN	P	-0.6415	µg/l	4/10/2017 14:28:40	
CCV1	S	Reactive CN	P	248.8426	µg/l	4/10/2017 14:28:41	
CCB1	S	Reactive CN	P	-0.2061	µg/l	4/10/2017 14:28:42	
PB98133BLS		Reactive CN	P	-0.6341	µg/l	4/10/2017 14:28:43	
I2582-02	S	Reactive CN	P	-0.5558	µg/l	4/10/2017 14:28:44	
I2582-02DIS		Reactive CN	P	-0.7056	µg/l	4/10/2017 14:28:45	
I2582-05	S	Reactive CN	P	-0.6237	µg/l	4/10/2017 14:28:46	
CCV2	S	Reactive CN	P	248.4658	µg/l	4/10/2017 14:28:47	
CCB2	S	Reactive CN	P	-0.3231	µg/l	4/10/2017 14:28:48	

LB-86793

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

4/10/2017 10:52


Test Total CN

Accepted 4/10/2017 10:52

Factor 1056
Bias 0.006

Coeff. of det. 0.999630

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-1.4343	0.0000	
2	5.0PPBCN	0.009	3.4615	5.0000	
3	10PPBCN	0.014	9.0311	10.0000	
4	50PPBCN	0.053	49.7320	50.0000	
5	100PPBCN	0.101	100.5041	100.0000	
6	250PPBCN	0.250	257.4898	250.0000	
7	500PPBCN	0.476	496.2159	500.0000	