

C387070

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

4/26/2017 12:07

Test: ^{Reactive} Total CNAK
4/26/17

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	92.612	0.0	0.089	
ICB1	-0.244	0.0	0.005	
CCV1	234.786	0.0	0.217	
CCB1	-0.335	0.0	0.005	
PB98454BL	-0.251	0.0	0.005	
PB98454BL	-0.255	0.0	0.005	
I2832-05	-0.152	0.0	0.005	
I2832-05DUP	-0.534	0.0	0.004	
I2832-11	-0.171	0.0	0.005	
CCV2	246.755	0.0	0.228	
CCB2	0.061	0.0	0.005	

AK 4/26/17

N	11
Mean	52.025
SD	97.3746
CV%	187.17

Aquakem v. 7.2AQ1

Results from time period:

Wed Apr 26 11:31:27 2017

Wed Apr 26 11:58:53 2017

AK

LB87070

Sample Id	Sam/Ctr/c	Test short na	Test type	Result	Result unit	Result date and tir	Stat
0.0PPBCN	A	Reactive CN	P	-0.7407	µg/l	4/26/2017 9:45	
5.0PPBCN	A	Reactive CN	P	4.2235	µg/l	4/26/2017 9:45	
10PPBCN	A	Reactive CN	P	9.6799	µg/l	4/26/2017 9:45	
50PPBCN	A	Reactive CN	P	49.1873	µg/l	4/26/2017 9:45	
100PPBCN	A	Reactive CN	P	100.8825	µg/l	4/26/2017 9:45	
250PPBCN	A	Reactive CN	P	253.6973	µg/l	4/26/2017 9:45	
500PPBCN	A	Reactive CN	P	498.0703	µg/l	4/26/2017 9:45	
ICV1	S	Reactive CN	P	92.6116	µg/l	4/26/2017 11:31	
ICB1	S	Reactive CN	P	-0.2438	µg/l	4/26/2017 11:31	
CCV1	S	Reactive CN	P	234.7862	µg/l	4/26/2017 11:58	
CCB1	S	Reactive CN	P	-0.3347	µg/l	4/26/2017 11:58	
PB98454BL	S	Reactive CN	P	-0.2515	µg/l	4/26/2017 11:58	
I2832-05	S	Reactive CN	P	-0.1518	µg/l	4/26/2017 11:58	
I2832-05DUP	S	Reactive CN	P	-0.5344	µg/l	4/26/2017 11:58	
I2832-11	S	Reactive CN	P	-0.1712	µg/l	4/26/2017 11:58	
CCV2	S	Reactive CN	P	246.7551	µg/l	4/26/2017 11:58	
CCB2	S	Reactive CN	P	0.0606	µg/l	4/26/2017 11:58	

=====

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

4/26/2017 12:09


Test Total CN

Accepted 4/26/2017 10:15

Factor 1107
Bias 0.005

Coeff. of det. 0.999902

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.7407	0.0000	
2	5.0PPBCN	0.009	4.2235	5.0000	
3	10PPBCN	0.014	9.6799	10.0000	
4	50PPBCN	0.049	49.1873	50.0000	
5	100PPBCN	0.096	100.8825	100.0000	
6	250PPBCN	0.234	253.6973	250.0000	
7	500PPBCN	0.455	498.0703	500.0000	