

UB87077

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

4/26/2017 16:07

Reactive

Test: ~~Total~~ CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.161	0.0	0.091	
ICB1	-0.105	0.0	0.005	
CCV1	244.336	0.0	0.226	
CCB1	0.397	0.0	0.005	
PB98485BL	0.369	0.0	0.005	
I2847-02	0.022	0.0	0.005	
I2847-02DUP	0.309	0.0	0.005	
CCV2	250.140	0.0	0.231	
CCB2	0.402	0.0	0.005	

N	9
Mean	65.670
SD	107.5369
CV%	163.75

Aquakem v. 7.2AQ1

Results from time period:

Wed Apr 26 16:03:10 2017

Wed Apr 26 16:03:18 2017

Alc LB 87077

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.7407	µg/l	4/26/2017 9:45:51	
5.OPPBCN	A	Reactive CN	P	4.2235	µg/l	4/26/2017 9:45:52	
10PPBCN	A	Reactive CN	P	9.6799	µg/l	4/26/2017 9:45:53	
50PPBCN	A	Reactive CN	P	49.1873	µg/l	4/26/2017 9:45:54	
100PPBCN	A	Reactive CN	P	100.8825	µg/l	4/26/2017 9:45:55	
250PPBCN	A	Reactive CN	P	253.6973	µg/l	4/26/2017 9:45:56	
500PPBCN	A	Reactive CN	P	498.0703	µg/l	4/26/2017 9:45:57	
ICV1	S	Reactive CN	P	95.1607	µg/l	4/26/2017 16:03:10	
ICB1	S	Reactive CN	P	-0.1047	µg/l	4/26/2017 16:03:11	
CCV1	S	Reactive CN	P	244.3364	µg/l	4/26/2017 16:03:12	
CCB1	S	Reactive CN	P	0.3966	µg/l	4/26/2017 16:03:13	
PB98485BLS		Reactive CN	P	0.3687	µg/l	4/26/2017 16:03:14	
I2847-02	S	Reactive CN	P	0.022	µg/l	4/26/2017 16:03:15	
I2847-02DIS		Reactive CN	P	0.3088	µg/l	4/26/2017 16:03:16	
CCV2	S	Reactive CN	P	250.1402	µg/l	4/26/2017 16:03:17	
CCB2	S	Reactive CN	P	0.4022	µg/l	4/26/2017 16:03:18	

LB87070 LB87077

=====
Calibration results Aquakem 7.2AQ1 Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

4/26/2017 12:09


Test Total CN

Accepted 4/26/2017 10:15

Factor 1107
Bias 0.005

Coeff. of det. 0.999902

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.7407	0.0000	
2	5.0PPBCN	0.009	4.2235	5.0000	
3	10PPBCN	0.014	9.6799	10.0000	
4	50PPBCN	0.049	49.1873	50.0000	
5	100PPBCN	0.096	100.8825	100.0000	
6	250PPBCN	0.234	253.6973	250.0000	
7	500PPBCN	0.455	498.0703	500.0000	