

LB87373

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

5/10/2017 15:50

Test: ~~Total~~ CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.596	0.0	0.106	
ICB1	-0.908	0.0	0.005	
CCV1	245.406	0.0	0.264	
CCB1	-1.145	0.0	0.004	
I3040-05	-1.740	0.0	0.004	
I3040-05DUP	-1.522	0.0	0.004	
I3046-09	-1.828	0.0	0.004	
I3046-10	-1.471	0.0	0.004	
I3046-11	-1.446	0.0	0.004	
I3048-04	-1.472	0.0	0.004	
I3049-04	-1.665	0.0	0.004	
I3059-03	-1.776	0.0	0.004	
I3062-16	-1.539	0.0	0.004	
I3062-17	-1.273	0.0	0.004	
CCV2	248.145	0.0	0.267	
CCB2	-0.741	0.0	0.005	
PB98843BL	-1.153	0.0	0.004	
I3062-18	-1.384	0.0	0.004	
I3062-19	-1.599	0.0	0.004	
I3062-20	-1.462	0.0	0.004	
I3062-21	-1.169	0.0	0.004	
I3062-22	-1.323	0.0	0.004	
CCV3	256.731	0.0	0.276	
CCB3	-0.640	0.0	0.005	

N	24
Mean	34.109
SD	85.7109
CV%	251.28

LB87873

AK

Aquakem v. 7.2AQ1

Results from time period:

Wed May 10 15:37:52 2017

Wed May 10 15:48:51 2017

Sample Id	Sam/Ctr/c/	Test short n	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-1.2671	µg/l	5/10/2017 13:42:28	
5.OPPBCN	A	Reactive CN	P	3.6615	µg/l	5/10/2017 13:42:29	
10PPBCN	A	Reactive CN	P	8.5219	µg/l	5/10/2017 13:42:30	
50PPBCN	A	Reactive CN	P	50.2121	µg/l	5/10/2017 13:42:31	
100PPBCN	A	Reactive CN	P	101.3832	µg/l	5/10/2017 13:42:32	
250PPBCN	A	Reactive CN	P	255.4866	µg/l	5/10/2017 13:42:33	
500PPBCN	A	Reactive CN	P	497.0018	µg/l	5/10/2017 13:42:34	
ICV1	S	Reactive CN	P	95.5962	µg/l	5/10/2017 15:37:52	
ICB1	S	Reactive CN	P	-0.9077	µg/l	5/10/2017 15:37:53	
CCV1	S	Reactive CN	P	245.406	µg/l	5/10/2017 15:37:54	
CCB1	S	Reactive CN	P	-1.1449	µg/l	5/10/2017 15:37:55	
I3040-05	S	Reactive CN	P	-1.7401	µg/l	5/10/2017 15:37:56	
I3040-05DUP	S	Reactive CN	P	-1.5222	µg/l	5/10/2017 15:37:57	
I3046-09	S	Reactive CN	P	-1.8276	µg/l	5/10/2017 15:37:58	
I3046-10	S	Reactive CN	P	-1.471	µg/l	5/10/2017 15:37:59	
I3046-11	S	Reactive CN	P	-1.4463	µg/l	5/10/2017 15:38:00	
I3048-04	S	Reactive CN	P	-1.4718	µg/l	5/10/2017 15:38:01	
I3049-04	S	Reactive CN	P	-1.6645	µg/l	5/10/2017 15:38:02	
I3059-03	S	Reactive CN	P	-1.7763	µg/l	5/10/2017 15:45:27	
I3062-16	S	Reactive CN	P	-1.5389	µg/l	5/10/2017 15:45:28	
I3062-17	S	Reactive CN	P	-1.2728	µg/l	5/10/2017 15:45:29	
CCV2	S	Reactive CN	P	248.1448	µg/l	5/10/2017 15:45:30	
CCB2	S	Reactive CN	P	-0.7406	µg/l	5/10/2017 15:45:31	
PB98843BL	S	Reactive CN	P	-1.1534	µg/l	5/10/2017 15:45:32	
I3062-18	S	Reactive CN	P	-1.3837	µg/l	5/10/2017 15:45:33	
I3062-19	S	Reactive CN	P	-1.5987	µg/l	5/10/2017 15:45:34	
I3062-20	S	Reactive CN	P	-1.4623	µg/l	5/10/2017 15:45:35	
I3062-21	S	Reactive CN	P	-1.1692	µg/l	5/10/2017 15:45:36	
I3062-22	S	Reactive CN	P	-1.3234	µg/l	5/10/2017 15:45:37	
CCV3	S	Reactive CN	P	256.7314	µg/l	5/10/2017 15:48:50	
CCB3	S	Reactive CN	P	-0.6398	µg/l	5/10/2017 15:48:51	

=====

Calibration results	Aquakem 7.2AQ1	Page: 1
---------------------	----------------	---------

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

5/10/2017 13:51

Test Total CN


Accepted 5/10/2017 13:51

Factor 947.6

Bias 0.005

Coeff. of det. 0.999773

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-1.2671	0.0000	
2	5.0PPBCN	0.009	3.6615	5.0000	
3	10PPBCN	0.014	8.5219	10.0000	
4	50PPBCN	0.058	50.2121	50.0000	
5	100PPBCN	0.112	101.3832	100.0000	
6	250PPBCN	0.275	255.4866	250.0000	
7	500PPBCN	0.530	497.0018	500.0000	