


245.1

LB 87440
INSTRUMENT ID: CVL

Linear

 μ Abs.:

32659



A= 0.0000e+000

B= 3.0430e-004

C= 2.9666e-002

Rho= 0.9999408

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.020	0.020	-32	0.000	-32				
0.2	0.200	0.206	0.006	578	0.0 %	578				
2.5	2.500	2.428	-0.072	7883	0.0 %	7883				
5.0	5.000	5.039	0.039	16461	0.0 %	16461				
7.5	7.500	7.540	0.040	24680	0.0 %	24680				
10.0	10.000	9.968	-0.032	32659	0.0 %	32659				

LB87440 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 15 May 2017 10:46:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-32	-	0.0000	245.1	PPB	15 May 2017 10:55:07	S	Std
0.2 - 1	50.2	578	-	0.2000	245.1	PPB	15 May 2017 10:57:15	S	Std
2.5 - 1	52.5	7883	-	2.5000	245.1	PPB	15 May 2017 10:59:24	S	Std
5.0 - 1	55	16461	-	5.0000	245.1	PPB	15 May 2017 11:01:32	S	Std
7.5 - 1	57.5	24680	-	7.5000	245.1	PPB	15 May 2017 11:03:40	S	Std
10.0 - 1	510	32659	-	10.0000	245.1	PPB	15 May 2017 11:05:47	S	Std
ICV54 - 1	ICV54	13077	4.0090	-245.1		PPB	15 May 2017 11:08:19	U	SMPL
ICB54 - 1	ICB54	-94	0.0011	-245.1		PPB	15 May 2017 11:10:25	U	SMPL
CCV96 - 1	CCV96	16300	4.9897	-245.1		PPB	15 May 2017 11:12:32	U	SMPL
CCB96 - 1	CCB96	-94	0.0011	-245.1		PPB	15 May 2017 11:14:38	U	SMPL
CRA - 1	CRA	566	0.2019	-245.1		PPB	15 May 2017 11:16:45	U	SMPL
HighStd - 1	HighStd	32818	10.0161	-245.1		PPB	15 May 2017 11:18:52	U	SMPL
ChkStd - 1	ChkStd	23287	7.1158	-245.1		PPB	15 May 2017 11:20:59	U	SMPL
PB98932BL - 1	PBW	-114	-0.0050	-245.1		PPB	15 May 2017 11:23:06	U	SMPL
PB98932BS - 1	LCSW	13556	4.1547	-245.1		PPB	15 May 2017 11:25:14	U	SMPL
I3045-01 - 1	EFFLUENT-DAY1METALS	227	0.0987	-245.1		PPB	15 May 2017 11:27:21	U	SMPL
I3062-15 - 1	TWP-02	8	0.0321	-245.1		PPB	15 May 2017 11:29:29	U	SMPL
I3062-15DUP - 1	TWP-02DUP	-17	0.0245	-245.1		PPB	15 May 2017 11:31:38	U	SMPL
I3062-15MS - 1	TWP-02MS	15335	4.6961	-245.1		PPB	15 May 2017 11:33:46	U	SMPL
I3062-15MSD - 1	TWP-02MSD	15382	4.7104	-245.1		PPB	15 May 2017 11:35:53	U	SMPL
CCV97 - 1	CCV97	16082	4.9234	-245.1		PPB	15 May 2017 11:38:00	U	SMPL
CCB97 - 1	CCB97	-84	0.0041	-245.1		PPB	15 May 2017 11:40:07	U	SMPL
I3062-15LX5 - 1		-40	0.0175	-245.1		PPB	15 May 2017 11:42:14	U	SMPL
I3062-15A - 1		13007	3.9877	-245.1		PPB	15 May 2017 11:44:21	U	SMPL
CCV98 - 1	CCV98	16634	5.0913	-245.1		PPB	15 May 2017 11:46:28	U	SMPL
CCB98 - 1	CCB98	-88	0.0029	-245.1		PPB	15 May 2017 11:48:35	U	SMPL