

LB87679

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

5/24/2017 10:25

Test: Phenolics-

mgl

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.986	0.0	0.108	
ICB1	0.006	0.0	0.003	
CCV1	1.005	0.0	0.110	
CCB1	0.003	0.0	0.002	
PB99229BL	0.002	0.0	0.002	
I3224-01	0.009	0.0	0.003	
I3245-01	0.003	0.0	0.002	
I3245-01DUP	0.003	0.0	0.002	
I3245-02	0.001	0.0	0.002	
I3245-03	0.003	0.0	0.002	
I3245-04	0.902	0.0	0.099	
I3245-05	0.897	0.0	0.099	
I3245-06	0.001	0.0	0.002	
I3282-01	0.005	0.0	0.002	
CCV2	0.998	0.0	0.110	
CCB2	0.001	0.0	0.002	
I3282-02	0.004	0.0	0.002	
I3282-03	-0.001	0.0	0.002	
I3282-04	0.001	0.0	0.002	
I3282-05	0.008	0.0	0.003	
I3282-06	0.002	0.0	0.002	
I3282-07	0.001	0.0	0.002	
PB99229BS	0.992	0.0	0.109	
CCV3	1.001	0.0	0.110	
CCB3	0.002	0.0	0.002	

N 25
Mean 0.273
SD 0.4433
CV% 162.15

CB87679

Aquakem v. 7.2AQ1

Results from time period:

AK

Wed May 24 09:28:16 2017

Wed May 24 10:18:15 2017

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0526	mg/l	5/24/2017 9:28:16	
0.1PPM	A	Phenolics-	P	0.1034	mg/l	5/24/2017 9:28:17	
0.25PPM	A	Phenolics-	P	0.2533	mg/l	5/24/2017 9:28:18	
0.50PPM	A	Phenolics-	P	0.4981	mg/l	5/24/2017 9:28:19	
1.0PPM	A	Phenolics-	P	0.9855	mg/l	5/24/2017 9:28:20	
2.0PPM	A	Phenolics-	P	2.0071	mg/l	5/24/2017 9:28:21	
ICV1	S	Phenolics-	P	0.9863	mg/l	5/24/2017 10:04:13	
ICB1	S	Phenolics-	P	0.0062	mg/l	5/24/2017 10:04:14	
CCV1	S	Phenolics-	P	1.0049	mg/l	5/24/2017 10:04:15	
CCB1	S	Phenolics-	P	0.0026	mg/l	5/24/2017 10:04:16	
PB99229BL	S	Phenolics-	P	0.0017	mg/l	5/24/2017 10:04:17	
I3224-01	S	Phenolics-	P	0.0088	mg/l	5/24/2017 10:04:18	
I3245-01	S	Phenolics-	P	0.0028	mg/l	5/24/2017 10:04:19	
I3245-01DUP	S	Phenolics-	P	0.0031	mg/l	5/24/2017 10:04:20	
I3245-02	S	Phenolics-	P	0.0009	mg/l	5/24/2017 10:04:21	
I3245-03	S	Phenolics-	P	0.0028	mg/l	5/24/2017 10:04:22	
I3245-04	S	Phenolics-	P	0.9022	mg/l	5/24/2017 10:04:23	
I3245-05	S	Phenolics-	P	0.8974	mg/l	5/24/2017 10:04:24	
I3245-06	S	Phenolics-	P	0.0005	mg/l	5/24/2017 10:14:15	
I3282-01	S	Phenolics-	P	0.005	mg/l	5/24/2017 10:14:16	
CCV2	S	Phenolics-	P	0.9983	mg/l	5/24/2017 10:14:17	
CCB2	S	Phenolics-	P	0.0014	mg/l	5/24/2017 10:14:18	
I3282-02	S	Phenolics-	P	0.0041	mg/l	5/24/2017 10:14:19	
I3282-03	S	Phenolics-	P	-0.0012	mg/l	5/24/2017 10:14:20	
I3282-04	S	Phenolics-	P	0.0011	mg/l	5/24/2017 10:14:21	
I3282-05	S	Phenolics-	P	0.0078	mg/l	5/24/2017 10:14:22	
I3282-06	S	Phenolics-	P	0.0022	mg/l	5/24/2017 10:14:23	
I3282-07	S	Phenolics-	P	0.0011	mg/l	5/24/2017 10:14:24	
PB99229BS	S	Phenolics-	P	0.9922	mg/l	5/24/2017 10:14:25	
CCV3	S	Phenolics-	P	1.0014	mg/l	5/24/2017 10:14:26	
CCB3	S	Phenolics-	P	0.0018	mg/l	5/24/2017 10:18:14	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AC

5/24/2017 9:32


Test Phenolics-

Accepted 5/24/2017 9:32

Factor 9.259
Bias 0.002

Coeff. of det. 0.999895

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0526	0.0500	
2	Phenol-2	0.013	0.1034	0.1000	
3	Phenol-2	0.029	0.2533	0.2500	
4	Phenol-2	0.056	0.4981	0.5000	
5	Phenol-2	0.108	0.9855	1.0000	
6	Phenol-2	0.219	2.0071	2.0000	