


ISM02.4

LB 87816
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42763



A= 0.0000e+000

B= 2.3369e-004 slope

C= 3.3462e-002 y-intercept

Rho= 0.9998997

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	q/d
0.00	0.000	0.032	0.032	-7	0.000	-7					-
0.2	0.200	0.226	0.026	825	0.0 %	825					13
2.5	2.500	2.486	-0.014	10494	0.0 %	10494					-1
5.0	5.000	4.890	-0.110	20780	0.0 %	20780					-2
7.5	7.500	7.540	0.040	32120	0.0 %	32120					1
10.0	10.000	10.027	0.027	42763	0.0 %	42763					0

NB
6/21/17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-7	-		0 ISM02.4	PPB	6/2/2017 11:59	Std
	0.2 S0.2	825	-		0.2 ISM02.4	PPB	6/2/2017 12:01	Std
	2.5 S2.5	10494	-		2.5 ISM02.4	PPB	6/2/2017 12:04	Std
	5 S5	20780	-		5 ISM02.4	PPB	6/2/2017 12:06	Std
	7.5 S7.5	32120	-		7.5 ISM02.4	PPB	6/2/2017 12:08	Std
	10 S10	42763	-		10 ISM02.4	PPB	6/2/2017 12:10	Std
ICV	ICV	16308	3.8445 -		ISM02.4	PPB	6/2/2017 12:13	SMPL
ICB	ICB	-77	0.0155 -		ISM02.4	PPB	6/2/2017 12:15	SMPL
CCV056	CCV056	20759	4.8847 -		ISM02.4	PPB	6/2/2017 12:18	SMPL
CCB056	CCB056	-67	0.0178 -		ISM02.4	PPB	6/2/2017 12:20	SMPL
PB99415BL	PBW003	49	0.0449 -		ISM02.4	PPB	6/2/2017 12:22	SMPL
I3374-01	MCOLE7	-50	0.0218 -		ISM02.4	PPB	6/2/2017 12:24	SMPL
I3374-02	MCOLE8	-69	0.0173 -		ISM02.4	PPB	6/2/2017 12:27	SMPL
I3374-03	MCOLF1	-36	0.025 -		ISM02.4	PPB	6/2/2017 12:29	SMPL
I3374-04	MCOLF2	-50	0.0218 -		ISM02.4	PPB	6/2/2017 12:31	SMPL
I3374-05	MCOLF2D	-53	0.0211 -		ISM02.4	PPB	6/2/2017 12:34	SMPL
I3374-06	MCOLF2S	4236	1.0234 -		ISM02.4	PPB	6/2/2017 12:36	SMPL
I3374-07	MCOLF7	-82	0.0143 -		ISM02.4	PPB	6/2/2017 12:38	SMPL
I3374-08	MCOLF8	-61	0.0192 -		ISM02.4	PPB	6/2/2017 12:40	SMPL
I3374-09	MCOLG2	-65	0.0183 -		ISM02.4	PPB	6/2/2017 12:43	SMPL
I3374-10	MCOLG3	-56	0.0204 -		ISM02.4	PPB	6/2/2017 12:45	SMPL
I3374-11	MCOLH0	-57	0.0201 -		ISM02.4	PPB	6/2/2017 12:47	SMPL
PB99416BL	PBW003	45	0.044 -		ISM02.4	PPB	6/2/2017 12:49	SMPL
I3375-01	MCOL93	-66	0.018 -		ISM02.4	PPB	6/2/2017 12:52	SMPL
I3375-02	MCOL97	-61	0.0192 -		ISM02.4	PPB	6/2/2017 12:54	SMPL
I3375-03	MCOLA3	-66	0.018 -		ISM02.4	PPB	6/2/2017 12:56	SMPL
I3375-04	MCOLA5	-52	0.0213 -		ISM02.4	PPB	6/2/2017 12:59	SMPL
I3375-05	MCOLA7	-66	0.018 -		ISM02.4	PPB	6/2/2017 13:01	SMPL
I3375-06	MCOLB1	-56	0.0204 -		ISM02.4	PPB	6/2/2017 13:03	SMPL
I3375-07	MCOLC9	-66	0.018 -		ISM02.4	PPB	6/2/2017 13:05	SMPL
CCV057	CCV057	20478	4.819 -		ISM02.4	PPB	6/2/2017 13:08	SMPL
CCB057	CCB057	-85	0.0136 -		ISM02.4	PPB	6/2/2017 13:10	SMPL
I3375-08	MCOLE5	-63	0.0187 -		ISM02.4	PPB	6/2/2017 13:12	SMPL
I3375-09	MCOLG7	-71	0.0169 -		ISM02.4	PPB	6/2/2017 13:15	SMPL
I3375-10	MCOLG5	-61	0.0192 -		ISM02.4	PPB	6/2/2017 13:17	SMPL
I3375-11	MCOLG9	-65	0.0183 -		ISM02.4	PPB	6/2/2017 13:19	SMPL
I3375-14	MCOLG5D	-71	0.0169 -		ISM02.4	PPB	6/2/2017 13:21	SMPL
I3375-15	MCOLG5S	4186	1.0117 -		ISM02.4	PPB	6/2/2017 13:24	SMPL
CCV058	CCV058	20678	4.8658 -		ISM02.4	PPB	6/2/2017 13:26	SMPL
CCB058	CCB058	-71	0.0169 -		ISM02.4	PPB	6/2/2017 13:28	SMPL