


ISM02.4

LB 87916
INSTRUMENT ID: CV4

Linear

 μ Abs.:

4101E



A= 0.0000e+000

B= 2.4168e-004

C= 2.3586e-002

Rho= 0.9994128

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.019	0.019	-19	0.000	-19					1
0.2	0.200	0.216	0.016	797	0.0 %	797					8
2.5	2.500	2.545	0.045	10432	0.0 %	10432					2
5.0	5.000	4.781	-0.219	19684	0.0 %	19684					-4
7.5	7.500	7.703	0.203	31775	0.0 %	31775					3
10.0	10.000	9.936	-0.064	41016	0.0 %	41016					-1

m3
617117

LB87916
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-19	-		0 ISM02.4	PPB	6/7/2017 13:55	Std
	0.2 S0.2	797	-		0.2 ISM02.4	PPB	6/7/2017 13:57	Std
	2.5 S2.5	10432	-		2.5 ISM02.4	PPB	6/7/2017 13:59	Std
	5 S5	19684	-		5 ISM02.4	PPB	6/7/2017 14:02	Std
	7.5 S7.5	31775	-		7.5 ISM02.4	PPB	6/7/2017 14:04	Std
	10 S10	41016	-		10 ISM02.4	PPB	6/7/2017 14:09	Std
ICV	ICV	16361	3.9777	-	ISM02.4	PPB	6/7/2017 14:12	SMPL
ICB	ICB	-74	0.0057	-	ISM02.4	PPB	6/7/2017 14:14	SMPL
CCV091	CCV091	21006	5.1003	-	ISM02.4	PPB	6/7/2017 14:16	SMPL
CCB091	CCB091	-97	0.0001	-	ISM02.4	PPB	6/7/2017 14:18	SMPL
PB99540BL	PBW003	-42	0.0134	-	ISM02.4	PPB	6/7/2017 14:21	SMPL
I3403-01	MJHTQ0	-51	0.0113	-	ISM02.4	PPB	6/7/2017 14:23	SMPL
I3403-02	MJHTQ1	-43	0.0132	-	ISM02.4	PPB	6/7/2017 14:25	SMPL
I3403-03	MJHTQ2	-81	0.004	-	ISM02.4	PPB	6/7/2017 14:28	SMPL
I3403-04	MJHTQ2D	-69	0.0069	-	ISM02.4	PPB	6/7/2017 14:30	SMPL
I3403-05	MJHTQ2S	4148	1.0261	-	ISM02.4	PPB	6/7/2017 14:32	SMPL
I3403-06	MJHTQ3	-49	0.0117	-	ISM02.4	PPB	6/7/2017 14:34	SMPL
I3403-07	MJHTQ4	-35	0.0151	-	ISM02.4	PPB	6/7/2017 14:37	SMPL
I3403-08	MJHTQ5	-109	-0.0028	-	ISM02.4	PPB	6/7/2017 14:39	SMPL
I3403-09	MJHTQ6	-57	0.0098	-	ISM02.4	PPB	6/7/2017 14:41	SMPL
I3403-10	MJHTQ7	-65	0.0079	-	ISM02.4	PPB	6/7/2017 14:44	SMPL
I3403-11	MJHTQ8	-79	0.0045	-	ISM02.4	PPB	6/7/2017 14:46	SMPL
I3403-12	MJHTQ9	-57	0.0098	-	ISM02.4	PPB	6/7/2017 14:48	SMPL
PB99541BL	PBW003	-43	0.0132	-	ISM02.4	PPB	6/7/2017 14:50	SMPL
I3404-01	MJHTR0	-68	0.0072	-	ISM02.4	PPB	6/7/2017 14:53	SMPL
I3404-02	MJHTR1	-66	0.0076	-	ISM02.4	PPB	6/7/2017 14:55	SMPL
I3404-03	MJHTR2	-76	0.0052	-	ISM02.4	PPB	6/7/2017 14:57	SMPL
I3404-04	MJHTR2D	-76	0.0052	-	ISM02.4	PPB	6/7/2017 14:59	SMPL
I3404-05	MJHTR2S	3926	0.9724	-	ISM02.4	PPB	6/7/2017 15:02	SMPL
I3404-06	MJHTR3	-58	0.0096	-	ISM02.4	PPB	6/7/2017 15:04	SMPL
CCV092	CCV092	20323	4.9352	-	ISM02.4	PPB	6/7/2017 15:06	SMPL
CCB092	CCB092	-80	0.0043	-	ISM02.4	PPB	6/7/2017 15:09	SMPL
I3404-07	MJHTR4	-62	0.0086	-	ISM02.4	PPB	6/7/2017 15:11	SMPL
I3404-08	MJHTR5	-94	0.0009	-	ISM02.4	PPB	6/7/2017 15:13	SMPL
I3404-09	MJHTR6	-93	0.0011	-	ISM02.4	PPB	6/7/2017 15:15	SMPL
I3404-10	MJHTR7	-81	0.004	-	ISM02.4	PPB	6/7/2017 15:18	SMPL
I3404-11	MJHTR8	-94	0.0009	-	ISM02.4	PPB	6/7/2017 15:20	SMPL
I3404-12	MJHTR9	-84	0.0033	-	ISM02.4	PPB	6/7/2017 15:22	SMPL
I3404-13	MJHTS0	-31	0.0161	-	ISM02.4	PPB	6/7/2017 15:24	SMPL
I3404-14	MJHTS1	-78	0.0047	-	ISM02.4	PPB	6/7/2017 15:27	SMPL
I3404-15	MJHTS2	-71	0.0064	-	ISM02.4	PPB	6/7/2017 15:29	SMPL
I3404-16	MJHTS3	-78	0.0047	-	ISM02.4	PPB	6/7/2017 15:31	SMPL
I3404-17	MJHTS4	-76	0.0052	-	ISM02.4	PPB	6/7/2017 15:34	SMPL
I3404-18	MJHTS5	-48	0.012	-	ISM02.4	PPB	6/7/2017 15:36	SMPL
I3404-19	MJHTS6	-40	0.0139	-	ISM02.4	PPB	6/7/2017 15:38	SMPL
I3404-20	MJHTS7	-41	0.0137	-	ISM02.4	PPB	6/7/2017 15:40	SMPL

LB87916

INSTRUMENT ID : CV1

CCV093	CCV093	20987	5.0957 -	ISM02.4	PPB	6/7/2017 15:43	SMPL
CCB093	CCB093	-58	0.0096 -	ISM02.4	PPB	6/7/2017 15:45	SMPL