


ISM02.4

LIB87980
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44353



A= 0.0000e+000

B= 2.2242e-004

C= 1.2994e-002

Rho= 0.9995946

Accept=Accepted

Slope
 $y = 1.2994e-002 x$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/D
0.00	0.000	0.018	0.018	24	0.000	24					—
0.2	0.200	0.221	0.021	937	0.0 %	937					11
2.5	2.500	2.352	-0.148	10514	0.0 %	10514					-6
5.0	5.000	5.084	0.084	22799	0.0 %	22799					2
7.5	7.500	7.647	0.147	34320	0.0 %	34320					2
10.0	10.000	9.878	-0.122	44353	0.0 %	44353					-1

NB
6/9/12

LB87980
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	24	-		0 ISM02.4	PPB	6/9/2017 14:08	Std
	0.2 S0.2	937	-		0.2 ISM02.4	PPB	6/9/2017 14:10	Std
	2.5 S2.5	10514	-		2.5 ISM02.4	PPB	6/9/2017 14:12	Std
	5 S5	22799	-		5 ISM02.4	PPB	6/9/2017 14:15	Std
	7.5 S7.5	34320	-		7.5 ISM02.4	PPB	6/9/2017 14:17	Std
	10 S10	44353	-		10 ISM02.4	PPB	6/9/2017 14:22	Std
ICV	ICV	17693	3.9483 -		ISM02.4	PPB	6/9/2017 14:25	SMPL
ICB	ICB	-49	0.0021 -		ISM02.4	PPB	6/9/2017 14:27	SMPL
CCV016	CCV016	22790	5.082 -		ISM02.4	PPB	6/9/2017 14:29	SMPL
CCB016	CCB016	-41	0.0039 -		ISM02.4	PPB	6/9/2017 14:31	SMPL
PB99608BL	PBS004	7	0.0146 -		ISM02.4	PPB	6/9/2017 14:34	SMPL
I3447-01	MBDC46	2493	0.5675 -		ISM02.4	PPB	6/9/2017 14:36	SMPL
I3447-02	MBDC46D	2182	0.4983 -		ISM02.4	PPB	6/9/2017 14:41	SMPL
I3447-03	MBDC46S	13695	3.0591 -		ISM02.4	PPB	6/9/2017 14:43	SMPL
I3447-04	MBDC47	1777	0.4082 -		ISM02.4	PPB	6/9/2017 14:45	SMPL
I3447-05	MBDC48	14604	3.2613 -		ISM02.4	PPB	6/9/2017 14:48	SMPL
I3447-06	MBDC49	1012	0.2381 -		ISM02.4	PPB	6/9/2017 14:50	SMPL
I3447-07	MBDC50	747	0.1791 -		ISM02.4	PPB	6/9/2017 14:52	SMPL
I3447-08	MBDC51	1623	0.374 -		ISM02.4	PPB	6/9/2017 14:54	SMPL
CCV017	CCV017	22654	5.0518 -		ISM02.4	PPB	6/9/2017 14:57	SMPL
CCB017	CCB017	-33	0.0057 -		ISM02.4	PPB	6/9/2017 14:59	SMPL