



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB88016

SDG No.: LB88016

Contract: _____ Lab Code: CHEM

Case No.: LB88016

SAS No.: LB88016

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	% RSD	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV1	Cyanide	0.094	0.099	94.9	0	85 - 115		06/12/2017	11:27	LB88016



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB88016

SDG No.: LB88016

Contract: _____ Lab Code: CHEM

Case No.: LB88016

SAS No.: LB88016

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	% RSD	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV1	Cyanide	0.24	0.25	96	0	90 - 110		06/12/2017	11:27	LB88016
CCV2	Cyanide	0.24	0.25	96	0	90 - 110		06/12/2017	11:35	LB88016
CCV3	Cyanide	0.24	0.25	96	0	90 - 110		06/12/2017	11:42	LB88016
CCV4	Cyanide	0.25	0.25	100	0	90 - 110		06/12/2017	11:50	LB88016



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB88016

SDG No.: LB88016

Contract:

Lab Code: CHEM

Case No.: LB88016

SAS No.: LB88016

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	Cyanide	-0.00023	+/-0.005	U		0.005		06/12/2017	11:27	LB88016



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB88016

SDG No.: LB88016

Contract:

Lab Code: CHEM

Case No.: LB88016

SAS No.: LB88016

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	Cyanide	0.00022	+/-0.005	U		0.005		06/12/2017	11:27	LB88016
CCB2	Cyanide	0.00042	+/-0.005	U		0.005		06/12/2017	11:35	LB88016
CCB3	Cyanide	0.00023	+/-0.005	U		0.005		06/12/2017	11:42	LB88016
CCB4	Cyanide	-0.0002	+/-0.005	U		0.005		06/12/2017	11:50	LB88016



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB88016

SDG No.: LB88016

Contract: Lab Code: CHEM

Case No.: LB88016

SAS No.: LB88016

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	-----	------	---	------------------	------------------	---------------



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB88016	Date Received:	
Client Sample ID:	HIGHPB99629	SDG No.:	LB88016
Lab Sample ID:	HIGHPB99629	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.48		1	0.003		0.005	mg/L		06/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB88016	Date Received:	
Client Sample ID:	HIGHPB99663	SDG No.:	LB88016
Lab Sample ID:	HIGHPB99663	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.47		1	0.003		0.005	mg/L		06/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/8/2017 12:00:00 AM
Project:	LB88016	Date Received:	6/8/2017 12:00:00 AM
Client Sample ID:	FK-PS-01-006	SDG No.:	LB88016
Lab Sample ID:	I3567-01	Matrix:	Solid
Level (low/med):	low	% Solid:	88.4

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.0065	U	1	0.036		0.27	mg/Kg	06/09/2017	06/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/8/2017 12:00:00 AM
Project:	LB88016	Date Received:	6/8/2017 12:00:00 AM
Client Sample ID:	HR-06-060817-A	SDG No.:	LB88016
Lab Sample ID:	I3575-01	Matrix:	Solid
Level (low/med):	low	% Solid:	91.3

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.041	U	1	0.035		0.26	mg/Kg	06/09/2017	06/12/2017	9012B

Report of Analysis

Client:		Date Collected:	6/8/2017 12:00:00 AM
Project:	LB88016	Date Received:	6/8/2017 12:00:00 AM
Client Sample ID:	HR-06-060817-B	SDG No.:	LB88016
Lab Sample ID:	I3575-04	Matrix:	Solid
Level (low/med):	low	% Solid:	88.4

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.016	U	1	0.036		0.27	mg/Kg	06/09/2017	06/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/8/2017 12:00:00 AM
Project:	LB88016	Date Received:	6/8/2017 12:00:00 AM
Client Sample ID:	HR-06-060817-C	SDG No.:	LB88016
Lab Sample ID:	I3575-07	Matrix:	Solid
Level (low/med):	low	% Solid:	88.5

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.044	U	1	0.036		0.27	mg/Kg	06/09/2017	06/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/8/2017 12:00:00 AM
Project:	LB88016	Date Received:	6/8/2017 12:00:00 AM
Client Sample ID:	TP-OGR1-OGR2	SDG No.:	LB88016
Lab Sample ID:	I3576-03	Matrix:	Solid
Level (low/med):	low	% Solid:	87.9

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.04	U	1	0.036		0.27	mg/Kg	06/09/2017	06/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/8/2017 12:00:00 AM
Project:	LB88016	Date Received:	6/8/2017 12:00:00 AM
Client Sample ID:	RO-2669	SDG No.:	LB88016
Lab Sample ID:	I3583-01	Matrix:	Solid
Level (low/med):	low	% Solid:	84.7

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.023	U	1	0.038		0.29	mg/Kg	06/10/2017	06/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/8/2017 12:00:00 AM
Project:	LB88016	Date Received:	6/8/2017 12:00:00 AM
Client Sample ID:	RO-2342	SDG No.:	LB88016
Lab Sample ID:	I3583-04	Matrix:	Solid
Level (low/med):	low	% Solid:	88.2

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.045	U	1	0.036		0.27	mg/Kg	06/10/2017	06/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/8/2017 12:00:00 AM
Project:	LB88016	Date Received:	6/8/2017 12:00:00 AM
Client Sample ID:	DUCT-BANK	SDG No.:	LB88016
Lab Sample ID:	I3584-01	Matrix:	Solid
Level (low/med):	low	% Solid:	94.1

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.048	U	1	0.033		0.25	mg/Kg	06/10/2017	06/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/9/2017 12:00:00 AM
Project:	LB88016	Date Received:	6/9/2017 12:00:00 AM
Client Sample ID:	SU-03-060917	SDG No.:	LB88016
Lab Sample ID:	I3592-01	Matrix:	Solid
Level (low/med):	low	% Solid:	92

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.046	U	1	0.034		0.26	mg/Kg	06/10/2017	06/12/2017	9012B

Report of Analysis

Client:		Date Collected:	6/9/2017 12:00:00 AM
Project:	LB88016	Date Received:	6/9/2017 12:00:00 AM
Client Sample ID:	EO-01-060917-A	SDG No.:	LB88016
Lab Sample ID:	I3594-01	Matrix:	Solid
Level (low/med):	low	% Solid:	93.1

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.019	U	1	0.034		0.26	mg/Kg	06/10/2017	06/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/9/2017 12:00:00 AM
Project:	LB88016	Date Received:	6/9/2017 12:00:00 AM
Client Sample ID:	EO-01-060917-B	SDG No.:	LB88016
Lab Sample ID:	I3594-04	Matrix:	Solid
Level (low/med):	low	% Solid:	89.4

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.039	U	1	0.036		0.27	mg/Kg	06/10/2017	06/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/9/2017 12:00:00 AM
Project:	LB88016	Date Received:	6/9/2017 12:00:00 AM
Client Sample ID:	EO-01-060917-C	SDG No.:	LB88016
Lab Sample ID:	I3594-07	Matrix:	Solid
Level (low/med):	low	% Solid:	90.9

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.051	U	1	0.035		0.27	mg/Kg	06/10/2017	06/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/9/2017 12:00:00 AM
Project:	LB88016	Date Received:	6/9/2017 12:00:00 AM
Client Sample ID:	VNJ-204-COMP	SDG No.:	LB88016
Lab Sample ID:	I3598-01	Matrix:	Solid
Level (low/med):	low	% Solid:	62.8

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.048	U	1	0.052		0.39	mg/Kg	06/10/2017	06/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/9/2017 12:00:00 AM
Project:	LB88016	Date Received:	6/9/2017 12:00:00 AM
Client Sample ID:	VNJ-226-COMP	SDG No.:	LB88016
Lab Sample ID:	I3598-05	Matrix:	Solid
Level (low/med):	low	% Solid:	55.6

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.077	U	1	0.057		0.43	mg/Kg	06/10/2017	06/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/9/2017 12:00:00 AM
Project:	LB88016	Date Received:	6/9/2017 12:00:00 AM
Client Sample ID:	HA4-WCA	SDG No.:	LB88016
Lab Sample ID:	I3615-01	Matrix:	Solid
Level (low/med):	low	% Solid:	79

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.37		1	0.039		0.3	mg/Kg	06/10/2017	06/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/9/2017 12:00:00 AM
Project:	LB88016	Date Received:	6/9/2017 12:00:00 AM
Client Sample ID:	HA4-WCB	SDG No.:	LB88016
Lab Sample ID:	I3615-05	Matrix:	Solid
Level (low/med):	low	% Solid:	79.9

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.036	U	1	0.039		0.29	mg/Kg	06/10/2017	06/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB88016	Date Received:	
Client Sample ID:	LOWPB99629	SDG No.:	LB88016
Lab Sample ID:	LOWPB99629	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.011		1	0.003		0.005	mg/L		06/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB88016	Date Received:	
Client Sample ID:	LOWPB99663	SDG No.:	LB88016
Lab Sample ID:	LOWPB99663	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.0088		1	0.003		0.005	mg/L		06/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB88016	Date Received:	
Client Sample ID:	LOWPB99663	SDG No.:	LB88016
Lab Sample ID:	LOWPB99663	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
-----	-----------	-------	------	----	-----	-----	------------	-------	-----------	-----------	----------

Color Before:	Clarity Before:	Texture:
Color After:	Clarity After:	Artifacts:
Comments:		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits



GENCHEM
- 3b -
PREPARATION BLANK SUMMARY

Client: LB88016

SDG No.: LB88016

Instrument: Konelab 20

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	LOD mg/Kg	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB99629BL	Cyanide	SOLID		Batch Number:		PB99629		Prep Date:	06/09/2017	
		0.0017	<0.25	U		0.25		06/12/2017	11:27	LB88016
Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	LOD mg/Kg	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB99663BL	Cyanide	SOLID		Batch Number:		PB99663		Prep Date:	06/10/2017	
		0.0058	<0.25	U		0.25		06/12/2017	11:35	LB88016



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB88016 level: low sdg no.: LB88016
contract: lab code: CHEM case no.: LB88016 sas no.: LB88016
matrix: Solid sample id: I3567-01 client id: FK-PS-01-006MS
Percent Solids for Sample: 88.4 Spiked ID: I3567-01MS Percent Solids for Spike Sample: 88.4

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	1.8		0.036	U	2.2	81.8		



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB88016 level: low sdg no.: LB88016
contract: lab code: CHEM case no.: LB88016 sas no.: LB88016
matrix: Solid sample id: I3567-01 client id: FK-PS-01-006MSD
Percent Solids for Sample: 88.4 Spiked ID: I3567-01MSD Percent Solids for Spike Sample: 88.4

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	1.8		0.036	U	2.2	81.8		



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB88016 level: low sdg no.: LB88016
contract: lab code: CHEM case no.: LB88016 sas no.: LB88016
matrix: Solid sample id: I3583-01 client id: RO-2669MS
Percent Solids for Sample: 84.7 Spiked ID: I3583-01MS Percent Solids for Spike Sample: 84.7

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	1.8		0.038	U	2.3	78.3		



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB88016 level: low sdg no.: LB88016
contract: lab code: CHEM case no.: LB88016 sas no.: LB88016
matrix: Solid sample id: I3583-01 client id: RO-2669MSD
Percent Solids for Sample: 84.7 Spiked ID: I3583-01MSD Percent Solids for Spike Sample: 84.7

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	1.8		0.038	U	2.3	78.3		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB88016 Level: LOW SDG No.: LB88016
 Contract: _____ Lab Code: CHEM Case No.: LB88016 SAS No.: LB88016
 Matrix: Solid Sample ID: I3567-01 Client ID: FK-PS-01-006DUP
 Percent Solids for Sample: 88.4 Duplicate ID I3567-01DUP Percent Solids for Spike Sample: 88.4

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	0.036	U	-0.037	U			

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB88016 **Level:** LOW **SDG No.:** LB88016
Contract: **Lab Code:** CHEM **Case No.:** LB88016 **SAS No.:** LB88016
Matrix: Solid **Sample ID:** I3567-01MS **Client ID:** FK-PS-01-006MSD
Percent Solids for Sample: 88.4 **Duplicate ID** I3567-01MSD **Percent Solids for Spike Sample:** 88.4

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	1.8		1.8		0		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB88016 **Level:** LOW **SDG No.:** LB88016
Contract: **Lab Code:** CHEM **Case No.:** LB88016 **SAS No.:** LB88016
Matrix: Solid **Sample ID:** I3583-01 **Client ID:** RO-2669DUP
Percent Solids for Sample: 84.7 **Duplicate ID** I3583-01DUP **Percent Solids for Spike Sample:** 84.7

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	0.038	U	-0.014	U			

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB88016 **Level:** LOW **SDG No.:** LB88016
Contract: **Lab Code:** CHEM **Case No.:** LB88016 **SAS No.:** LB88016
Matrix: Solid **Sample ID:** I3583-01MS **Client ID:** RO-2669MSD
Percent Solids for Sample: 84.7 **Duplicate ID** I3583-01MSD **Percent Solids for Spike Sample:** 84.7

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	1.8		1.8		0		



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB88016

SDG No.: LB88016

Contract:

Lab Code: CHEM

Case No.: LB88016

SAS No.: LB88016

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB99629BS Cyanide	mg/Kg	4.9	4.7		95.9	85 - 115	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB88016

SDG No.: LB88016

Contract:

Lab Code: CHEM

Case No.: LB88016

SAS No.: LB88016

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB99663BS Cyanide	mg/Kg	4.9	4.8		98	85 - 115	