

LB88021

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

6/12/2017 13:03

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	93.596	0.0	0.091	
ICB1	-0.173	0.0	0.004	
CCV1	235.885	0.0	0.223	
CCB1	-0.377	0.0	0.004	
PB99630BL	-0.261	0.0	0.004	
MIDPB99630	242.695	0.0	0.229	
I3550-01	-0.560	0.0	0.004	
I3550-01DUP	-0.732	0.0	0.004	
I3550-01MS	34.860	0.0	0.037	
I3550-01MSD	34.736	0.0	0.037	
PB99630BS	100.200	0.0	0.097	
CCV2	256.757	0.0	0.242	
CCB2	-0.146	0.0	0.005	

N	13
Mean	76.652
SD	102.1874
CV%	133.31

LB88021

AK

Aquakem v. 7.2AQ1

Results from time period:

Mon Jun 12 12:28:19 2017

Mon Jun 12 12:54:55 2017

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.4252	µg/l	6/12/2017 10:34:12	
5.OPPBCN	A	Total CN	P	3.6977	µg/l	6/12/2017 10:34:13	
10PPBCN	A	Total CN	P	8.9878	µg/l	6/12/2017 10:34:14	
50PPBCN	A	Total CN	P	50.5506	µg/l	6/12/2017 10:34:15	
100PPBCN	A	Total CN	P	101.2416	µg/l	6/12/2017 10:34:16	
250PPBCN	A	Total CN	P	252.4352	µg/l	6/12/2017 10:34:17	
500PPBCN	A	Total CN	P	498.5123	µg/l	6/12/2017 10:34:18	
ICV1	S	Total CN	P	93.5959	µg/l	6/12/2017 12:28:19	
ICB1	S	Total CN	P	-0.1734	µg/l	6/12/2017 12:28:20	
CCV1	S	Total CN	P	235.8849	µg/l	6/12/2017 12:28:21	
CCB1	S	Total CN	P	-0.3772	µg/l	6/12/2017 12:28:22	
PB99630BL	S	Total CN	P	-0.261	µg/l	6/12/2017 12:28:23	
MIDPB99630	S	Total CN	P	242.6947	µg/l	6/12/2017 12:28:25	
I3550-01	S	Total CN	P	-0.5601	µg/l	6/12/2017 12:28:26	
I3550-01DUP	S	Total CN	P	-0.7317	µg/l	6/12/2017 12:28:27	
I3550-01MS	S	Total CN	P	34.8603	µg/l	6/12/2017 12:28:28	
I3550-01MSD	S	Total CN	P	34.7362	µg/l	6/12/2017 12:28:29	
PB99630BS	S	Total CN	P	100.1995	µg/l	6/12/2017 12:48:44	
CCV2	S	Total CN	P	256.7569	µg/l	6/12/2017 12:48:45	
CCB2	S	Total CN	P	-0.1455	µg/l	6/12/2017 12:48:46	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AL

6/12/2017 10:39


Test Total CN

Accepted 6/12/2017 10:39

Factor 1083
 Bias 0.005

Coeff. of det. 0.999937

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.4252	0.0000	
2	5.0PPBCN	0.008	3.6977	5.0000	
3	10PPBCN	0.013	8.9878	10.0000	
4	50PPBCN	0.051	50.5506	50.0000	
5	100PPBCN	0.098	101.2416	100.0000	
6	250PPBCN	0.238	252.4352	250.0000	
7	500PPBCN	0.465	498.5123	500.0000	