

688082

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AK

6/14/2017 14:43

Test: Reactive Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.876	0.0	0.091	
ICB1	-0.890	0.0	0.004	
CCV1	244.128	0.0	0.226	
CCB1	-0.663	0.0	0.004	
PB99779BL	-0.788	0.0	0.004	
I3660-03	-1.266	0.0	0.003	
I3660-03DUP	-1.111	0.0	0.004	
I3660-07	-1.367	0.0	0.003	
I3660-11	-0.985	0.0	0.004	
I3660-15	-0.970	0.0	0.004	
I3660-19	-0.814	0.0	0.004	
I3665-05	-1.047	0.0	0.004	
I3665-10	-1.035	0.0	0.004	
I3670-04	-0.568	0.0	0.004	
CCV2	243.591	0.0	0.226	
CCB2	-0.784	0.0	0.004	
PB99776BL	-1.182	0.0	0.004	
I3644-02	-1.017	0.0	0.004	
I3644-02DUP	-1.088	0.0	0.004	
CCV3	249.628	0.0	0.231	
CCB3	-0.186	0.0	0.004	

N	21
Mean	38.879
SD	89.0290
CV%	228.99

CB88082

AK

Aquakem v. 7.2AQ1

Results from time period:

Wed Jun 14 12:37:38 2017

Wed Jun 14 14:31:33 2017

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	-0.9482	µg/l	6/14/2017 13:57:47	
5.0PPBCN	A	Reactive CN	P	3.7158	µg/l	6/14/2017 13:57:48	
10PPBCN	A	Reactive CN	P	9.061	µg/l	6/14/2017 13:57:49	
50PPBCN	A	Reactive CN	P	48.7836	µg/l	6/14/2017 13:57:50	
100PPBCN	A	Reactive CN	P	102.0725	µg/l	6/14/2017 13:57:51	
250PPBCN	A	Reactive CN	P	255.153	µg/l	6/14/2017 13:57:52	
500PPBCN	A	Reactive CN	P	497.1622	µg/l	6/14/2017 13:57:53	
ICV1	S	Reactive CN	P	94.8765	µg/l	6/14/2017 14:24:17	
ICB1	S	Reactive CN	P	-0.8903	µg/l	6/14/2017 14:24:18	
CCV1	S	Reactive CN	P	244.1281	µg/l	6/14/2017 14:24:19	
CCB1	S	Reactive CN	P	-0.6633	µg/l	6/14/2017 14:24:20	
PB99779BL	S	Reactive CN	P	-0.7876	µg/l	6/14/2017 14:24:21	
I3660-03	S	Reactive CN	P	-1.2657	µg/l	6/14/2017 14:24:22	
I3660-03DUP	S	Reactive CN	P	-1.1108	µg/l	6/14/2017 14:24:23	
I3660-07	S	Reactive CN	P	-1.3669	µg/l	6/14/2017 14:24:24	
I3660-11	S	Reactive CN	P	-0.9847	µg/l	6/14/2017 14:24:25	
I3660-15	S	Reactive CN	P	-0.9695	µg/l	6/14/2017 14:24:26	
I3660-19	S	Reactive CN	P	-0.814	µg/l	6/14/2017 14:24:27	
I3665-05	S	Reactive CN	P	-1.0475	µg/l	6/14/2017 14:31:24	
I3665-10	S	Reactive CN	P	-1.035	µg/l	6/14/2017 14:31:25	
I3670-04	S	Reactive CN	P	-0.5684	µg/l	6/14/2017 14:31:26	
CCV2	S	Reactive CN	P	243.5906	µg/l	6/14/2017 14:31:27	
CCB2	S	Reactive CN	P	-0.7838	µg/l	6/14/2017 14:31:28	
PB99776BL	S	Reactive CN	P	-1.1819	µg/l	6/14/2017 14:31:29	
I3644-02	S	Reactive CN	P	-1.0167	µg/l	6/14/2017 14:31:30	
I3644-02DUP	S	Reactive CN	P	-1.0883	µg/l	6/14/2017 14:31:31	
CCV3	S	Reactive CN	P	249.6284	µg/l	6/14/2017 14:31:32	
CCB3	S	Reactive CN	P	-0.1859	µg/l	6/14/2017 14:31:33	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

6/14/2017 13:58


Test Total CN

Accepted 6/14/2017 13:58

Factor 1102
Bias 0.005

Coeff. of det. 0.999787

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.9482	0.0000	
2	5.0PPBCN	0.008	3.7158	5.0000	
3	10PPBCN	0.013	9.0610	10.0000	
4	50PPBCN	0.049	48.7836	50.0000	
5	100PPBCN	0.097	102.0725	100.0000	
6	250PPBCN	0.236	255.1530	250.0000	
7	500PPBCN	0.456	497.1622	500.0000	