



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB88086	Date Received:	
Client Sample ID:	HIGHPB99783	SDG No.:	LB88086
Lab Sample ID:	HIGHPB99783	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.49		1	0.003		0.005	mg/L		06/14/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/13/2017 12:00:00 AM
Project:	LB88086	Date Received:	6/12/2017 12:00:00 AM
Client Sample ID:	DRUMS-M47409-COMP	SDG No.:	LB88086
Lab Sample ID:	I3644-01	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.0016	U	1	0.003		0.005	mg/L	06/14/2017	06/14/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/13/2017 12:00:00 AM
Project:	LB88086	Date Received:	6/13/2017 12:00:00 AM
Client Sample ID:	RO-132	SDG No.:	LB88086
Lab Sample ID:	I3660-01	Matrix:	Solid
Level (low/med):	low	% Solid:	95.8

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.017	U	1	0.033		0.25	mg/Kg	06/14/2017	06/14/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/13/2017 12:00:00 AM
Project:	LB88086	Date Received:	6/13/2017 12:00:00 AM
Client Sample ID:	RO-153	SDG No.:	LB88086
Lab Sample ID:	I3660-05	Matrix:	Solid
Level (low/med):	low	% Solid:	97.6

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.057	U	1	0.033		0.25	mg/Kg	06/14/2017	06/14/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/13/2017 12:00:00 AM
Project:	LB88086	Date Received:	6/13/2017 12:00:00 AM
Client Sample ID:	RO-351	SDG No.:	LB88086
Lab Sample ID:	I3660-09	Matrix:	Solid
Level (low/med):	low	% Solid:	95.3

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.02	U	1	0.033		0.25	mg/Kg	06/14/2017	06/14/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/13/2017 12:00:00 AM
Project:	LB88086	Date Received:	6/13/2017 12:00:00 AM
Client Sample ID:	RO-350	SDG No.:	LB88086
Lab Sample ID:	I3660-13	Matrix:	Solid
Level (low/med):	low	% Solid:	94.9

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.033	U	1	0.033		0.25	mg/Kg	06/14/2017	06/14/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/13/2017 12:00:00 AM
Project:	LB88086	Date Received:	6/13/2017 12:00:00 AM
Client Sample ID:	RO-72-11941	SDG No.:	LB88086
Lab Sample ID:	I3660-17	Matrix:	Solid
Level (low/med):	low	% Solid:	92.6

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.062	U	1	0.034		0.26	mg/Kg	06/14/2017	06/14/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/13/2017 12:00:00 AM
Project:	LB88086	Date Received:	6/13/2017 12:00:00 AM
Client Sample ID:	ONSITE-001-A	SDG No.:	LB88086
Lab Sample ID:	I3663-01	Matrix:	Solid
Level (low/med):	low	% Solid:	88.4

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.23	J	1	0.036		0.27	mg/Kg	06/14/2017	06/14/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/13/2017 12:00:00 AM
Project:	LB88086	Date Received:	6/13/2017 12:00:00 AM
Client Sample ID:	ONSITE-001-B	SDG No.:	LB88086
Lab Sample ID:	I3663-03	Matrix:	Solid
Level (low/med):	low	% Solid:	86.8

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.65		1	0.037		0.28	mg/Kg	06/14/2017	06/14/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/13/2017 12:00:00 AM
Project:	LB88086	Date Received:	6/13/2017 12:00:00 AM
Client Sample ID:	ONSITE-001A	SDG No.:	LB88086
Lab Sample ID:	I3664-01	Matrix:	Solid
Level (low/med):	low	% Solid:	92

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.017	U	1	0.035		0.26	mg/Kg	06/14/2017	06/14/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/13/2017 12:00:00 AM
Project:	LB88086	Date Received:	6/13/2017 12:00:00 AM
Client Sample ID:	ONSITE-001B	SDG No.:	LB88086
Lab Sample ID:	I3664-03	Matrix:	Solid
Level (low/med):	low	% Solid:	91.4

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.064	U	1	0.034		0.26	mg/Kg	06/14/2017	06/14/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/13/2017 12:00:00 AM
Project:	LB88086	Date Received:	6/13/2017 12:00:00 AM
Client Sample ID:	HA3-WC1	SDG No.:	LB88086
Lab Sample ID:	I3665-04	Matrix:	Solid
Level (low/med):	low	% Solid:	76.6

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.038	U	1	0.042		0.32	mg/Kg	06/14/2017	06/14/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/13/2017 12:00:00 AM
Project:	LB88086	Date Received:	6/13/2017 12:00:00 AM
Client Sample ID:	HA3-WC2	SDG No.:	LB88086
Lab Sample ID:	I3665-09	Matrix:	Solid
Level (low/med):	low	% Solid:	82.2

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.055	U	1	0.039		0.3	mg/Kg	06/14/2017	06/14/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	6/13/2017 12:00:00 AM
Project:	LB88086	Date Received:	6/13/2017 12:00:00 AM
Client Sample ID:	TP-CLINTON-AVE-1-2	SDG No.:	LB88086
Lab Sample ID:	I3670-01	Matrix:	Solid
Level (low/med):	low	% Solid:	88.1

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.046	U	1	0.036		0.27	mg/Kg	06/14/2017	06/14/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB88086	Date Received:	
Client Sample ID:	LOWPB99783	SDG No.:	LB88086
Lab Sample ID:	LOWPB99783	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.0094		1	0.003		0.005	mg/L		06/14/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB88086	Date Received:	
Client Sample ID:	LOWPB99783	SDG No.:	LB88086
Lab Sample ID:	LOWPB99783	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
-----	-----------	-------	------	----	-----	-----	------------	-------	-----------	-----------	----------

Color Before:	Clarity Before:	Texture:
Color After:	Clarity After:	Artifacts:
Comments:		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB88086

SDG No.: LB88086

Contract: _____ Lab Code: CHEM

Case No.: LB88086

SAS No.: LB88086

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	% RSD	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV1	Cyanide	0.24	0.25	96	0	90 - 110		06/14/2017	15:26	LB88086
CCV2	Cyanide	0.24	0.25	96	0	90 - 110		06/14/2017	15:33	LB88086
CCV3	Cyanide	0.25	0.25	100	0	90 - 110		06/14/2017	15:38	LB88086
CCV4	Cyanide	0.25	0.25	100	0	90 - 110		06/14/2017	15:58	LB88086



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB88086 **SDG No.:** LB88086
Contract: **Lab Code:** CHEM **Case No.:** LB88086 **SAS No.:** LB88086

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	Cyanide	-0.00059	+/-0.005	U		0.005		06/14/2017	15:26	LB88086



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB88086

SDG No.: LB88086

Contract:

Lab Code: CHEM

Case No.: LB88086

SAS No.: LB88086

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	Cyanide	-0.00056	+/-0.005	U		0.005		06/14/2017	15:26	LB88086
CCB2	Cyanide	-0.00059	+/-0.005	U		0.005		06/14/2017	15:33	LB88086
CCB3	Cyanide	-0.00031	+/-0.005	U		0.005		06/14/2017	15:44	LB88086
CCB4	Cyanide	-0.00063	+/-0.005	U		0.005		06/14/2017	15:58	LB88086



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB88086

SDG No.: LB88086

Contract: Lab Code: CHEM

Case No.: LB88086

SAS No.: LB88086

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	-----	------	---	------------------	------------------	---------------



GENCHEM
- 3b -
PREPARATION BLANK SUMMARY

Client: LB88086

SDG No.: LB88086

Instrument: Konelab 20

Sample ID	Analyte	Result (mg/L)	Acceptance Limit	Conc Qual	LOD mg/L	CRQL mg/L	M	Analysis Date	Analysis Time	Run
PB99782BL	Cyanide	WATER		Batch Number:		PB99782		Prep Date:	06/14/2017	
		-0.00044	<0.005	U		0.005		06/14/2017	15:38	LB88086
Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	LOD mg/Kg	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB99783BL	Cyanide	SOLID		Batch Number:		PB99783		Prep Date:	06/14/2017	
		-0.023	<0.25	U		0.25		06/14/2017	15:26	LB88086



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB88086 level: low sdg no.: LB88086
contract: lab code: CHEM case no.: LB88086 sas no.: LB88086
matrix: WATER sample id: I3644-01 client id: DRUMS-M47409-COMPMS
Percent Solids for Sample: 0 Spiked ID: I3644-01MS Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/L	75 - 125	0.026		0.003	U	0.04	65	*	



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB88086 level: low sdg no.: LB88086
contract: lab code: CHEM case no.: LB88086 sas no.: LB88086
matrix: WATER sample id: I3644-01 client id: DRUMS-M47409-COMPMSD
Percent Solids for Sample: 0 Spiked ID: I3644-01MSD Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/L	75 - 125	0.025		0.003	U	0.04	62.5	*	



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB88086 level: low sdg no.: LB88086
contract: lab code: CHEM case no.: LB88086 sas no.: LB88086
matrix: Solid sample id: I3660-01 client id: RO-132MS
Percent Solids for Sample: 95.8 Spiked ID: I3660-01MS Percent Solids for Spike Sample: 95.8

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	1.8		0.033	U	2	90		



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB88086 level: low sdg no.: LB88086
contract: lab code: CHEM case no.: LB88086 sas no.: LB88086
matrix: Solid sample id: I3660-01 client id: RO-132MSD
Percent Solids for Sample: 95.8 Spiked ID: I3660-01MSD Percent Solids for Spike Sample: 95.8

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	1.7		0.033	U	2	85		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB88086 **Level:** LOW **SDG No.:** LB88086
Contract: **Lab Code:** CHEM **Case No.:** LB88086 **SAS No.:** LB88086
Matrix: WATER **Sample ID:** I3644-01 **Client ID:** DRUMS-M47409-COMPDUP
Percent Solids for Sample: 0 **Duplicate ID** I3644-01DUP **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/L	20	0.003	U	0.0018	U			

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB88086 **Level:** LOW **SDG No.:** LB88086
Contract: **Lab Code:** CHEM **Case No.:** LB88086 **SAS No.:** LB88086
Matrix: WATER **Sample ID:** I3644-01MS **Client ID:** DRUMS-M47409-COMPMSD
Percent Solids for Sample: 0 **Duplicate ID** I3644-01MSD **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/L	20	0.026		0.025		4		

GENCHEM**- 6 -****DUPLICATE SAMPLE SUMMARY**

Client: LB88086 **Level:** LOW **SDG No.:** LB88086
Contract: **Lab Code:** CHEM **Case No.:** LB88086 **SAS No.:** LB88086
Matrix: Solid **Sample ID:** I3660-01 **Client ID:** RO-132DUP
Percent Solids for Sample: 95.8 **Duplicate ID** I3660-01DUP **Percent Solids for Spike Sample:** 95.8

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	0.033	U	-0.036	U			

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB88086 **Level:** LOW **SDG No.:** LB88086
Contract: **Lab Code:** CHEM **Case No.:** LB88086 **SAS No.:** LB88086
Matrix: Solid **Sample ID:** I3660-01MS **Client ID:** RO-132MSD
Percent Solids for Sample: 95.8 **Duplicate ID** I3660-01MSD **Percent Solids for Spike Sample:** 95.8

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	1.8		1.7		6		



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB88086

SDG No.: LB88086

Contract:

Lab Code: CHEM

Case No.: LB88086

SAS No.: LB88086

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB99782BS Cyanide	mg/L	0.1	0.1		100	85 - 115	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB88086

SDG No.: LB88086

Contract:

Lab Code: CHEM

Case No.: LB88086

SAS No.: LB88086

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB99783BS Cyanide	mg/Kg	4.9	4.8		98	85 - 115	

Daily Analysis Runlog For Sequence/QC Batch ID # LB88086

Review By		Review On				
SubDirectory		LB88086		Test : Cyanide		
STD. NAME		STD REF.#				
ICAL Standard		WP55636,WP55637,WP55638WP55639WP55640,WP55641,WP55642,WP55643				
ICV Standard		W1541				
CCV Standard		WP55645				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		WP54521				
Chk Standard		WP55644,WP51626,WP55355				

Sr#	SampleID	ClientID	QcType	Date	Comment	Status
1	0.0PPBCN	0.0PPBCN	CAL1	06/14/17 13:57		OK
2	5.0PPBCN	5.0PPBCN	CAL2	06/14/17 13:57		OK
3	10PPBCN	10PPBCN	CAL3	06/14/17 13:57		OK
4	50PPBCN	50PPBCN	CAL4	06/14/17 13:57		OK
5	100PPBCN	100PPBCN	CAL5	06/14/17 13:57		OK
6	250PPBCN	250PPBCN	CAL6	06/14/17 13:57		OK
7	500PPBCN	500PPBCN	CAL7	06/14/17 13:57		OK
8	ICV1	ICV1	ICV	06/14/17 15:26		OK
9	ICB1	ICB1	ICB	06/14/17 15:26		OK
10	CCV1	CCV1	CCV	06/14/17 15:26		OK
11	CCB1	CCB1	CCB	06/14/17 15:26		OK
12	PB99783BL	PB99783BL	MB	06/14/17 15:26		OK
13	PB99783BS	PB99783BS	LCS	06/14/17 15:26		OK
14	HIGHPB99783	HIGHPB99783	SAM	06/14/17 15:26		OK
15	I3660-01	RO-132	SAM	06/14/17 15:26		OK
16	I3660-01DUP	RO-132DUP	DUP	06/14/17 15:26		OK
17	I3660-01MS	RO-132MS	MS	06/14/17 15:26		OK
18	I3660-01MSD	RO-132MSD	MSD	06/14/17 15:33		OK
19	I3660-05	RO-153	SAM	06/14/17 15:33		OK
20	I3660-09	RO-351	SAM	06/14/17 15:33		OK
21	CCV2	CCV2	CCV	06/14/17 15:33		OK

Daily Analysis Runlog For Sequence/QC Batch ID # LB88086

Review By		Review On				
SubDirectory		LB88086		Test : Cyanide		
STD. NAME		STD REF.#				
ICAL Standard		WP55636,WP55637,WP55638WP55639WP55640,WP55641,WP55642,WP55643				
ICV Standard		W1541				
CCV Standard		WP55645				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		WP54521				
Chk Standard		WP55644,WP51626,WP55355				
22	CCB2	CCB2	CCB	06/14/17 15:33		OK
23	I3660-13	RO-350	SAM	06/14/17 15:33		OK
24	I3660-17	RO-72-11941	SAM	06/14/17 15:33		OK
25	I3663-01	ONSITE-001-A	SAM	06/14/17 15:33		OK
26	I3663-03	ONSITE-001-B	SAM	06/14/17 15:33		OK
27	I3664-01	ONSITE-001A	SAM	06/14/17 15:33		OK
28	I3664-03	ONSITE-001B	SAM	06/14/17 15:33		OK
29	I3665-04	HA3-WC1	SAM	06/14/17 15:38		OK
30	I3665-09	HA3-WC2	SAM	06/14/17 15:38		OK
31	I3670-01	TP-CLINTON-AVE-1-	SAM	06/14/17 15:38		OK
32	PB99782BL	PB99782BL	MB	06/14/17 15:38		OK
33	CCV3	CCV3	CCV	06/14/17 15:38		OK
34	CCB3	CCB3	CCB	06/14/17 15:44		OK
35	PB99782BS	PB99782BS	LCS	06/14/17 15:44		OK
36	I3644-01	DRUMS-M47409-CO	SAM	06/14/17 15:44		OK
37	I3644-01DUP	DRUMS-M47409-CO	DUP	06/14/17 15:44		OK
38	I3644-01MS	DRUMS-M47409-CO	MS	06/14/17 15:44		OK
39	I3644-01MSD	DRUMS-M47409-CO	MSD	06/14/17 15:49		OK
40	LOWPB99783	LOWPB99783	SAM	06/14/17 15:56		OK
41	CCV4	CCV4	CCV	06/14/17 15:58		OK
42	CCB4	CCB4	CCB	06/14/17 15:58		OK