

LB88248

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

6/21/2017 12:27

Test: Phenolics-

AK
6-21-17

mjil

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.964	0.0	0.106	
ICB1	0.009	0.0	0.003	
CCV1	0.988	0.0	0.108	
CCB1	0.007	0.0	0.003	
PB99957BL	0.004	0.0	0.003	
PB99957BS	0.971	0.0	0.106	
I3757-05	0.010	0.0	0.003	
PB99958BL	0.000	0.0	0.002	
PB99958BS	0.962	0.0	0.105	
I3757-01	0.011	0.0	0.004	
I3757-06	0.100	0.0	0.013	
I3757-02	0.102	0.0	0.013	
CCV2	0.993	0.0	0.109	
CCB2	0.007	0.0	0.003	

N 14
Mean 0.366
SD 0.4724
CV% 129.01

Aquakem v. 7.2AQ1

Results from time period:

Wed Jun 21 10:16:42 2017

Wed Jun 21 11:52:33 2017

AK

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0557	mg/l	6/21/2017 10:16:42	
0.1PPM	A	Phenolics-	P	0.1001	mg/l	6/21/2017 10:16:43	
0.25PPM	A	Phenolics-	P	0.2544	mg/l	6/21/2017 10:16:44	
0.50PPM	A	Phenolics-	P	0.4932	mg/l	6/21/2017 10:16:45	
1.0PPM	A	Phenolics-	P	0.9911	mg/l	6/21/2017 10:16:46	
2.0PPM	A	Phenolics-	P	2.0055	mg/l	6/21/2017 10:16:47	
ICV1	S	Phenolics-	P	0.9636	mg/l	6/21/2017 10:52:32	
ICB1	S	Phenolics-	P	0.0093	mg/l	6/21/2017 10:52:33	
CCV1	S	Phenolics-	P	0.9875	mg/l	6/21/2017 10:52:34	
CCB1	S	Phenolics-	P	0.0071	mg/l	6/21/2017 10:52:35	
PB99957BL	S	Phenolics-	P	0.0038	mg/l	6/21/2017 10:52:36	
PB99957BS	S	Phenolics-	P	0.9708	mg/l	6/21/2017 10:52:37	
I3757-05	S	Phenolics-	P	0.0101	mg/l	6/21/2017 10:52:39	
PB99958BL	S	Phenolics-	P	0.0002	mg/l	6/21/2017 10:52:42	
PB99958BS	S	Phenolics-	P	0.9622	mg/l	6/21/2017 10:52:43	
I3757-01	S	Phenolics-	P	0.0114	mg/l	6/21/2017 10:58:50	
I3757-06	S	Phenolics-	P	0.0997	mg/l	6/21/2017 11:52:30	
I3757-02	S	Phenolics-	P	0.1017	mg/l	6/21/2017 11:52:31	
CCV2	S	Phenolics-	P	0.9925	mg/l	6/21/2017 11:52:32	
CCB2	S	Phenolics-	P	0.0067	mg/l	6/21/2017 11:52:33	

UB 88248

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

6/21/2017 10:24

Test Phenolics-


Accepted 6/21/2017 10:24

Factor 9.343

Bias 0.002

Coeff. of det. 0.999926

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0557	0.0500	
2	Phenol-2	0.013	0.1001	0.1000	
3	Phenol-2	0.030	0.2544	0.2500	
4	Phenol-2	0.055	0.4932	0.5000	
5	Phenol-2	0.109	0.9911	1.0000	
6	Phenol-2	0.217	2.0055	2.0000	