

CB 88441

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092

Reviewed by : AK

6/29/2017 15:14

Test: Phenolics-

Sample Id	Result ^{mg/L}	Dil. 1 +	Response	Errors
ICV1	0.949	0.0	0.104	
ICB1	0.003	0.0	0.002	
CCV1	0.946	0.0	0.104	
CCB1	0.001	0.0	0.002	
PB100251BL	0.005	0.0	0.003	
PB100251BS	0.950	0.0	0.105	
I3952-01	0.005	0.0	0.003	
I3952-01DUP	0.003	0.0	0.002	
I3952-01MS	0.988	0.0	0.109	
I3952-01MSD	0.984	0.0	0.108	
CCV2	0.947	0.0	0.104	
CCB2	-0.974	0.0	-0.103	

N 12
Mean 0.401
SD 0.6441
CV% 160.75

Aquakem v. 7.2AQ1

Results from time period:

Thu Jun 29 14:38:08 2017

Thu Jun 29 15:13:13 2017

AK

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0531	mg/l	6/29/2017 14:38:08	
0.1PPM	A	Phenolics-	P	0.1056	mg/l	6/29/2017 14:38:09	
0.25PPM	A	Phenolics-	P	0.2522	mg/l	6/29/2017 14:38:10	
0.50PPM	A	Phenolics-	P	0.4942	mg/l	6/29/2017 14:38:11	
1.0PPM	A	Phenolics-	P	0.9882	mg/l	6/29/2017 14:38:12	
2.0PPM	A	Phenolics-	P	2.0067	mg/l	6/29/2017 14:38:13	
ICV1	S	Phenolics-	P	0.9494	mg/l	6/29/2017 15:13:02	
ICB1	S	Phenolics-	P	0.0031	mg/l	6/29/2017 15:13:03	
CCV1	S	Phenolics-	P	0.9462	mg/l	6/29/2017 15:13:04	
CCB1	S	Phenolics-	P	0.0014	mg/l	6/29/2017 15:13:05	
PB100251BL	S	Phenolics-	P	0.0046	mg/l	6/29/2017 15:13:06	
PB100251BS	S	Phenolics-	P	0.9505	mg/l	6/29/2017 15:13:07	
I3952-01	S	Phenolics-	P	0.0048	mg/l	6/29/2017 15:13:08	
I3952-01DUP	S	Phenolics-	P	0.0032	mg/l	6/29/2017 15:13:09	
I3952-01MS	S	Phenolics-	P	0.9878	mg/l	6/29/2017 15:13:10	
I3952-01MSD	S	Phenolics-	P	0.9842	mg/l	6/29/2017 15:13:11	
CCV2	S	Phenolics-	P	0.947	mg/l	6/29/2017 15:13:12	
CCB2	S	Phenolics-	P	-0.974	mg/l	6/29/2017 15:13:13	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

6/29/2017 16:01


Test Phenolics-

Accepted 6/29/2017 14:45

Factor 9.274
Bias 0.002

Coeff. of det. 0.999906

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0531	0.0500	
2	Phenol-2	0.013	0.1056	0.1000	
3	Phenol-2	0.029	0.2522	0.2500	
4	Phenol-2	0.055	0.4942	0.5000	
5	Phenol-2	0.109	0.9882	1.0000	
6	Phenol-2	0.218	2.0067	2.0000	