

7471A


LB 88580

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38307



A= 0.0000e+000

B= 2.6089e-004

C= 7.1399e-002

Rho= 0.9998398

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.024	0.024	-181	0.000	-181				
0.2	0.200	0.257	0.057	711	0.0 %	711				
2.5	2.500	2.485	-0.015	9253	0.0 %	9253				
5.0	5.000	4.866	-0.134	18379	0.0 %	18379				
7.5	7.500	7.502	0.002	28480	0.0 %	28480	0			
10.0	10.000	10.065	0.065	38307	0.0 %	38307				

LB88580 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 07 Jul 2017 10:43:39

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-181	-	0.0007471A	PPB		07 Jul 2017 10:59:55	S	Std
0.2 - 1	50.2	711	-	0.20007471A	PPB		07 Jul 2017 11:02:12	S	Std
2.5 - 1	52.5	9253	-	2.50007471A	PPB		07 Jul 2017 11:04:30	S	Std
5.0 - 1	55	18379	-	5.00007471A	PPB		07 Jul 2017 11:06:48	S	Std
7.5 - 1	57.8	28480	-	7.50007471A	PPB		07 Jul 2017 11:11:41	S	Std
10.0 - 1	510	38307	-	10.00007471A	PPB		07 Jul 2017 11:16:37	S	Std
ICV51 - 1	ICV51	14828	3.9399	-7471A	PPB		07 Jul 2017 11:20:32	U	SMPL
ICB51 - 1	ICB51	35	0.0805	-7471A	PPB		07 Jul 2017 11:22:48	U	SMPL
CCV59 - 1	CCV59	19163	5.0709	-7471A	PPB		07 Jul 2017 11:25:03	U	SMPL
CCB59 - 1	CCB59	-104	0.0443	-7471A	PPB		07 Jul 2017 11:27:19	U	SMPL
CRA - 1	CRA	570	0.2201	-7471A	PPB		07 Jul 2017 11:32:16	U	SMPL
HighStd - 1	HighStd	39291	10.3222	-7471A	PPB		07 Jul 2017 11:34:31	U	SMPL
ChkStd - 1	ChkStd	25766	6.7936	-7471A	PPB		07 Jul 2017 11:36:47	U	SMPL
PB100462BL - 1	PBS	-102	0.0448	-7471A	PPB		07 Jul 2017 11:39:05	U	SMPL
PB100462BS - 1	LCSS	13466	3.5846	-7471A	PPB		07 Jul 2017 11:41:22	U	SMPL
I3992-01 - 1	CLINTON-NJ-QUARRY-STONE-DUST	1014	0.3359	-7471A	PPB		07 Jul 2017 11:43:39	U	SMPL
I4061-01 - 1	SP-1-A	1306	0.4121	-7471A	PPB		07 Jul 2017 11:45:56	U	SMPL
I4061-04 - 1	SP-1-B	3397	0.9577	-7471A	PPB		07 Jul 2017 11:48:14	U	SMPL
I4061-07 - 1	SP-1-C	1050	0.3453	-7471A	PPB		07 Jul 2017 11:50:31	U	SMPL
I4061-10 - 1	SP-1-D	1586	0.4852	-7471A	PPB		07 Jul 2017 11:52:47	U	SMPL
CCV60 - 1	CCV60	19429	5.1403	-7471A	PPB		07 Jul 2017 11:55:03	U	SMPL
CCB60 - 1	CCB60	-65	0.0544	-7471A	PPB		07 Jul 2017 11:57:19	U	SMPL
I4062-01 - 1	SU-02-070517	5650	1.5454	-7471A	PPB		07 Jul 2017 11:59:35	U	SMPL
I4063-01 - 1	NB-01-070517-A	421202	HIGH	-7471A	PPB		07 Jul 2017 12:01:51	X	
I4063-04 - 1	NB-01-070517-B	18	0.0761	-7471A	PPB		07 Jul 2017 12:06:47	U	SMPL
I4063-07 - 1	NB-01-070517-C	478	0.1961	-7471A	PPB		07 Jul 2017 12:09:03	U	SMPL
I4064-01 - 1	PET-BF001-01	283	0.1452	-7471A	PPB		07 Jul 2017 12:11:20	U	SMPL
I4064-01DUP - 1	PET-BF001-01DUP	338	0.1596	-7471A	PPB		07 Jul 2017 12:13:37	U	SMPL
I4064-02 - 1	I4064-01MS	14912	3.9618	-7471A	PPB		07 Jul 2017 12:15:54	U	SMPL
I4064-03 - 1	I4064-01MSD	14904	3.9598	-7471A	PPB		07 Jul 2017 12:18:12	U	SMPL
I4064-04 - 1	PET-BF001-02	260	0.1392	-7471A	PPB		07 Jul 2017 12:20:29	U	SMPL
I4064-05 - 1	PET-BF002-01	374	0.1690	-7471A	PPB		07 Jul 2017 12:22:46	U	SMPL
CCV61 - 1	CCV61	19373	5.1257	-7471A	PPB		07 Jul 2017 12:25:02	U	SMPL
CCB61 - 1	CCB61	-58	0.0563	-7471A	PPB		07 Jul 2017 12:27:18	U	SMPL
I4064-06 - 1	PET-BF003-01	397	0.1750	-7471A	PPB		07 Jul 2017 12:29:35	U	SMPL
I4064-07 - 1	PET-BF004-01	150	0.1105	-7471A	PPB		07 Jul 2017 12:31:51	U	SMPL
I4064-08 - 1	PET-BF005-01	649	0.2407	-7471A	PPB		07 Jul 2017 12:34:08	U	SMPL
I4064-09 - 1	PET-BF006-01	201	0.1238	-7471A	PPB		07 Jul 2017 12:36:24	U	SMPL
I4064-10 - 1	PET-BF007-01	166	0.1147	-7471A	PPB		07 Jul 2017 12:38:41	U	SMPL
I4064-11 - 1	PET-BF008-01	558	0.2170	-7471A	PPB		07 Jul 2017 12:40:58	U	SMPL
I4064-12 - 1	PET-BF009-01	373	0.1687	-7471A	PPB		07 Jul 2017 12:43:15	U	SMPL
PB100463BL - 1	PBS	-21	0.0659	-7471A	PPB		07 Jul 2017 12:45:33	U	SMPL
PB100463BS - 1	LCSS	13937	3.7075	-7471A	PPB		07 Jul 2017 12:47:51	U	SMPL
I3757-01 - 1	LOD-SOIL-01-QT3-2017	52	0.0850	-7471A	PPB		07 Jul 2017 12:50:07	U	SMPL
CCV62 - 1	CCV62	19285	5.1027	-7471A	PPB		07 Jul 2017 12:55:23	U	SMPL
CCB62 - 1	CCB62	-103	0.0445	-7471A	PPB		07 Jul 2017 12:57:40	U	SMPL
I3757-02 - 1	LOQ-SOIL-02-QT3-2017	1506	0.4643	-7471A	PPB		07 Jul 2017 12:59:56	U	SMPL
I4069-01 - 1	SB-1(10.5-11.0)	207	0.1254	-7471A	PPB		07 Jul 2017 13:02:13	U	SMPL
I4069-02 - 1	SB-1	330	0.1575	-7471A	PPB		07 Jul 2017 13:04:29	U	SMPL
I4069-03 - 1	SB-2(12.0-12.5)	764	0.2707	-7471A	PPB		07 Jul 2017 13:06:45	U	SMPL
I4069-03DUP - 1	SB-2(12.0-12.5)DUP	897	0.3054	-7471A	PPB		07 Jul 2017 13:09:02	U	SMPL
I4069-03MS - 1	SB-2(12.0-12.5)MS	15205	4.0383	-7471A	PPB		07 Jul 2017 13:11:19	U	SMPL
I4069-03MSD - 1	SB-2(12.0-12.5)MSD	15039	3.9950	-7471A	PPB		07 Jul 2017 13:13:37	U	SMPL
I4069-04 - 1	SB-2	47	0.0837	-7471A	PPB		07 Jul 2017 13:15:54	U	SMPL
I4069-05 - 1	SB-3(6.5-7.0)	38000	9.9854	-7471A	PPB		07 Jul 2017 13:18:12	U	SMPL
I4069-06 - 1	SB-3	41996	11.0279	-7471A	PPB		07 Jul 2017 13:20:28	U	SMPL
CCV63 - 1	CCV63	18365	4.8627	-7471A	PPB		07 Jul 2017 13:22:46	U	SMPL
CCB63 - 1	CCB63	-154	0.0312	-7471A	PPB		07 Jul 2017 13:25:04	U	SMPL
I4070-01 - 1	AU-05-070517-A	1468	0.4544	-7471A	PPB		07 Jul 2017 13:27:21	U	SMPL
I4070-04 - 1	AU-05-070517-B	1914	0.5707	-7471A	PPB		07 Jul 2017 13:29:37	U	SMPL
I4070-07 - 1	AU-05-070517-C	864	0.2968	-7471A	PPB		07 Jul 2017 13:31:53	U	SMPL
I4070-10 - 1	AU-05-070517-D	748	0.2665	-7471A	PPB		07 Jul 2017 13:34:10	U	SMPL
I4070-13 - 1	AU-05-070517-E	2241	0.6561	-7471A	PPB		07 Jul 2017 13:36:27	U	SMPL
I4072-01 - 1	TR-05-070517-A	3293	0.9305	-7471A	PPB		07 Jul 2017 13:38:43	U	SMPL
I4072-04 - 1	TR-05-070517-B	1124	0.3646	-7471A	PPB		07 Jul 2017 13:41:01	U	SMPL
I4074-01 - 1	TOPSOIL-01	2356	0.6861	-7471A	PPB		07 Jul 2017 13:43:18	U	SMPL
I4075-01 - 1	TP8-MH2104-COMP	2162	0.6355	-7471A	PPB		07 Jul 2017 13:45:36	U	SMPL
I4075-06 - 1	TP7-MH2103-COMP	1351	0.4239	-7471A	PPB		07 Jul 2017 13:47:53	U	SMPL
CCV64 - 1	CCV64	18412	4.8750	-7471A	PPB		07 Jul 2017 13:50:10	U	SMPL
CCB64 - 1	CCB64	-103	0.0445	-7471A	PPB		07 Jul 2017 13:52:26	U	SMPL
I4083-01 - 1	STOCKPILE	7183	1.9454	-7471A	PPB		07 Jul 2017 13:54:43	U	SMPL
I4092-01 - 1	FK-PS-01-010	679	0.2485	-7471A	PPB		07 Jul 2017 13:56:59	U	SMPL
I4095-01 - 1	FK-PS-01-011	443	0.1870	-7471A	PPB		07 Jul 2017 13:59:16	U	SMPL
I4063-01DLX10 - 1		60783	15.9293	-7471A	PPB		07 Jul 2017 14:01:33	U	SMPL
I4064-01LX5 - 1		-209	0.0169	-7471A	PPB		07 Jul 2017 14:03:49	U	SMPL
I4064-01A - 1		15645	4.1531	-7471A	PPB		07 Jul 2017 14:06:08	U	SMPL
I4069-03LX5 - 1		88	0.0944	-7471A	PPB		07 Jul 2017 14:08:26	U	SMPL
I4069-03A - 1		16623	4.4082	-7471A	PPB		07 Jul 2017 14:10:43	U	SMPL
I4063-01DL2X50 - 1		12370	3.2987	-7471A	PPB		07 Jul 2017 14:13:01	U	SMPL
I4069-06DLX2 - 1		20916	5.5283	-7471A	PPB		07 Jul 2017 14:15:18	U	SMPL
CCV65 - 1	CCV65	18278	4.8400	-7471A	PPB		07 Jul 2017 14:17:35	U	SMPL
CCB65 - 1	CCB65	-112	0.0422	-7471A	PPB		07 Jul 2017 14:19:53	U	SMPL