

Daily Analysis Runlog For Sequence/QC Batch ID # LB88582

Review By

SubDirectory

Review On

Test : Anions

LB88582

STD. NAME		STD REF.#				
ICAL Standard		WP55544,WP55545,WP55546,WP55547,WP55548,WP55549,WP55550				
ICV Standard		WP55551				
CCV Standard		WP56212				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		WP56213				
Chk Standard		WP55553,WP55554				

Sr#	SampleID	ClientID	QcType	Date	Comment	Status
1	STD1	STD1	CAL1	06/12/17 11:03		OK
2	STD2	STD2	CAL2	06/12/17 11:32		OK
3	STD3	STD3	CAL3	06/12/17 12:01		OK
4	STD4	STD4	CAL4	06/12/17 12:29		OK
5	STD5	STD5	CAL5	06/12/17 12:58		OK
6	STD6	STD6	CAL6	06/12/17 13:27		OK
7	STD7	STD7	CAL7	06/12/17 13:56		OK
8	ICV1	ICV1	ICV	06/12/17 16:02		OK
9	ICB1	ICB1	ICB	06/12/17 16:31		OK
10	CCV1	CCV1	CCV	07/06/17 12:09		OK
11	CCB1	CCB1	CCB	07/06/17 13:16		OK
12	LB88582BLW	LB88582BLW	MB	07/06/17 13:45		OK
13	LB88582BSW	LB88582BSW	LCS	07/06/17 14:14		OK
14	I4058-07	7-BOD5-NITRATE	SAM	07/06/17 14:42		OK
15	I4058-07DUP	7-BOD5-NITRATEDU	DUP	07/06/17 15:11		OK
16	I4058-07MS	7-BOD5-NITRATES	MS	07/06/17 15:40		OK
17	CCV2	CCV2	CCV	07/06/17 16:09		OK
18	CCB2	CCB2	CCB	07/06/17 16:38		OK



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	7/5/2017 12:00:00 AM
Project:	LB88582	Date Received:	7/5/2017 12:00:00 AM
Client Sample ID:	7-BOD5-NITRATE	SDG No.:	LB88582
Lab Sample ID:	I4058-07	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
	Bromide	0.69		1	0.066		0.5	mg/L		07/06/2017	300.0
	Chloride	669	OR	1	0.075		0.15	mg/L		07/06/2017	300.0
	Fluoride	0.54		1	0.043		0.1	mg/L		07/06/2017	300.0
	Nitrate	0.39		1	0.027		0.13	mg/L		07/06/2017	300.0
	Nitrite	0	U	1	0.022		0.15	mg/L		07/06/2017	300.0
	Phosphorus, Ortho	6.9		1	0.01		0.25	mg/L		07/06/2017	300.0
	Sulfate	84.6	OR	1	0.13		0.75	mg/L		07/06/2017	300.0



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	7/5/2017 12:00:00 AM
Project:	LB88582	Date Received:	7/5/2017 12:00:00 AM
Client Sample ID:	7-BOD5-NITRATE	SDG No.:	LB88582
Lab Sample ID:	I4058-07	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
-----	-----------	-------	------	----	-----	-----	------------	-------	-----------	-----------	----------

Color Before:

Clarity Before:

Texture:

Color After:

Clarity After:

Artifacts:

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB88582

SDG No.: LB88582

Contract: _____ Lab Code: CHEM

Case No.: LB88582

SAS No.: LB88582

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	% RSD	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV1	Bromide	9.8	10	98	0	90 - 110		06/12/2017	16:02	LB88582
	Chloride	2.9	3	96.7	0	90 - 110		06/12/2017	16:02	LB88582
	Fluoride	2.1	2	105	0	90 - 110		06/12/2017	16:02	LB88582
	Phosphorus, Ortho	5.1	5	102	0	90 - 110		06/12/2017	16:02	LB88582
	Nitrite	2.9	3	96.7	0	90 - 110		06/12/2017	16:02	LB88582
	Nitrate	2.4	2.5	96	0	90 - 110		06/12/2017	16:02	LB88582
	Sulfate	14.5	15	96.7	0	90 - 110		06/12/2017	16:02	LB88582

Sample ID	Analyte	Result	True Value	%	%	Acceptance	M	Analysis	Analysis	Run
		mg/L								
CCV1	Bromide	10.3	10	103	0	90 - 110		07/06/2017	12:09	LB88582
	Chloride	2.9	3	96.7	0	90 - 110		07/06/2017	12:09	LB88582
	Fluoride	2.2	2	110	0	90 - 110		07/06/2017	12:09	LB88582
	Phosphorus, Ortho	5.9	5	118	0	90 - 110		07/06/2017	12:09	LB88582
	Nitrite	2.9	3	96.7	0	90 - 110		07/06/2017	12:09	LB88582
	Nitrate	2.5	2.5	100	0	90 - 110		07/06/2017	12:09	LB88582
	Sulfate	14.8	15	98.7	0	90 - 110		07/06/2017	12:09	LB88582
CCV2	Bromide	10.3	10	103	0	90 - 110		07/06/2017	16:09	LB88582
	Chloride	3	3	100	0	90 - 110		07/06/2017	16:09	LB88582
	Fluoride	2	2	100	0	90 - 110		07/06/2017	16:09	LB88582
	Phosphorus, Ortho	4.6	5	92	0	90 - 110		07/06/2017	16:09	LB88582
	Nitrite	3	3	100	0	90 - 110		07/06/2017	16:09	LB88582
	Nitrate	2.5	2.5	100	0	90 - 110		07/06/2017	16:09	LB88582
	Sulfate	14.9	15	99.3	0	90 - 110		07/06/2017	16:09	LB88582



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB88582

SDG No.: LB88582

Contract:

Lab Code: CHEM

Case No.: LB88582

SAS No.: LB88582

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	Bromide	0.066	+/-0.5	U		0.5		06/12/2017	16:31	LB88582
	Chloride	0.075	+/-0.15	U		0.15		06/12/2017	16:31	LB88582
	Fluoride	0.043	+/-0.1	U		0.1		06/12/2017	16:31	LB88582
	Phosphorus, Ortho	0.01	+/-0.25	U		0.25		06/12/2017	16:31	LB88582
	Nitrite	0.022	+/-0.15	U		0.15		06/12/2017	16:31	LB88582
	Nitrate	0.027	+/-0.13	U		0.13		06/12/2017	16:31	LB88582
	Sulfate	0.13	+/-0.75	U		0.75		06/12/2017	16:31	LB88582



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB88582

SDG No.: LB88582

Contract:

Lab Code: CHEM

Case No.: LB88582

SAS No.: LB88582

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	Bromide	0.066	+/-0.5	U		0.5		07/06/2017	13:16	LB88582
	Chloride	0.075	+/-0.15	U		0.15		07/06/2017	13:16	LB88582
	Fluoride	0.043	+/-0.1	U		0.1		07/06/2017	13:16	LB88582
	Phosphorus, Ortho	0.01	+/-0.25	U		0.25		07/06/2017	13:16	LB88582
	Nitrite	0.022	+/-0.15	U		0.15		07/06/2017	13:16	LB88582
	Nitrate	0.027	+/-0.13	U		0.13		07/06/2017	13:16	LB88582
	Sulfate	0.13	+/-0.75	U		0.75		07/06/2017	13:16	LB88582
CCB2	Bromide	0.066	+/-0.5	U		0.5		07/06/2017	16:38	LB88582
	Chloride	0.075	+/-0.15	U		0.15		07/06/2017	16:38	LB88582
	Fluoride	0.043	+/-0.1	U		0.1		07/06/2017	16:38	LB88582
	Phosphorus, Ortho	0.01	+/-0.25	U		0.25		07/06/2017	16:38	LB88582
	Nitrite	0.022	+/-0.15	U		0.15		07/06/2017	16:38	LB88582
	Nitrate	0.027	+/-0.13	U		0.13		07/06/2017	16:38	LB88582
	Sulfate	0.13	+/-0.75	U		0.75		07/06/2017	16:38	LB88582



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB88582

SDG No.: LB88582

Contract: Lab Code: CHEM

Case No.: LB88582

SAS No.: LB88582

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	-----	------	---	------------------	------------------	---------------



GENCHEM
- 3b -
PREPARATION BLANK SUMMARY

Client: LB88582

SDG No.: LB88582

Instrument: IC-1

Sample ID	Analyte	Result (mg/L)	Acceptance Limit	Conc Qual	LOD mg/L	CRQL mg/L	M	Analysis Date	Analysis Time	Run
LB88582BLW		WATER		Batch Number:				Prep Date:		
	Bromide	0.066	<0.5	U		0.5		07/06/2017	13:45	LB88582
	Chloride	0.075	<0.15	U		0.15		07/06/2017	13:45	LB88582
	Fluoride	0.043	<0.1	U		0.1		07/06/2017	13:45	LB88582
	Phosphorus, Ortho	0.01	<0.25	U		0.25		07/06/2017	13:45	LB88582
	Nitrite	0.022	<0.15	U		0.15		07/06/2017	13:45	LB88582
	Nitrate	0.027	<0.13	U		0.13		07/06/2017	13:45	LB88582
	Sulfate	0.13	<0.75	U		0.75		07/06/2017	13:45	LB88582



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB88582 level: low sdg no.: LB88582
contract: _____ lab code: CHEM case no.: LB88582 sas no.: LB88582
matrix: WATER sample id: I4058-07 client id: 7-BOD5-NITRATEMS
Percent Solids for Sample: 0 Spiked ID: I4058-07MS Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Bromide	mg/L	80 - 120	11.1		0.69		10	104		
Chloride	mg/L	80 - 120	595	OR	669	OR	3	-2467	*	
Fluoride	mg/L	80 - 120	2.1		0.54		2	78	*	
Phosphorus, Ortho	mg/L	80 - 120	8.7		6.9		5	36	*	
Nitrite	mg/L	80 - 120	0	U	0.022	U	3			
Nitrate	mg/L	80 - 120	2.9		0.39		2.5	100		
Sulfate	mg/L	80 - 120	97.3	OR	84.6	OR	15	85		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB88582 **Level:** LOW **SDG No.:** LB88582
Contract: _____ **Lab Code:** CHEM **Case No.:** LB88582 **SAS No.:** LB88582
Matrix: WATER **Sample ID:** I4058-07 **Client ID:** 7-BOD5-NITRATEDUP
Percent Solids for Sample: 0 **Duplicate ID** I4058-07DUP **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Bromide	mg/L	20	0.69		0.91		28	*	
Chloride	mg/L	20	669	OR	682	OR	2		
Fluoride	mg/L	20	0.54		0.56		4		
Phosphorus, Ortho	mg/L	20	6.9		4.6		40	*	
Nitrite	mg/L	20	0.022	U	0	U			
Nitrate	mg/L	20	0.39		0.4		3		
Sulfate	mg/L	20	84.6	OR	84.9	OR	0		



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB88582

SDG No.: LB88582

Contract:

Lab Code: CHEM

Case No.: LB88582

SAS No.: LB88582

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
LB88582BSW							
Bromide	mg/L	10	10.2		102	90 - 110	
Chloride	mg/L	3	3		100	90 - 110	
Fluoride	mg/L	2	2.2		110	90 - 110	
Phosphorus, Ortho	mg/L	5	6.2		124	90 - 110	
Nitrite	mg/L	3	2.9		96.7	90 - 110	
Nitrate	mg/L	2.5	2.4		96	90 - 110	
Sulfate	mg/L	15	14.8		98.7	90 - 110	
