

Chemtech Consulting Group

Analytical Review Report

Date Printed : 7/10/17

Analyst : AK

Data File : lb88599.csv

Approved By :

Approved Date :

Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter			PPB			%Rec	Method	Method			Line 2
Chloride											
STD1	STD1				W	6/12/17					
Chloride		PASS	0.000		0.000						mg/L
STD2	STD2				W	6/12/17					
Chloride		PASS	0.202		0.202						mg/L
STD3	STD3				W	6/12/17					
Chloride		PASS	0.612		0.612						mg/L
STD4	STD4				W	6/12/17					
Chloride		PASS	1.149		1.15						mg/L
STD5	STD5				W	6/12/17					
Chloride		PASS	2.994		2.99						mg/L
STD6	STD6				W	6/12/17					
Chloride		PASS	5.930		5.93						mg/L
STD7	STD7				W	6/12/17					
Chloride		PASS	7.563		7.56						mg/L
ICV1	ICV1				W	6/12/17					
Chloride		PASS	2.949		2.95	98.0	90	110			mg/L
ICB1	ICB1				W	6/12/17					
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L
CCV1	CCV1				W	7/7/17					
Chloride		PASS	2.972		2.97	99.0	90	110			mg/L
CCB1	CCB1				W	7/7/17					
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L
LB88599BLS	LB88599BLS				S	7/7/17					
Chloride		PASS	0.000		0.000			+/-0.0320			mg/Kg
LB88599BSS	LB88599BSS				S	7/7/17					
Chloride		PASS	3.017		60.20	100.0	90.00	110.00			mg/Kg
I4118-01	NAA-PIPE-SAND			1	S	7/7/17					
Chloride		PASS	0.242		5.06						mg/Kg
I4118-01D	NAA-PIPE-SANDD			1	S	7/7/17					
Chloride		PASS	0.248		5.20				2.7	20	mg/Kg
I4118-01S	NAA-PIPE-SANDS			1	S	7/7/17					
Chloride		PASS	3.194		67.1	98.0	80	120			mg/Kg
I4118-01SD	NAA-PIPE-SANDSD			1	S	7/7/17					
Chloride		PASS	3.081		64.6	95.0	80	120	3.8	15	mg/Kg
CCV2	CCV2				W	7/7/17					
Chloride		PASS	2.971		2.97	99.0	90	110			mg/L
CCB2	CCB2				W	7/7/17					
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L