


245.1

L1388679
INSTRUMENT ID: CV2

Linear

 μ Abs.:

29552



A= 0.0000e+000

B= 3.4132e-004

C= 1.8062e-002

Rho= 0.9994670

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	-81	0.000	-81				
0.2	0.200	0.240	0.040	651	0.0 %	651				
2.5	2.500	2.640	0.140	7682	0.0 %	7682				
5.0	5.000	4.771	-0.229	13924	0.0 %	13924				
7.5	7.500	7.454	-0.046	21785	0.0 %	21785				
10.0	10.000	10.105	0.105	29552	0.0 %	29552				

LB88679 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 13 Jul 2017 08:25:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	58	-81	-	0.0000	245.1	PPB	13 Jul 2017 08:33:27	S	Std
0.2 - 1	50.2	651	-	0.2000	245.1	PPB	13 Jul 2017 08:35:35	S	Std
2.5 - 1	52.5	7682	-	2.5000	245.1	PPB	13 Jul 2017 08:37:44	S	Std
5.0 - 1	55	13924	-	5.0000	245.1	PPB	13 Jul 2017 08:39:52	S	Std
7.5 - 1	57.5	21785	-	7.5000	245.1	PPB	13 Jul 2017 08:41:59	S	Std
10.0 - 1	510	29552	-	10.0000	245.1	PPB	13 Jul 2017 08:46:49	S	Std
ICV59 - 1	ICV59	11512	3.9474	-245.1		PPB	13 Jul 2017 08:55:10	U	SMPL
ICB59 - 1	ICB59	-107	-0.0185	-245.1		PPB	13 Jul 2017 08:57:16	U	SMPL
CCV05 - 1	CCV05	13971	4.7867	-245.1		PPB	13 Jul 2017 08:59:24	U	SMPL
CCB05 - 1	CCB05	-256	-0.0693	-245.1		PPB	13 Jul 2017 09:01:31	U	SMPL
CRA - 1	CRA	530	0.1990	-245.1		PPB	13 Jul 2017 09:06:04	U	SMPL
HighStd - 1	HighStd	28985	9.9113	-245.1		PPB	13 Jul 2017 09:08:11	U	SMPL
ChkStd - 1	ChkStd	20387	6.9766	-245.1		PPB	13 Jul 2017 09:10:18	U	SMPL
PB100602BL - 1	PBW	-239	-0.0635	-245.1		PPB	13 Jul 2017 09:12:27	U	SMPL
PB100602BS - 1	LCSW	11664	3.9992	-245.1		PPB	13 Jul 2017 09:14:36	U	SMPL
I4100-04 - 1	EFFLUENT-4-7-5	-120	-0.0229	-245.1		PPB	13 Jul 2017 09:16:44	U	SMPL
I4100-04DUP - 1	EFFLUENT-4-7-5DUP	-21	0.0109	-245.1		PPB	13 Jul 2017 09:18:52	U	SMPL
I4100-04MS - 1	EFFLUENT-4-7-5MS	13427	4.6010	-245.1		PPB	13 Jul 2017 09:21:00	U	SMPL
I4100-04MSD - 1	EFFLUENT-4-7-5MSD	14268	4.8880	-245.1		PPB	13 Jul 2017 09:23:09	U	SMPL
CCV06 - 1	CCV06	14205	4.8665	-245.1		PPB	13 Jul 2017 09:25:16	U	SMPL
CCB06 - 1	CCB06	-169	-0.0396	-245.1		PPB	13 Jul 2017 09:27:24	U	SMPL
I4100-04LX5 - 1		-60	-0.0024	-245.1		PPB	13 Jul 2017 09:29:31	U	SMPL
I4100-04A - 1		11241	3.8549	-245.1		PPB	13 Jul 2017 09:34:08	U	SMPL
CCV07 - 1	CCV07	13867	4.7512	-245.1		PPB	13 Jul 2017 09:36:14	U	SMPL
CCB07 - 1	CCB07	-184	-0.0447	-245.1		PPB	13 Jul 2017 09:38:22	U	SMPL