

CB88696

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

7/13/2017 15:10

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.247	0.0	0.098	
ICB1	1.407	0.0	0.004	
CCV1	234.772	0.0	0.241	
CCB1	1.100	0.0	0.004	
PB100583BL	0.863	0.0	0.004	
PB100583BS	94.292	0.0	0.098	
MIDPB100583	238.345	0.0	0.244	
I4141-01	0.800	0.0	0.004	
I4141-01DUP	0.993	0.0	0.004	
I4141-01MS	32.822	0.0	0.036	
I4141-01MSD	33.102	0.0	0.036	
CCV2	246.620	0.0	0.253	
CCB2	0.939	0.0	0.004	

N	13
Mean	75.408
SD	99.4897
CV%	131.94

Aquakem v. 7.2AQ1

AK

Results from time period:

Thu Jul 13 15:04:09 2017

Thu Jul 13 15:08:02 2017

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	1.4474	µg/l	7/13/2017 11:40:12	
5.OPPBCN	A	Total CN	P	5.837	µg/l	7/13/2017 11:40:13	
10PPBCN	A	Total CN	P	10.2738	µg/l	7/13/2017 11:40:14	
50PPBCN	A	Total CN	P	48.7851	µg/l	7/13/2017 11:40:15	
100PPBCN	A	Total CN	P	98.2155	µg/l	7/13/2017 11:40:16	
250PPBCN	A	Total CN	P	249.9534	µg/l	7/13/2017 11:40:17	
500PPBCN	A	Total CN	P	500.4879	µg/l	7/13/2017 11:40:18	
ICV1	S	Total CN	P	94.2473	µg/l	7/13/2017 15:04:09	
ICB1	S	Total CN	P	1.4073	µg/l	7/13/2017 15:04:10	
CCV1	S	Total CN	P	234.7723	µg/l	7/13/2017 15:04:11	
CCB1	S	Total CN	P	1.0998	µg/l	7/13/2017 15:04:12	
PB100583BL	S	Total CN	P	0.863	µg/l	7/13/2017 15:04:13	
PB100583BS	S	Total CN	P	94.2915	µg/l	7/13/2017 15:04:14	
MIDPB100583	S	Total CN	P	238.3452	µg/l	7/13/2017 15:04:15	
I4141-01	S	Total CN	P	0.7996	µg/l	7/13/2017 15:04:16	
I4141-01DUP	S	Total CN	P	0.9932	µg/l	7/13/2017 15:04:17	
I4141-01MS	S	Total CN	P	32.8223	µg/l	7/13/2017 15:04:18	
I4141-01MSD	S	Total CN	P	33.102	µg/l	7/13/2017 15:04:19	
CCV2	S	Total CN	P	246.6201	µg/l	7/13/2017 15:08:00	
CCB2	S	Total CN	P	0.9385	µg/l	7/13/2017 15:08:01	

LB 88696

=====

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092

Reviewed by : _____

7/13/2017 11:42

Test Total CN


Accepted 7/13/2017 11:42

Factor 986.6

Bias 0.003

Coeff. of det. 0.999962

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	1.4474	0.0000	
2	5.0PPBCN	0.009	5.8370	5.0000	
3	10PPBCN	0.013	10.2738	10.0000	
4	50PPBCN	0.052	48.7851	50.0000	
5	100PPBCN	0.102	98.2155	100.0000	
6	250PPBCN	0.256	249.9534	250.0000	
7	500PPBCN	0.510	500.4879	500.0000	