

LB88714

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

7/14/2017 10:46

Test: Phenolics-

Sample Id	Result ^{mg/L}	Dil. 1 +	Response	Errors
ICV1	0.973	0.0	0.103	
ICB1	-0.003	0.0	0.002	
CCV1	0.981	0.0	0.104	
CCB1	0.001	0.0	0.002	
PB100632BL	0.000	0.0	0.002	
PB100632BS	0.984	0.0	0.104	
I4185-01	-0.001	0.0	0.002	
I4185-01DUP	-0.001	0.0	0.002	
I4185-01MS	0.922	0.0	0.098	
I4185-01MSD	0.922	0.0	0.098	
CCV2	0.982	0.0	0.104	
CCB2	-0.000	0.0	0.002	

N	12
Mean	0.480
SD	0.5024
CV%	104.65

Aquakem v. 7.2AQ1

Results from time period:

Fri Jul 14 10:18:11 2017

Fri Jul 14 10:44:52 2017

AIL

Sample Id	Sam/Ctr/c/	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0524	mg/l	7/14/2017 10:18:11	
0.1PPM	A	Phenolics-	P	0.1008	mg/l	7/14/2017 10:18:12	
0.25PPM	A	Phenolics-	P	0.2519	mg/l	7/14/2017 10:18:13	
0.50PPM	A	Phenolics-	P	0.497	mg/l	7/14/2017 10:18:14	
1.0PPM	A	Phenolics-	P	0.9949	mg/l	7/14/2017 10:18:15	
2.0PPM	A	Phenolics-	P	2.003	mg/l	7/14/2017 10:18:16	
ICV1	S	Phenolics-	P	0.9725	mg/l	7/14/2017 10:44:41	
ICB1	S	Phenolics-	P	-0.0034	mg/l	7/14/2017 10:44:42	
CCV1	S	Phenolics-	P	0.9811	mg/l	7/14/2017 10:44:43	
CCB1	S	Phenolics-	P	0.0014	mg/l	7/14/2017 10:44:44	
PB100632BL	S	Phenolics-	P	0.0004	mg/l	7/14/2017 10:44:45	
PB100632BS	S	Phenolics-	P	0.9836	mg/l	7/14/2017 10:44:46	
I4185-01	S	Phenolics-	P	-0.0008	mg/l	7/14/2017 10:44:47	
I4185-01DUP	S	Phenolics-	P	-0.0006	mg/l	7/14/2017 10:44:48	
I4185-01MS	S	Phenolics-	P	0.9224	mg/l	7/14/2017 10:44:49	
I4185-01MSD	S	Phenolics-	P	0.9223	mg/l	7/14/2017 10:44:50	
CCV2	S	Phenolics-	P	0.9825	mg/l	7/14/2017 10:44:51	
CCB2	S	Phenolics-	P	-0.0001	mg/l	7/14/2017 10:44:52	

LB88714

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

7/14/2017 10:20


Test Phenolics-

Accepted 7/14/2017 10:20

Factor 9.622
Bias 0.002

Coeff. of det. 0.999981

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0524	0.0500	
2	Phenol-2	0.013	0.1008	0.1000	
3	Phenol-2	0.028	0.2519	0.2500	
4	Phenol-2	0.054	0.4970	0.5000	
5	Phenol-2	0.106	0.9949	1.0000	
6	Phenol-2	0.210	2.0030	2.0000	