

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AK

7/24/2017 8:46

Test CNEPA-NEW


Instrument ID: CW

Accepted 7/24/2017 8:46

Factor (Slope) 925.7
 Bias (Y-intercept) 0.006

Coeff. of det. 0.999969

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors	Percentage
1	0.0PPBCN5.0	0.008	1.8499	0.0000		
2	5.0PPBCN55.0	0.011	4.8473	5.0000		96.94%
3	10PPBCN10.0	0.017	9.6278	10.0000		96.27%
4	100PPBCN100.0	0.113	98.4392	100.0000		98.43%
5	250PPBCN250.0	0.276	249.8293	250.0000		99.93%
6	500PPBCN550.0	0.547	500.4065	500.0000		100%

AK
 7-24-17

UB88934

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

7/24/2017 9:40

Test: CNEPA-NEW

Sample Id	Result	Dil. 1 +	Response	Errors
ICV ICV	92.079	0.0	0.106	
ICB ICB	1.931	0.0	0.008	
CCV001 CCV001	238.353	0.0	0.264	
CCB001 CCB001	2.219	0.0	0.009	
PB100809BL PBW005	1.989	0.0	0.008	
I4259-01 MC0AC7	1.540	0.0	0.008	
I4259-02 MC0AC8	0.226	0.0	0.006	
I4259-03 MC0AC9	-0.029	0.0	0.006	
I4259-04 MC0AC9D	0.118	0.0	0.006	
I4259-06 MC0AD1	0.202	0.0	0.006	
I4259-07 MC0AD2	-0.027	0.0	0.006	
I4259-08 MC0AD3	1.958	0.0	0.008	
I4259-05 MC0AC9S	79.462	0.0	0.092	
CCV002 CCV002	255.742	0.0	0.283	
CCB002 CCB002	1.830	0.0	0.008	

N 15
Mean 45.173
SD 87.1870
CV% 193.01

AK
7-24-18

Aquakem v. 7.2AQ1

Results from time period:

Mon Jul 24 08:43:44 2017

Mon Jul 24 09:38:07 2017

LB 88934

AK

Sample Id	Sam/Ctr/c/	Test short n	Test type	Result	Result unit	Result date and time
S0.0	A	CNEPA-NEW P		1.8499	µg/l	7/24/2017 8:43:44
S5.0	A	CNEPA-NEW P		4.8473	µg/l	7/24/2017 8:43:45
S10.0	A	CNEPA-NEW P		9.6278	µg/l	7/24/2017 8:43:46
S100.0	A	CNEPA-NEW P		98.4392	µg/l	7/24/2017 8:43:47
S250.0	A	CNEPA-NEW P		249.8293	µg/l	7/24/2017 8:43:48
S500.0	A	CNEPA-NEW P		500.4065	µg/l	7/24/2017 8:43:49
ICV ICV	S	CNEPA-NEW P		92.0791	µg/l	7/24/2017 9:18:31
ICB ICB	S	CNEPA-NEW P		1.9308	µg/l	7/24/2017 9:18:32
CCV001 CCV001	S	CNEPA-NEW P		238.3527	µg/l	7/24/2017 9:18:33
CCB001 CCB001	S	CNEPA-NEW P		2.2191	µg/l	7/24/2017 9:18:34
PB100809BL PBW005	S	CNEPA-NEW P		1.9891	µg/l	7/24/2017 9:18:35
I4259-01 MC0AC7	S	CNEPA-NEW P		1.5399	µg/l	7/24/2017 9:18:36
I4259-02 MC0AC8	S	CNEPA-NEW P		0.2256	µg/l	7/24/2017 9:18:37
I4259-03 MC0AC9	S	CNEPA-NEW P		-0.0286	µg/l	7/24/2017 9:18:38
I4259-04 MC0AC9D	S	CNEPA-NEW P		0.1184	µg/l	7/24/2017 9:18:39
I4259-06 MC0AD1	S	CNEPA-NEW P		0.2019	µg/l	7/24/2017 9:18:41
I4259-07 MC0AD2	S	CNEPA-NEW P		-0.0267	µg/l	7/24/2017 9:23:18
I4259-08 MC0AD3	S	CNEPA-NEW P		1.9582	µg/l	7/24/2017 9:23:19
I4259-05 MC0AC9S	S	CNEPA-NEW P		79.4617	µg/l	7/24/2017 9:36:15
CCV002 CCV002	S	CNEPA-NEW P		255.7417	µg/l	7/24/2017 9:36:16
CCB002 CCB002	S	CNEPA-NEW P		1.8297	µg/l	7/24/2017 9:38:07