

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AE

7/28/2017 15:11

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.022	0.0	0.113	
ICB1	0.007	0.0	0.004	
CCV1	0.998	0.0	0.111	
CCB1	0.007	0.0	0.004	
PB101032BL	0.006	0.0	0.004	
PB101032BS	1.030	0.0	0.114	
I4346-16	1.724	0.0	0.189	
CCV2	1.002	0.0	0.111	
CCB2	0.003	0.0	0.003	

N	9
Mean	0.644
SD	0.6463
CV%	100.30

Aquakem v. 7.2AQ1

Results from time period:

Fri Jul 28 15:07:14 2017

Fri Jul 28 15:07:23 2017

LB 89105

AK

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.054	mg/l	7/28/2017 14:03:59	
0.1PPM	A	Phenolics-	P	0.1086	mg/l	7/28/2017 14:04:00	
0.25PPM	A	Phenolics-	P	0.2564	mg/l	7/28/2017 14:04:01	
0.50PPM	A	Phenolics-	P	0.5021	mg/l	7/28/2017 14:04:02	
1.0PPM	A	Phenolics-	P	0.9616	mg/l	7/28/2017 14:04:03	
2.0PPM	A	Phenolics-	P	2.0173	mg/l	7/28/2017 14:04:04	
ICV1	S	Phenolics-	P	1.0219	mg/l	7/28/2017 15:07:14	
ICB1	S	Phenolics-	P	0.0069	mg/l	7/28/2017 15:07:15	
CCV1	S	Phenolics-	P	0.9983	mg/l	7/28/2017 15:07:16	
CCB1	S	Phenolics-	P	0.0068	mg/l	7/28/2017 15:07:17	
PB101032BL	S	Phenolics-	P	0.0059	mg/l	7/28/2017 15:07:18	
PB101032BS	S	Phenolics-	P	1.03	mg/l	7/28/2017 15:07:19	
I4346-16	S	Phenolics-	P	1.7238	mg/l	7/28/2017 15:07:20	
CCV2	S	Phenolics-	P	1.0024	mg/l	7/28/2017 15:07:22	
CCB2	S	Phenolics-	P	0.0033	mg/l	7/28/2017 15:07:23	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

7/28/2017 14:05


Test Phenolics-

Accepted 7/28/2017 14:05

Factor 9.252
Bias 0.003

Coeff. of det. 0.999317

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0540	0.0500	
2	Phenol-2	0.015	0.1086	0.1000	
3	Phenol-2	0.031	0.2564	0.2500	
4	Phenol-2	0.057	0.5021	0.5000	
5	Phenol-2	0.107	0.9616	1.0000	
6	Phenol-2	0.221	2.0173	2.0000	