

UB89153

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092

Reviewed by : AK

7/31/2017 14:35

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.983	0.0	0.108	
ICB1	0.003	0.0	0.002	
CCV1	0.998	0.0	0.109	
CCB1	0.003	0.0	0.003	
PB101063BL	0.004	0.0	0.003	
PB101063BS	0.986	0.0	0.108	
I4508-01	0.029	0.0	0.005	
I4508-01DUP	0.029	0.0	0.005	
I4508-01MS	0.873	0.0	0.096	
I4508-01MSD	0.863	0.0	0.095	
CCV2	1.011	0.0	0.111	
CCB2	0.002	0.0	0.002	

N	12
Mean	0.482
SD	0.4933
CV%	102.34

CB 8915 3

Aquakem v. 7.2AQ1

Results from time period:

Mon Jul 31 13:24:34 2017

Mon Jul 31 14:34:56 2017

AK

Sample Id	Sam/Ctr/c/	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.051	mg/l	7/31/2017 13:43:56	
0.1PPM	A	Phenolics-	P	0.1046	mg/l	7/31/2017 13:43:57	
0.25PPM	A	Phenolics-	P	0.251	mg/l	7/31/2017 13:43:58	
0.50PPM	A	Phenolics-	P	0.5053	mg/l	7/31/2017 13:43:59	
1.0PPM	A	Phenolics-	P	0.9798	mg/l	7/31/2017 13:44:00	
2.0PPM	A	Phenolics-	P	2.0084	mg/l	7/31/2017 13:44:01	
ICV1	S	Phenolics-	P	0.9825	mg/l	7/31/2017 14:32:08	
ICB1	S	Phenolics-	P	0.0028	mg/l	7/31/2017 14:32:09	
CCV1	S	Phenolics-	P	0.9975	mg/l	7/31/2017 14:32:10	
CCB1	S	Phenolics-	P	0.0035	mg/l	7/31/2017 14:32:11	
PB101063BL	S	Phenolics-	P	0.0036	mg/l	7/31/2017 14:32:12	
PB101063BS	S	Phenolics-	P	0.9864	mg/l	7/31/2017 14:32:13	
I4508-01	S	Phenolics-	P	0.0294	mg/l	7/31/2017 14:32:14	
I4508-01DUP	S	Phenolics-	P	0.0286	mg/l	7/31/2017 14:32:15	
I4508-01MS	S	Phenolics-	P	0.8734	mg/l	7/31/2017 14:32:16	
I4508-01MSD	S	Phenolics-	P	0.8634	mg/l	7/31/2017 14:32:17	
CCV2	S	Phenolics-	P	1.0106	mg/l	7/31/2017 14:32:18	
CCB2	S	Phenolics-	P	0.0024	mg/l	7/31/2017 14:32:19	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

7/31/2017 13:53


Test Phenolics-

Accepted 7/31/2017 13:53

Factor 9.32
Bias 0.002

Coeff. of det. 0.999811

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0510	0.0500	
2	Phenol-2	0.013	0.1046	0.1000	
3	Phenol-2	0.029	0.2510	0.2500	
4	Phenol-2	0.056	0.5053	0.5000	
5	Phenol-2	0.107	0.9798	1.0000	
6	Phenol-2	0.218	2.0084	2.0000	