

CB89154

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK

7/31/2017 15:26

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.990	0.0	0.108	
ICB1	0.002	0.0	0.002	
CCV1	1.005	0.0	0.110	
CCB1	0.002	0.0	0.002	
PB101064BL	0.001	0.0	0.002	
PB101064BS	0.992	0.0	0.109	
I4512-01	0.002	0.0	0.002	
I4512-01DUP	0.003	0.0	0.002	
I4512-01MS	0.946	0.0	0.104	
I4512-01MSD	0.926	0.0	0.102	
CCV2	1.004	0.0	0.110	
CCB2	0.004	0.0	0.003	

N	12
Mean	0.490
SD	0.5095
CV%	104.02

CB 89154

Aquakem v. 7.2AQ1

Results from time period:

Mon Jul 31 15:01:04 2017

Mon Jul 31 15:01:15 2017

AC

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.051	mg/l	7/31/2017 13:43:56	
0.1PPM	A	Phenolics-	P	0.1046	mg/l	7/31/2017 13:43:57	
0.25PPM	A	Phenolics-	P	0.251	mg/l	7/31/2017 13:43:58	
0.50PPM	A	Phenolics-	P	0.5053	mg/l	7/31/2017 13:43:59	
1.0PPM	A	Phenolics-	P	0.9798	mg/l	7/31/2017 13:44:00	
2.0PPM	A	Phenolics-	P	2.0084	mg/l	7/31/2017 13:44:01	
ICV1	S	Phenolics-	P	0.9902	mg/l	7/31/2017 15:01:04	
ICB1	S	Phenolics-	P	0.0024	mg/l	7/31/2017 15:01:05	
CCV1	S	Phenolics-	P	1.0047	mg/l	7/31/2017 15:01:06	
CCB1	S	Phenolics-	P	0.0024	mg/l	7/31/2017 15:01:07	
PB101064BL	S	Phenolics-	P	0.0011	mg/l	7/31/2017 15:01:08	
PB101064BS	S	Phenolics-	P	0.9924	mg/l	7/31/2017 15:01:09	
I4512-01	S	Phenolics-	P	0.0017	mg/l	7/31/2017 15:01:10	
I4512-01DUP	S	Phenolics-	P	0.0027	mg/l	7/31/2017 15:01:11	
I4512-01MS	S	Phenolics-	P	0.9455	mg/l	7/31/2017 15:01:12	
I4512-01MSD	S	Phenolics-	P	0.9264	mg/l	7/31/2017 15:01:13	
CCV2	S	Phenolics-	P	1.0036	mg/l	7/31/2017 15:01:14	
CCB2	S	Phenolics-	P	0.0045	mg/l	7/31/2017 15:01:15	

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AK7/31/2017 13:53


Test Phenolics-

Accepted 7/31/2017 13:53

Factor 9.32
Bias 0.002

Coeff. of det. 0.999811

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0510	0.0500	
2	Phenol-2	0.013	0.1046	0.1000	
3	Phenol-2	0.029	0.2510	0.2500	
4	Phenol-2	0.056	0.5053	0.5000	
5	Phenol-2	0.107	0.9798	1.0000	
6	Phenol-2	0.218	2.0084	2.0000	