


Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : 

8/1/2017 9:43

Test: Total CN

Sample Id	Result ^{ug/L}	Dil. 1 +	Response	Errors
ICV1	95.535	0.0	0.102	
ICB1	0.908	0.0	0.004	
CCV1	239.607	0.0	0.252	
CCB1	1.306	0.0	0.005	
PB101101BL	0.501	0.0	0.004	
PB101101BS	95.656	0.0	0.103	
LOWPB101101	10.132	0.0	0.014	
HIGHPB101101	483.805	0.0	0.504	
I4511-02	0.941	0.0	0.005	
I4511-02DUP	0.303	0.0	0.004	
I4511-02MS	26.896	0.0	0.031	
I4511-02MSD	27.381	0.0	0.032	
I4512-01	0.371	0.0	0.004	
I4524-01	2.228	0.0	0.006	
CCV2	250.530	0.0	0.263	
CCB2	0.615	0.0	0.004	

N	16
Mean	77.295
SD	136.2630
CV%	176.29

Aquakem v. 7.2AQ1

Results from time period:

Tue Aug 01 08:45:30 2017

Tue Aug 01 09:39:24 2017

D.P

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and tin	Stat
0.0PPBCN	A	Total CN	P	0.177	µg/l	8/1/2017 8:45:30	
5.0PPBCN	A	Total CN	P	5.074	µg/l	8/1/2017 8:45:31	
10PPBCN	A	Total CN	P	10.1503	µg/l	8/1/2017 8:45:32	
50PPBCN	A	Total CN	P	49.2779	µg/l	8/1/2017 8:45:33	
100PPBCN	A	Total CN	P	100.3265	µg/l	8/1/2017 8:45:34	
250PPBCN	A	Total CN	P	249.9822	µg/l	8/1/2017 8:45:35	
500PPBCN	A	Total CN	P	500.0121	µg/l	8/1/2017 8:45:36	
ICV1	S	Total CN	P	95.5348	µg/l	8/1/2017 9:33:35	
ICB1	S	Total CN	P	0.9083	µg/l	8/1/2017 9:33:36	
CCV1	S	Total CN	P	239.607	µg/l	8/1/2017 9:33:37	
CCB1	S	Total CN	P	1.3057	µg/l	8/1/2017 9:33:38	
PB101101BL	S	Total CN	P	0.5007	µg/l	8/1/2017 9:33:39	
PB101101BS	S	Total CN	P	95.6561	µg/l	8/1/2017 9:33:40	
LOWPB101101	S	Total CN	P	10.1318	µg/l	8/1/2017 9:33:41	
HIGHPB101101	S	Total CN	P	483.8045	µg/l	8/1/2017 9:33:42	
I4511-02	S	Total CN	P	0.9414	µg/l	8/1/2017 9:33:43	
I4511-02DUP	S	Total CN	P	0.3031	µg/l	8/1/2017 9:33:44	
I4511-02MS	S	Total CN	P	26.8962	µg/l	8/1/2017 9:33:45	
I4511-02MSD	S	Total CN	P	27.3815	µg/l	8/1/2017 9:39:18	
I4512-01	S	Total CN	P	0.3714	µg/l	8/1/2017 9:39:19	
I4524-01	S	Total CN	P	2.2278	µg/l	8/1/2017 9:39:20	
CCV2	S	Total CN	P	250.5302	µg/l	8/1/2017 9:39:21	
CCB2	S	Total CN	P	0.6154	µg/l	8/1/2017 9:39:22	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : DP

8/1/2017 9:05


Test Total CN

Accepted 8/1/2017 9:05

Factor 966
Bias 0.004

Coeff. of det. 0.999997

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	0.1770	0.0000	
2	5.0PPBCN	0.009	5.0740	5.0000	
3	10PPBCN	0.014	10.1503	10.0000	
4	50PPBCN	0.055	49.2779	50.0000	
5	100PPBCN	0.107	100.3265	100.0000	
6	250PPBCN	0.262	249.9822	250.0000	
7	500PPBCN	0.521	500.0121	500.0000	