

LB89357

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : AR

8/9/2017 10:38

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.997	0.0	0.110	
ICB1	0.006	0.0	0.003	
CCV1	1.007	0.0	0.111	
CCB1	0.006	0.0	0.003	
PB101363BL	0.005	0.0	0.003	
PB101363BS	0.993	0.0	0.110	
I4655-01	0.005	0.0	0.003	
I4655-01DUP	0.005	0.0	0.002	
I4655-01MS	0.863	0.0	0.096	
I4655-01MSD	0.861	0.0	0.096	
I4655-02	0.005	0.0	0.002	
I4655-03	0.008	0.0	0.003	
CCV2	1.003	0.0	0.111	
CCB2	0.003	0.0	0.002	

N	14
Mean	0.412
SD	0.4892
CV%	118.77

Aquakem v. 7.2AQ1

Results from time period:

Wed Aug 09 09:55:10 2017

Wed Aug 09 10:33:51 2017

AK

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0548	mg/l	8/9/2017 9:55:10	
0.1PPM	A	Phenolics-	P	0.1041	mg/l	8/9/2017 9:55:11	
0.25PPM	A	Phenolics-	P	0.2539	mg/l	8/9/2017 9:55:12	
0.50PPM	A	Phenolics-	P	0.4997	mg/l	8/9/2017 9:55:13	
1.0PPM	A	Phenolics-	P	0.9764	mg/l	8/9/2017 9:55:14	
2.0PPM	A	Phenolics-	P	2.011	mg/l	8/9/2017 9:55:15	
ICV1	S	Phenolics-	P	0.9971	mg/l	8/9/2017 10:28:54	
ICB1	S	Phenolics-	P	0.0056	mg/l	8/9/2017 10:28:55	
CCV1	S	Phenolics-	P	1.0068	mg/l	8/9/2017 10:28:56	
CCB1	S	Phenolics-	P	0.0058	mg/l	8/9/2017 10:28:57	
PB101363BL	S	Phenolics-	P	0.0052	mg/l	8/9/2017 10:28:58	
PB101363BS	S	Phenolics-	P	0.9932	mg/l	8/9/2017 10:28:59	
I4655-01	S	Phenolics-	P	0.0049	mg/l	8/9/2017 10:29:00	
I4655-01DUP	S	Phenolics-	P	0.0046	mg/l	8/9/2017 10:29:01	
I4655-01MS	S	Phenolics-	P	0.8627	mg/l	8/9/2017 10:29:02	
I4655-01MSD	S	Phenolics-	P	0.8614	mg/l	8/9/2017 10:29:03	
I4655-02	S	Phenolics-	P	0.0046	mg/l	8/9/2017 10:29:04	
I4655-03	S	Phenolics-	P	0.0085	mg/l	8/9/2017 10:29:05	
CCV2	S	Phenolics-	P	1.003	mg/l	8/9/2017 10:33:48	
CCB2	S	Phenolics-	P	0.0032	mg/l	8/9/2017 10:33:49	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AK

8/9/2017 9:58


Test Phenolics-

Accepted 8/9/2017 9:58

Factor 9.2
Bias 0.002

Coeff. of det. 0.999737

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0548	0.0500	
2	Phenol-2	0.013	0.1041	0.1000	
3	Phenol-2	0.030	0.2539	0.2500	
4	Phenol-2	0.056	0.4997	0.5000	
5	Phenol-2	0.108	0.9764	1.0000	
6	Phenol-2	0.221	2.0110	2.0000	