

LB89553

Result archive

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : DP

8/17/2017 9:20

Sample id	Test id	Result ^{mg/L}	Status	Date	Errors
ICV1	Phenolics-	0.975	man acc	8/16/2017	
ICB1	Phenolics-	0.012	man acc	8/16/2017	
CCV1	Phenolics-	1.016	man acc	8/16/2017	
CCB1	Phenolics-	0.011	man acc	8/16/2017	
PB101586PBL	Phenolics-	0.010	man acc	8/16/2017	
PB101586PBS	Phenolics-	0.993	man acc	8/16/2017	
I4792-11	Phenolics-	0.010	man acc	8/16/2017	
I4792-11DUP	Phenolics-	0.011	man acc	8/16/2017	
I4792-11MS	Phenolics-	0.951	man acc	8/16/2017	
I4792-11MSD	Phenolics-	0.930	man acc	8/16/2017	
CCV2	Phenolics-	1.010	man acc	8/16/2017	
CCB2	Phenolics-	0.008	man acc	8/16/2017	

LB89555

Aquakem v. 7.2AQ1

Results from time period:

Wed Aug 16 17:06:08 2017

Wed Aug 16 18:06:53 2017

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0599	mg/l	8/16/2017 17:06:08	
0.1PPM	A	Phenolics-	P	0.1068	mg/l	8/16/2017 17:06:09	
0.25PPM	A	Phenolics-	P	0.2562	mg/l	8/16/2017 17:06:10	
0.50PPM	A	Phenolics-	P	0.4961	mg/l	8/16/2017 17:06:11	
1.0PPM	A	Phenolics-	P	0.9627	mg/l	8/16/2017 17:06:12	
2.0PPM	A	Phenolics-	P	2.0183	mg/l	8/16/2017 17:06:13	
ICV1	S	Phenolics-	P	0.9755	mg/l	8/16/2017 18:06:42	
ICB1	S	Phenolics-	P	0.0118	mg/l	8/16/2017 18:06:43	
CCV1	S	Phenolics-	P	1.0157	mg/l	8/16/2017 18:06:44	
CCB1	S	Phenolics-	P	0.0113	mg/l	8/16/2017 18:06:45	
PB101586BL	S	Phenolics-	P	0.0097	mg/l	8/16/2017 18:06:46	
PB101586BS	S	Phenolics-	P	0.993	mg/l	8/16/2017 18:06:47	
I4792-11	S	Phenolics-	P	0.0103	mg/l	8/16/2017 18:06:48	
I4792-11DUP	S	Phenolics-	P	0.0106	mg/l	8/16/2017 18:06:49	
I4792-11MS	S	Phenolics-	P	0.9509	mg/l	8/16/2017 18:06:50	
I4792-11MSD	S	Phenolics-	P	0.93	mg/l	8/16/2017 18:06:51	
CCV2	S	Phenolics-	P	1.0097	mg/l	8/16/2017 18:06:52	
CCB2	S	Phenolics-	P	0.0081	mg/l	8/16/2017 18:06:53	

LB89555

=====

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : DP

8/16/2017 17:09

Test Phenolics-


Accepted 8/16/2017 17:09

Factor 8.644

Bias 0.002

Coeff. of det. 0.999311

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0599	0.0500	
2	Phenol-2	0.014	0.1068	0.1000	
3	Phenol-2	0.031	0.2562	0.2500	
4	Phenol-2	0.059	0.4961	0.5000	
5	Phenol-2	0.113	0.9627	1.0000	
6	Phenol-2	0.235	2.0183	2.0000	