

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : DP

8/25/2017 14:41

m *mg/L*

Sample id	Test id	Result	Status	Date	Errors
ICV1	Phenolics-	1.031	man acc	8/25/2017	
ICB1	Phenolics-	-0.003	man acc	8/25/2017	
CCV1	Phenolics-	1.016	man acc	8/25/2017	
CCB1	Phenolics-	-0.001	man acc	8/25/2017	
PB101770BL	Phenolics-	0.003	man acc	8/25/2017	
PB101770BS	Phenolics-	1.020	man acc	8/25/2017	
I4923-01	Phenolics-	0.002	man acc	8/25/2017	
I4923-01DUP	Phenolics-	-0.002	man acc	8/25/2017	
I4923-01MS	Phenolics-	0.935	man acc	8/25/2017	
I4923-01MSD	Phenolics-	0.979	man acc	8/25/2017	
I4923-02	Phenolics-	0.004	man acc	8/25/2017	
CCV2	Phenolics-	0.974	man acc	8/25/2017	
CCB2	Phenolics-	-0.002	man acc	8/25/2017	

Aquakem v. 7.2AQ1

Results from time period:

Fri Aug 25 13:39:39 2017

Fri Aug 25 14:28:53 2017

~~LB89718~~
LB89721

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0505	mg/l	8/25/2017 13:39:39	
0.1PPM	A	Phenolics-	P	0.098	mg/l	8/25/2017 13:39:40	
0.25PPM	A	Phenolics-	P	0.2516	mg/l	8/25/2017 13:39:41	
0.50PPM	A	Phenolics-	P	0.4916	mg/l	8/25/2017 13:39:42	
1.0PPM	A	Phenolics-	P	1.0125	mg/l	8/25/2017 13:39:43	
2.0PPM	A	Phenolics-	P	1.9957	mg/l	8/25/2017 13:39:44	
ICV1	S	Phenolics-	P	1.0313	mg/l	8/25/2017 14:22:30	
ICB1	S	Phenolics-	P	-0.0035	mg/l	8/25/2017 14:22:31	
CCV1	S	Phenolics-	P	1.0165	mg/l	8/25/2017 14:22:33	
CCB1	S	Phenolics-	P	-0.001	mg/l	8/25/2017 14:22:36	
PB101770BL	S	Phenolics-	P	0.0027	mg/l	8/25/2017 14:22:38	
PB101770BS	S	Phenolics-	P	1.0202	mg/l	8/25/2017 14:22:40	
I4923-01	S	Phenolics-	P	0.0023	mg/l	8/25/2017 14:28:47	
I4923-01DUP	S	Phenolics-	P	-0.0024	mg/l	8/25/2017 14:28:48	
I4923-01MS	S	Phenolics-	P	0.9349	mg/l	8/25/2017 14:28:49	
I4923-01MSD	S	Phenolics-	P	0.9794	mg/l	8/25/2017 14:28:50	
I4923-02	S	Phenolics-	P	0.0043	mg/l	8/25/2017 14:28:51	
CCV2	S	Phenolics-	P	0.9739	mg/l	8/25/2017 14:28:52	
CCB2	S	Phenolics-	P	-0.002	mg/l	8/25/2017 14:28:53	

=====

Calibration results

LB89721

Aquakem 7.2AQ1

LB89718

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : DP

8/25/2017 13:58


Test Phenolics-

Accepted 8/25/2017 13:58

Factor 9.201
Bias 0.003

Coeff. of det. 0.999910

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0505	0.0500	
2	Phenol-2	0.014	0.0980	0.1000	
3	Phenol-2	0.031	0.2516	0.2500	
4	Phenol-2	0.057	0.4916	0.5000	
5	Phenol-2	0.113	1.0125	1.0000	
6	Phenol-2	0.220	1.9957	2.0000	