

2B90005

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH

284 Sheffield Street,
Mountainside, NJ 07092Reviewed by : DP

9/11/2017 17:13

Test: Phenolics-

mg/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.028	0.0	0.117	
ICB1	0.003	0.0	0.003	
CCV1	1.045	0.0	0.118	
CCB1	0.002	0.0	0.003	
PB102219BL	0.002	0.0	0.003	
PB102219BS	1.012	0.0	0.115	
I5164-01	0.008	0.0	0.003	
I5164-01DUP	0.003	0.0	0.003	
I5164-01MS	0.990	0.0	0.112	
I5164-01MSD	0.973	0.0	0.111	
CCV2	1.020	0.0	0.116	
CCB2	-0.001	0.0	0.002	

N	12
Mean	0.507
SD	0.5269
CV%	103.91

LB90005

Aquakem v. 7.2AQ1

Results from time period:

Mon Sep 11 16:45:02 2017

Mon Sep 11 17:12:32 2017

Sample Id	Sam/Ctr/c/	Test short na	Test type	Result	Result unit	Result date	Stat	Accept stat
0.05PPM	A	Phenolics-	P	0.0476	mg/l	42989.7		MA
0.05PPM	A	Phenolics-	P	0.1012	mg/l	42989.7		MA
0.05PPM	A	Phenolics-	P	0.2561	mg/l	42989.7		MA
0.05PPM	A	Phenolics-	P	0.5106	mg/l	42989.7		MA
0.05PPM	A	Phenolics-	P	0.976	mg/l	42989.7		MA
0.05PPM	A	Phenolics-	P	2.0086	mg/l	42989.7		MA
ICV1	S	Phenolics-	P	1.0282	mg/l	42989.71		MA
ICB1	S	Phenolics-	P	0.0028	mg/l	42989.71		MA
CCV1	S	Phenolics-	P	1.0448	mg/l	42989.71		MA
CCB1	S	Phenolics-	P	0.0021	mg/l	42989.71		MA
PB102219BL	S	Phenolics-	P	0.002	mg/l	42989.71		MA
PB102219BS	S	Phenolics-	P	1.0119	mg/l	42989.71		MA
I5164-01	S	Phenolics-	P	0.0083	mg/l	42989.72		MA
I5164-01DUP	S	Phenolics-	P	0.0033	mg/l	42989.72		MA
I5164-01MS	S	Phenolics-	P	0.9898	mg/l	42989.72		MA
I5164-01MSD	S	Phenolics-	P	0.973	mg/l	42989.72		MA
CCV2	S	Phenolics-	P	1.02	mg/l	42989.72		MA
CCB2	S	Phenolics-	P	-0.001	mg/l	42989.72		MA

=====

Calibration results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : dp

9/11/2017 16:45


Test Phenolics-

Accepted 9/11/2017 16:45

Factor 9.014
Bias 0.003

Coeff. of det. 0.999712

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.008	0.0476	0.0500	
2	Phenol-2	0.014	0.1012	0.1000	
3	Phenol-2	0.031	0.2561	0.2500	
4	Phenol-2	0.059	0.5106	0.5000	
5	Phenol-2	0.111	0.9760	1.0000	
6	Phenol-2	0.225	2.0086	2.0000	