

Daily Analysis Runlog For Sequence/QC Batch ID # LB90030

Review By		Review On				
SubDirectory		LB90030		Test : Cyanide		
STD. NAME		STD REF.#				
ICAL Standard		WP57860,WP57861,WP57862,WP57863,WP57864,WP57865,WP57866				
ICV Standard		wp57868				
CCV Standard		wp57869				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		N/A				
Chk Standard		WP57867,WP55355,WP55836				
Sr#	SampleId	ClientID	QcType	Date	Comment	Status
1	0.0PPBCN	0.0PPBCN	CAL1	09/12/17 12:41		OK
2	5.0PPBCN	5.0PPBCN	CAL2	09/12/17 12:41		OK
3	10PPBCN	10PPBCN	CAL3	09/12/17 12:41		OK
4	50PPBCN	50PPBCN	CAL4	09/12/17 12:41		OK
5	100PPBCN	100PPBCN	CAL5	09/12/17 12:41		OK
6	250PPBCN	250PPBCN	CAL6	09/12/17 12:41		OK
7	500PPBCN	500PPBCN	CAL7	09/12/17 12:41		OK
8	ICV1	ICV1	ICV	09/12/17 16:02		OK
9	ICB1	ICB1	ICB	09/12/17 16:02		OK
10	CCV1	CCV1	CCV	09/12/17 16:02		OK
11	CCB1	CCB1	CCB	09/12/17 16:02		OK
12	PB102274BL	PB102274BL	MB	09/12/17 16:02		OK
13	PB102274BS	PB102274BS	LCS	09/12/17 16:02		OK
14	LOWPB102274	LOWPB102274	SAM	09/12/17 16:10		OK
15	HIGHPB102274	HIGHPB102274	SAM	09/12/17 16:10		OK
16	I5201-01	SU-03-091117	SAM	09/12/17 16:10		OK
17	I5201-01DUP	SU-03-091117DUP	DUP	09/12/17 16:10		OK
18	I5201-01MS	SU-03-091117MS	MS	09/12/17 16:10		OK
19	I5201-01MSD	SU-03-091117MSD	MSD	09/12/17 16:10		OK
20	I5203-01	A4129	SAM	09/12/17 16:10		OK
21	I5207-01	20170610-COMPOS	SAM	09/12/17 16:10		OK

Daily Analysis Runlog For Sequence/QC Batch ID # LB90030

Review By		Review On				
SubDirectory		LB90030		Test : Cyanide		
STD. NAME		STD REF.#				
ICAL Standard		WP57860,WP57861,WP57862,WP57863,WP57864,WP57865,WP57866				
ICV Standard		wp57868				
CCV Standard		wp57869				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		N/A				
Chk Standard		WP57867,WP55355,WP55836				
22	CCV2	CCV2	CCV	09/12/17 16:17		OK
23	CCB2	CCB2	CCB	09/12/17 16:17		OK
24	I5207-02	20170611-COMPOS	SAM	09/12/17 16:17		OK
25	I5207-03	20170530-LOC-6	SAM	09/12/17 16:17		OK
26	I5208-01	CL-02-091117-A	SAM	09/12/17 16:17		OK
27	I5208-04	CL-02-091117-B	SAM	09/12/17 16:17		OK
28	I5208-07	CL-02-091117-C	SAM	09/12/17 16:17		OK
29	I5208-10	CL-02-091117-D	SAM	09/12/17 16:17		OK
30	I5208-13	CL-02-091117-E	SAM	09/12/17 16:17		OK
31	I5209-01	OR-03-91117	SAM	09/12/17 16:17		OK
32	PB102275BL	PB102275BL	MB	09/12/17 16:17		OK
33	PB102275BS	PB102275BS	LCS	09/12/17 16:25		OK
34	CCV3	CCV3	CCV	09/12/17 16:25		OK
35	CCB3	CCB3	CCB	09/12/17 16:25		OK
36	I5202-01	A5972-OT	SAM	09/12/17 16:25		OK
37	I5202-01DUP	A5972-OTDUP	DUP	09/12/17 16:25		OK
38	I5202-01MS	A5972-OTMS	MS	09/12/17 16:25		OK
39	I5202-01MSD	A5972-OTMSD	MSD	09/12/17 16:25		OK
40	CCV4	CCV4	CCV	09/12/17 16:25		OK
41	CCB4	CCB4	CCB	09/12/17 16:25		OK



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB90030	Date Received:	
Client Sample ID:	HIGHPB102274	SDG No.:	LB90030
Lab Sample ID:	HIGHPB102274	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.49		1	0.003		0.005	mg/L		09/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/11/2017 12:00:00 AM
Project:	LB90030	Date Received:	9/11/2017 12:00:00 AM
Client Sample ID:	SU-03-091117	SDG No.:	LB90030
Lab Sample ID:	I5201-01	Matrix:	Solid
Level (low/med):	low	% Solid:	89.7

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.029	U	1	0.036		0.27	mg/Kg	09/12/2017	09/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/11/2017 12:00:00 AM
Project:	LB90030	Date Received:	9/11/2017 12:00:00 AM
Client Sample ID:	A5972-OT	SDG No.:	LB90030
Lab Sample ID:	I5202-01	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.00045	U	1	0.003		0.005	mg/L	09/12/2017	09/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/11/2017 12:00:00 AM
Project:	LB90030	Date Received:	9/11/2017 12:00:00 AM
Client Sample ID:	A4129	SDG No.:	LB90030
Lab Sample ID:	I5203-01	Matrix:	Solid
Level (low/med):	low	% Solid:	18.6

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.14	U	1	0.17		1.3	mg/Kg	09/12/2017	09/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/5/2017 12:00:00 AM
Project:	LB90030	Date Received:	9/11/2017 12:00:00 AM
Client Sample ID:	20170610-COMPOSITE-A	SDG No.:	LB90030
Lab Sample ID:	I5207-01	Matrix:	Solid
Level (low/med):	low	% Solid:	57.9

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.07	U	1	0.054		0.41	mg/Kg	09/12/2017	09/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/5/2017 12:00:00 AM
Project:	LB90030	Date Received:	9/11/2017 12:00:00 AM
Client Sample ID:	20170611-COMPOSITE-B	SDG No.:	LB90030
Lab Sample ID:	I5207-02	Matrix:	Solid
Level (low/med):	low	% Solid:	57.5

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.073	U	1	0.055		0.42	mg/Kg	09/12/2017	09/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/5/2017 12:00:00 AM
Project:	LB90030	Date Received:	9/11/2017 12:00:00 AM
Client Sample ID:	20170530-LOC-6	SDG No.:	LB90030
Lab Sample ID:	I5207-03	Matrix:	Solid
Level (low/med):	low	% Solid:	46.5

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.081	U	1	0.068		0.52	mg/Kg	09/12/2017	09/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/11/2017 12:00:00 AM
Project:	LB90030	Date Received:	9/11/2017 12:00:00 AM
Client Sample ID:	CL-02-091117-A	SDG No.:	LB90030
Lab Sample ID:	I5208-01	Matrix:	Solid
Level (low/med):	low	% Solid:	91.6

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.016	U	1	0.035		0.26	mg/Kg	09/12/2017	09/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/11/2017 12:00:00 AM
Project:	LB90030	Date Received:	9/11/2017 12:00:00 AM
Client Sample ID:	CL-02-091117-B	SDG No.:	LB90030
Lab Sample ID:	I5208-04	Matrix:	Solid
Level (low/med):	low	% Solid:	93.4

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	-0.027	U	1	0.034		0.26	mg/Kg	09/12/2017	09/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/11/2017 12:00:00 AM
Project:	LB90030	Date Received:	9/11/2017 12:00:00 AM
Client Sample ID:	CL-02-091117-C	SDG No.:	LB90030
Lab Sample ID:	I5208-07	Matrix:	Solid
Level (low/med):	low	% Solid:	93.9

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.026	U	1	0.034		0.26	mg/Kg	09/12/2017	09/12/2017	9012B

Report of Analysis

Client:		Date Collected:	9/11/2017 12:00:00 AM
Project:	LB90030	Date Received:	9/11/2017 12:00:00 AM
Client Sample ID:	CL-02-091117-D	SDG No.:	LB90030
Lab Sample ID:	I5208-10	Matrix:	Solid
Level (low/med):	low	% Solid:	93.3

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.039	J	1	0.034		0.26	mg/Kg	09/12/2017	09/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/11/2017 12:00:00 AM
Project:	LB90030	Date Received:	9/11/2017 12:00:00 AM
Client Sample ID:	CL-02-091117-E	SDG No.:	LB90030
Lab Sample ID:	I5208-13	Matrix:	Solid
Level (low/med):	low	% Solid:	92.2

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.021	U	1	0.034		0.26	mg/Kg	09/12/2017	09/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/11/2017 12:00:00 AM
Project:	LB90030	Date Received:	9/11/2017 12:00:00 AM
Client Sample ID:	OR-03-91117	SDG No.:	LB90030
Lab Sample ID:	I5209-01	Matrix:	Solid
Level (low/med):	low	% Solid:	92.8

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.04	J	1	0.034		0.26	mg/Kg	09/12/2017	09/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB90030	Date Received:	
Client Sample ID:	LOWPB102274	SDG No.:	LB90030
Lab Sample ID:	LOWPB102274	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.0093		1	0.003		0.005	mg/L		09/12/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB90030	Date Received:	
Client Sample ID:	LOWPB102274	SDG No.:	LB90030
Lab Sample ID:	LOWPB102274	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
-----	-----------	-------	------	----	-----	-----	------------	-------	-----------	-----------	----------

Color Before:	Clarity Before:	Texture:
Color After:	Clarity After:	Artifacts:
Comments:		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB90030

SDG No.: LB90030

Contract: _____ Lab Code: CHEM

Case No.: LB90030

SAS No.: LB90030

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	% RSD	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV1	Cyanide	0.099	0.099	100	0	85 - 115		09/12/2017	16:02	LB90030



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB90030

SDG No.: LB90030

Contract: _____ Lab Code: CHEM

Case No.: LB90030

SAS No.: LB90030

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	% RSD	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV1	Cyanide	0.25	0.25	100	0	90 - 110		09/12/2017	16:02	LB90030
CCV2	Cyanide	0.25	0.25	100	0	90 - 110		09/12/2017	16:17	LB90030
CCV3	Cyanide	0.25	0.25	100	0	90 - 110		09/12/2017	16:25	LB90030
CCV4	Cyanide	0.25	0.25	100	0	90 - 110		09/12/2017	16:25	LB90030



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB90030 **SDG No.:** LB90030
Contract: **Lab Code:** CHEM **Case No.:** LB90030 **SAS No.:** LB90030

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	Cyanide	-0	+/-0.005	U		0.005		09/12/2017	16:02	LB90030

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	Cyanide	-0.00019	+/-0.005	U		0.005		09/12/2017	16:02	LB90030
CCB2	Cyanide	-0.00024	+/-0.005	U		0.005		09/12/2017	16:17	LB90030
CCB3	Cyanide	-0.00059	+/-0.005	U		0.005		09/12/2017	16:25	LB90030
CCB4	Cyanide	0.00025	+/-0.005	U		0.005		09/12/2017	16:25	LB90030



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB90030

SDG No.: LB90030

Contract: Lab Code: CHEM

Case No.: LB90030

SAS No.: LB90030

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	-----	------	---	------------------	------------------	---------------



GENCHEM
- 3b -
PREPARATION BLANK SUMMARY

Client: LB90030

SDG No.: LB90030

Instrument: Konelab 20

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	LOD mg/Kg	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB102274BL	Cyanide	SOLID		Batch Number:		PB102274		Prep Date:	09/12/2017	
		0.0054	<0.25	U		0.25		09/12/2017	16:02	LB90030
Sample ID	Analyte	Result (mg/L)	Acceptance Limit	Conc Qual	LOD mg/L	CRQL mg/L	M	Analysis Date	Analysis Time	Run
PB102275BL	Cyanide	WATER		Batch Number:		PB102275		Prep Date:	09/12/2017	
		-0.0005	<0.005	U		0.005		09/12/2017	16:17	LB90030



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB90030 level: low sdg no.: LB90030
contract: lab code: CHEM case no.: LB90030 sas no.: LB90030
matrix: Solid sample id: I5201-01 client id: SU-03-091117MS
Percent Solids for Sample: 89.7 Spiked ID: I5201-01MS Percent Solids for Spike Sample: 89.7

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	1.9		0.036	U	2.2	86.4		



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB90030 level: low sdg no.: LB90030
contract: lab code: CHEM case no.: LB90030 sas no.: LB90030
matrix: Solid sample id: I5201-01 client id: SU-03-091117MSD
Percent Solids for Sample: 89.7 Spiked ID: I5201-01MSD Percent Solids for Spike Sample: 89.7

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	1.9		0.036	U	2.2	86.4		



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB90030 level: low sdg no.: LB90030
contract: lab code: CHEM case no.: LB90030 sas no.: LB90030
matrix: WATER sample id: I5202-01 client id: A5972-OTMS
Percent Solids for Sample: 0 Spiked ID: I5202-01MS Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/L	75 - 125	0.036		0.003	U	0.04	90		



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB90030 level: low sdg no.: LB90030
contract: lab code: CHEM case no.: LB90030 sas no.: LB90030
matrix: WATER sample id: I5202-01 client id: A5972-OTMSD
Percent Solids for Sample: 0 Spiked ID: I5202-01MSD Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/L	75 - 125	0.035		0.003	U	0.04	87.5		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB90030 **Level:** LOW **SDG No.:** LB90030
Contract: **Lab Code:** CHEM **Case No.:** LB90030 **SAS No.:** LB90030
Matrix: Solid **Sample ID:** I5201-01 **Client ID:** SU-03-091117DUP
Percent Solids for Sample: 89.7 **Duplicate ID** I5201-01DUP **Percent Solids for Spike Sample:** 89.7

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	0.036	U	-0.034	U			

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB90030 Level: LOW SDG No.: LB90030
Contract: _____ Lab Code: CHEM Case No.: LB90030 SAS No.: LB90030
Matrix: Solid Sample ID: I5201-01MS Client ID: SU-03-091117MSD
Percent Solids for Sample: 89.7 Duplicate ID I5201-01MSD Percent Solids for Spike Sample: 89.7

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	1.9		1.9		0		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB90030 **Level:** LOW **SDG No.:** LB90030
Contract: **Lab Code:** CHEM **Case No.:** LB90030 **SAS No.:** LB90030
Matrix: WATER **Sample ID:** I5202-01 **Client ID:** A5972-OTDUP
Percent Solids for Sample: 0 **Duplicate ID** I5202-01DUP **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/L	20	0.003	U	0.00081	U			

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB90030 **Level:** LOW **SDG No.:** LB90030
Contract: **Lab Code:** CHEM **Case No.:** LB90030 **SAS No.:** LB90030
Matrix: WATER **Sample ID:** I5202-01MS **Client ID:** A5972-OTMSD
Percent Solids for Sample: 0 **Duplicate ID** I5202-01MSD **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/L	20	0.036		0.035		3		



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB90030

SDG No.: LB90030

Contract:

Lab Code: CHEM

Case No.: LB90030

SAS No.: LB90030

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB102274BS Cyanide	mg/Kg	4.9	5.2		106.1	85 - 115	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB90030

SDG No.: LB90030

Contract:

Lab Code: CHEM

Case No.: LB90030

SAS No.: LB90030

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB102275BS Cyanide	mg/L	0.1	0.1		100	85 - 115	