

Chemtech Consulting Group

Analytical Review Report

Date Printed : 9/29/17
Analyst : AK
Data File : LB90316.csv

Approved By : 
Approved Date : 9/29/18
Worksheet # :

M9056A - Inorganic Anions-13

Lab Sample ID	Client ID		Raw Amt PPB	Dil	Matrix	A. Date	Prep Method	Analysis Method			Line 1	
Parameter						%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Sulfate												
STD1	STD1				W	9/6/17						
Sulfate		PASS	0.000		0.000						mg/L	
STD2	STD2				W	9/6/17						
Sulfate		PASS	1.436		1.44						mg/L	
STD3	STD3				W	9/6/17						
Sulfate		PASS	3.325		3.32						mg/L	
STD4	STD4				W	9/6/17						
Sulfate		PASS	5.940		5.94						mg/L	
STD5	STD5				W	9/6/17						
Sulfate		PASS	14.169		14.2						mg/L	
STD6	STD6				W	9/6/17						
Sulfate		PASS	29.252		29.3						mg/L	
STD7	STD7				W	9/6/17						
Sulfate		PASS	38.229		38.2						mg/L	
ICV1	ICV1				W	9/6/17						
Sulfate		PASS	14.593		14.60	97.0	90	110			mg/L	
ICB1	ICB1				W	9/6/17						
Sulfate		PASS	0.000		0.000			+/-0.1320			mg/L	
CCV1	CCV1				W	9/26/17						
Sulfate		PASS	14.051		14.10	94.0	90	110			mg/L	
CCB1	CCB1				W	9/26/17						
Sulfate		PASS	0.000		0.000			+/-0.1320			mg/L	
LB90316BLS	LB90316BLS				S	9/26/17						
Sulfate		PASS	0.000		0.000			+/-0.0960			mg/Kg	
LB90316BSS	LB90316BSS				S	9/26/17						
Sulfate		PASS	14.340		286.00	95.0	90.00	110.00			mg/Kg	
I5417-01	B-2-2-4		1		S	9/26/17						
Sulfate		PASS	2.174		48.6						mg/Kg	
I5417-01D	B-2-2-4D		1		S	9/26/17						
Sulfate		PASS	2.179		48.80				0.4	20	mg/Kg	
I5417-01S	B-2-2-4S		1		S	9/26/17						
Sulfate		PASS	17.580		393.0	103.0	80	120			mg/Kg	
I5417-01SD	B-2-2-4SD		1		S	9/26/17						
Sulfate		PASS	17.621		394.0	103.0	80	120	0.3	15	mg/Kg	
CCV2	CCV2				W	9/26/17						
Sulfate		PASS	14.320		14.30	95.0	90	110			mg/L	
CCB2	CCB2				W	9/26/17						
Sulfate		PASS	0.000		0.000			+/-0.1320			mg/L	