

Daily Analysis Runlog For Sequence/QC Batch ID # LB90526

Review By		Review On				
SubDirectory		LB90526		Test : Cyanide		
STD. NAME		STD REF.#				
ICAL Standard		WP58478,WP58479,WP58480,WP58481,WP58482,WP58483,WP58476				
ICV Standard		wp58485				
CCV Standard		wp58484				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		N/A				
Chk Standard		WP58486,WP55355,WP55836				
Sr#	SampleId	ClientID	QcType	Date	Comment	Status
1	0.0PPBCN	0.0PPBCN	CAL1	10/03/17 09:59		OK
2	5.0PPBCN	5.0PPBCN	CAL2	10/03/17 09:59		OK
3	10PPBCN	10PPBCN	CAL3	10/03/17 09:59		OK
4	50PPBCN	50PPBCN	CAL4	10/03/17 09:59		OK
5	100PPBCN	100PPBCN	CAL5	10/03/17 09:59		OK
6	250PPBCN	250PPBCN	CAL6	10/03/17 09:59		OK
7	500PPBCN	500PPBCN	CAL7	10/03/17 09:59		OK
8	ICV1	ICV1	ICV	10/03/17 14:25		OK
9	ICB1	ICB1	ICB	10/03/17 14:25		OK
10	CCV1	CCV1	CCV	10/03/17 14:25		OK
11	CCB1	CCB1	CCB	10/03/17 14:25		OK
12	PB102820BL	PB102820BL	MB	10/03/17 14:25		OK
13	PB102820BS	PB102820BS	LCS	10/03/17 14:25		OK
14	LOWPB102820	LOWPB102820	SAM	10/03/17 14:33		OK
15	HIGHPB102820	HIGHPB102820	SAM	10/03/17 14:33		OK
16	I5540-01MS	MH-6MS	MS	10/03/17 14:33		OK
17	I5540-01MSD	MH-6MSD	MSD	10/03/17 14:33		OK
18	PB102868BL	PB102868BL	MB	10/03/17 14:33		OK
19	PB102868BS	PB102868BS	LCS	10/03/17 14:33		OK
20	CCV2	CCV2	CCV	10/03/17 14:40		OK
21	CCB2	CCB2	CCB	10/03/17 14:40		OK

Daily Analysis Runlog For Sequence/QC Batch ID # LB90526

Review By		Review On				
SubDirectory		LB90526		Test : Cyanide		
STD. NAME		STD REF.#				
ICAL Standard		WP58478,WP58479,WP58480,WP58481,WP58482,WP58483,WP58476				
ICV Standard		wp58485				
CCV Standard		wp58484				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		N/A				
Chk Standard		WP58486,WP55355,WP55836				
22	LOWPB102868	LOWPB102868	SAM	10/03/17 14:40		OK
23	HIGHPB102868	HIGHPB102868	SAM	10/03/17 14:40		OK
24	I5432-01	FK-BPM-01-E7-E8-F	SAM	10/03/17 14:40		OK
25	I5584-01	TP-1-COMP	SAM	10/03/17 14:40		OK
26	I5584-01DUP	TP-1-COMPDUP	DUP	10/03/17 14:40		OK
27	I5584-01MS	TP-1-COMPMS	MS	10/03/17 14:40		OK
28	I5584-01MSD	TP-1-COMPMSD	MSD	10/03/17 14:40		OK
29	I5585-01	G67-0-1	SAM	10/03/17 14:40		OK
30	I5585-03	G67-1-2	SAM	10/03/17 14:40		OK
31	I5585-05	G68-0-1.5	SAM	10/03/17 14:48		OK
32	CCV3	CCV3	CCV	10/03/17 14:48		OK
33	CCB3	CCB3	CCB	10/03/17 14:48		OK
34	I5585-07	G68-1.5-3	SAM	10/03/17 14:48		OK
35	I5585-09	G69-0-1.5	SAM	10/03/17 14:48		OK
36	I5585-11	G69-1.5-3	SAM	10/03/17 14:48		OK
37	I5585-13	G70-0-2.5	SAM	10/03/17 14:48		OK
38	I5585-15	G70-2.5-5	SAM	10/03/17 14:48		OK
39	I5585-17	G71-0-1.5	SAM	10/03/17 14:48		OK
40	I5585-19	G71-1.5-3	SAM	10/03/17 14:48		OK
41	I5540-01	MH-6	SAM	10/03/17 15:19		OK
42	I5540-01DUP	MH-6DUP	DUP	10/03/17 15:19		OK
43	CCV4	CCV4	CCV	10/03/17 15:19		OK



Instrument ID : KONELAB

284 Sheffield Street, Mountainside, New Jersey 07092 Phone : 908 789 8900 Fax : 908 789 8922

Daily Analysis Runlog For Sequence/QCBatch ID # LB90526

Review By		Review On				
SubDirectory		LB90526		Test : Cyanide		
STD. NAME		STD REF.#				
ICAL Standard		WP58478,WP58479,WP58480,WP58481,WP58482,WP58483,WP58476				
ICV Standard		wp58485				
CCV Standard		wp58484				
ICSA Standard		N/A				
CRI Standard		N/A				
LCS Standard		N/A				
Chk Standard		WP58486,WP55355,WP55836				
44	CCB4	CCB4	CCB	10/03/17 15:19		OK



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB90526	Date Received:	
Client Sample ID:	HIGHPB102820	SDG No.:	LB90526
Lab Sample ID:	HIGHPB102820	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.49		1	0.003		0.005	mg/L		10/03/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB90526	Date Received:	
Client Sample ID:	HIGHPB102868	SDG No.:	LB90526
Lab Sample ID:	HIGHPB102868	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.48		1	0.003		0.005	mg/L		10/03/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/26/2017 12:00:00 AM
Project:	LB90526	Date Received:	9/26/2017 12:00:00 AM
Client Sample ID:	FK-BPM-01-E7-E8-F7-F8	SDG No.:	LB90526
Lab Sample ID:	I5432-01	Matrix:	Solid
Level (low/med):	low	% Solid:	91.9

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.043	J	1	0.035		0.26	mg/Kg	10/03/2017	10/03/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/28/2017 12:00:00 AM
Project:	LB90526	Date Received:	9/29/2017 12:00:00 AM
Client Sample ID:	MH-6	SDG No.:	LB90526
Lab Sample ID:	I5540-01	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.0028	U	1	0.003		0.005	mg/L	10/02/2017	10/03/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/29/2017 12:00:00 AM
Project:	LB90526	Date Received:	9/29/2017 12:00:00 AM
Client Sample ID:	TP-1-COMP	SDG No.:	LB90526
Lab Sample ID:	I5584-01	Matrix:	Solid
Level (low/med):	low	% Solid:	86

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.039	J	1	0.037		0.28	mg/Kg	10/03/2017	10/03/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/29/2017 12:00:00 AM
Project:	LB90526	Date Received:	9/29/2017 12:00:00 AM
Client Sample ID:	G67-0-1	SDG No.:	LB90526
Lab Sample ID:	I5585-01	Matrix:	Solid
Level (low/med):	low	% Solid:	93.8

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.097	J	1	0.034		0.25	mg/Kg	10/03/2017	10/03/2017	9012B

Report of Analysis

Client:		Date Collected:	9/29/2017 12:00:00 AM
Project:	LB90526	Date Received:	9/29/2017 12:00:00 AM
Client Sample ID:	G67-1-2	SDG No.:	LB90526
Lab Sample ID:	I5585-03	Matrix:	Solid
Level (low/med):	low	% Solid:	92.5

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.097	J	1	0.034		0.26	mg/Kg	10/03/2017	10/03/2017	9012B

Report of Analysis

Client:		Date Collected:	9/29/2017 12:00:00 AM
Project:	LB90526	Date Received:	9/29/2017 12:00:00 AM
Client Sample ID:	G68-0-1.5	SDG No.:	LB90526
Lab Sample ID:	I5585-05	Matrix:	Solid
Level (low/med):	low	% Solid:	93.6

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.21	J	1	0.034		0.26	mg/Kg	10/03/2017	10/03/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/29/2017 12:00:00 AM
Project:	LB90526	Date Received:	9/29/2017 12:00:00 AM
Client Sample ID:	G68-1.5-3	SDG No.:	LB90526
Lab Sample ID:	I5585-07	Matrix:	Solid
Level (low/med):	low	% Solid:	94.1

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.073	J	1	0.034		0.26	mg/Kg	10/03/2017	10/03/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/29/2017 12:00:00 AM
Project:	LB90526	Date Received:	9/29/2017 12:00:00 AM
Client Sample ID:	G69-0-1.5	SDG No.:	LB90526
Lab Sample ID:	I5585-09	Matrix:	Solid
Level (low/med):	low	% Solid:	93.6

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.15	J	1	0.034		0.26	mg/Kg	10/03/2017	10/03/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/29/2017 12:00:00 AM
Project:	LB90526	Date Received:	9/29/2017 12:00:00 AM
Client Sample ID:	G69-1.5-3	SDG No.:	LB90526
Lab Sample ID:	I5585-11	Matrix:	Solid
Level (low/med):	low	% Solid:	94.4

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.015	U	1	0.034		0.26	mg/Kg	10/03/2017	10/03/2017	9012B

Report of Analysis

Client:		Date Collected:	9/29/2017 12:00:00 AM
Project:	LB90526	Date Received:	9/29/2017 12:00:00 AM
Client Sample ID:	G70-0-2.5	SDG No.:	LB90526
Lab Sample ID:	I5585-13	Matrix:	Solid
Level (low/med):	low	% Solid:	92.6

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.25	J	1	0.034		0.26	mg/Kg	10/03/2017	10/03/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/29/2017 12:00:00 AM
Project:	LB90526	Date Received:	9/29/2017 12:00:00 AM
Client Sample ID:	G70-2.5-5	SDG No.:	LB90526
Lab Sample ID:	I5585-15	Matrix:	Solid
Level (low/med):	low	% Solid:	96.5

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.05	J	1	0.033		0.25	mg/Kg	10/03/2017	10/03/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	9/29/2017 12:00:00 AM
Project:	LB90526	Date Received:	9/29/2017 12:00:00 AM
Client Sample ID:	G71-0-1.5	SDG No.:	LB90526
Lab Sample ID:	I5585-17	Matrix:	Solid
Level (low/med):	low	% Solid:	94.5

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.12	J	1	0.034		0.26	mg/Kg	10/03/2017	10/03/2017	9012B

Report of Analysis

Client:		Date Collected:	9/29/2017 12:00:00 AM
Project:	LB90526	Date Received:	9/29/2017 12:00:00 AM
Client Sample ID:	G71-1.5-3	SDG No.:	LB90526
Lab Sample ID:	I5585-19	Matrix:	Solid
Level (low/med):	low	% Solid:	94.7

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.07	J	1	0.034		0.25	mg/Kg	10/03/2017	10/03/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB90526	Date Received:	
Client Sample ID:	LOWPB102820	SDG No.:	LB90526
Lab Sample ID:	LOWPB102820	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.0097		1	0.003		0.005	mg/L		10/03/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB90526	Date Received:	
Client Sample ID:	LOWPB102868	SDG No.:	LB90526
Lab Sample ID:	LOWPB102868	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
57-12-5	Cyanide	0.0094		1	0.003		0.005	mg/L		10/03/2017	9012B



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Report of Analysis

Client:		Date Collected:	
Project:	LB90526	Date Received:	
Client Sample ID:	LOWPB102868	SDG No.:	LB90526
Lab Sample ID:	LOWPB102868	Matrix:	WATER
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
-----	-----------	-------	------	----	-----	-----	------------	-------	-----------	-----------	----------

Color Before:	Clarity Before:	Texture:
Color After:	Clarity After:	Artifacts:
Comments:		

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
D = Dilution
Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
* = indicates the duplicate analysis is not within control limits.
E = Indicates the reported value is estimated because of the presence of interference.
OR = Over Range
N = Spiked sample recovery not within control limits



GENCHEM

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: LB90526

SDG No.: LB90526

Contract: _____ Lab Code: CHEM

Case No.: LB90526

SAS No.: LB90526

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result mg/L	True Value	% Recovery	% RSD	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV1	Cyanide	0.24	0.25	96	0	90 - 110		10/03/2017	14:25	LB90526
CCV2	Cyanide	0.24	0.25	96	0	90 - 110		10/03/2017	14:40	LB90526
CCV3	Cyanide	0.24	0.25	96	0	90 - 110		10/03/2017	14:48	LB90526
CCV4	Cyanide	0.26	0.25	104	0	90 - 110		10/03/2017	15:19	LB90526



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB90526

SDG No.: LB90526

Contract:

Lab Code: CHEM

Case No.: LB90526

SAS No.: LB90526

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB1	Cyanide	0.001	+/-0.005	U		0.005		10/03/2017	14:25	LB90526



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB90526

SDG No.: LB90526

Contract:

Lab Code: CHEM

Case No.: LB90526

SAS No.: LB90526

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB1	Cyanide	0.00079	+/-0.005	U		0.005		10/03/2017	14:25	LB90526
CCB2	Cyanide	0.0012	+/-0.005	U		0.005		10/03/2017	14:40	LB90526
CCB3	Cyanide	0.00073	+/-0.005	U		0.005		10/03/2017	14:48	LB90526
CCB4	Cyanide	0.00066	+/-0.005	U		0.005		10/03/2017	15:19	LB90526



GENCHEM

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: LB90526

SDG No.: LB90526

Contract: Lab Code: CHEM

Case No.: LB90526

SAS No.: LB90526

Sample ID	Analyte	Result mg/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
-----------	---------	----------------	---------------------	--------------	-----	------	---	------------------	------------------	---------------



GENCHEM
- 3b -
PREPARATION BLANK SUMMARY

Client: LB90526

SDG No.: LB90526

Instrument: Konelab 20

Sample ID	Analyte	Result (mg/L)	Acceptance Limit	Conc Qual	LOD mg/L	CRQL mg/L	M	Analysis Date	Analysis Time	Run
PB102820BL	Cyanide	WATER		Batch Number:		PB102820		Prep Date:	10/02/2017	
		0.00093	<0.005	U		0.005		10/03/2017	14:25	LB90526
Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	LOD mg/Kg	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB102868BL	Cyanide	SOLID		Batch Number:		PB102868		Prep Date:	10/03/2017	
		0.058	<0.25	J		0.25		10/03/2017	14:33	LB90526



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB90526 level: low sdg no.: LB90526
contract: lab code: CHEM case no.: LB90526 sas no.: LB90526
matrix: WATER sample id: I5540-01 client id: MH-6MS
Percent Solids for Sample: 0 Spiked ID: I5540-01MS Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/L	75 - 125	0.033		0.003	U	0.04	82.5		



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB90526 level: low sdg no.: LB90526
contract: lab code: CHEM case no.: LB90526 sas no.: LB90526
matrix: WATER sample id: I5540-01 client id: MH-6MSD
Percent Solids for Sample: 0 Spiked ID: I5540-01MSD Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/L	75 - 125	0.033		0.003	U	0.04	82.5		



GENCHEM
- 5a -
MATRIX SPIKE SUMMARY

client: LB90526 level: low sdg no.: LB90526
contract: lab code: CHEM case no.: LB90526 sas no.: LB90526
matrix: Solid sample id: I5584-01 client id: TP-1-COMPMS
Percent Solids for Sample: 86 Spiked ID: I5584-01MS Percent Solids for Spike Sample: 86

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	1.8		0.039	J	2.3	77		



GENCHEM
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: LB90526 level: low sdg no.: LB90526
contract: lab code: CHEM case no.: LB90526 sas no.: LB90526
matrix: Solid sample id: I5584-01 client id: TP-1-COMPMSD
Percent Solids for Sample: 86 Spiked ID: I5584-01MSD Percent Solids for Spike Sample: 86

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Cyanide	mg/Kg	75 - 125	1.8		0.039	J	2.3	77		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB90526 **Level:** LOW **SDG No.:** LB90526
Contract: **Lab Code:** CHEM **Case No.:** LB90526 **SAS No.:** LB90526
Matrix: WATER **Sample ID:** I5540-01 **Client ID:** MH-6DUP
Percent Solids for Sample: 0 **Duplicate ID** I5540-01DUP **Percent Solids for Spike Sample:** 0

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/L	20	0.003	U	0.003	J	200.0		



GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB90526 Level: LOW SDG No.: LB90526
Contract: _____ Lab Code: CHEM Case No.: LB90526 SAS No.: LB90526
Matrix: WATER Sample ID: I5540-01MS Client ID: MH-6MSD
Percent Solids for Sample: 0 Duplicate ID I5540-01MSD Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/L	20	0.033		0.033		0		

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB90526 **Level:** LOW **SDG No.:** LB90526
Contract: **Lab Code:** CHEM **Case No.:** LB90526 **SAS No.:** LB90526
Matrix: Solid **Sample ID:** I5584-01 **Client ID:** TP-1-COMPDUP
Percent Solids for Sample: 86 **Duplicate ID** I5584-01DUP **Percent Solids for Spike Sample:** 86

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	0.039	J	0.024	U	200.0	*	

GENCHEM

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: LB90526 **Level:** LOW **SDG No.:** LB90526
Contract: **Lab Code:** CHEM **Case No.:** LB90526 **SAS No.:** LB90526
Matrix: Solid **Sample ID:** I5584-01MS **Client ID:** TP-1-COMPMSD
Percent Solids for Sample: 86 **Duplicate ID** I5584-01MSD **Percent Solids for Spike Sample:** 86

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Cyanide	mg/Kg	20	1.8		1.8		0		



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB90526 SDG No.: LB90526
Contract: Lab Code: CHEM Case No.: LB90526 SAS No.: LB90526

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB102820BS Cyanide	mg/L	0.1	0.096		96	85 - 115	



GENCHEM

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: LB90526

SDG No.: LB90526

Contract:

Lab Code: CHEM

Case No.: LB90526

SAS No.: LB90526

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB102868BS Cyanide	mg/Kg	4.9	4.9		100	85 - 115	