

IC-2 300.0
10/10/17 AK


Clear table

ident	concentrati on F-	concentrati on CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2017-09-18_	9/18/17 11:14
STD2	0.1400	0.2300	0.2670	0.8860	0.2300	-0.0150	1.4070	_2017-09-18_	9/18/17 11:43
STD3	0.4340	0.6550	0.6670	2.2590	0.5620	1.5520	3.2660	_2017-09-18_	9/18/17 12:12
STD4	0.7900	1.1620	1.1920	4.0390	0.9720	3.4220	5.6960	_2017-09-18_	9/18/17 12:41
STD5	2.0010	3.0010	2.9470	9.8370	2.3700	6.2830	14.3960	_2017-09-18_	9/18/17 13:10
STD6	3.9040	5.9100	5.8430	19.5270	4.8540	12.1460	28.5870	_2017-09-18_	9/18/17 13:39
STD7	5.1160	7.5820	7.6010	25.2570	6.3140	12.8120	38.2920	_2017-09-18_	9/18/17 14:08
ICV	1.9640	2.8510	2.8810	9.9890	2.4840	4.8410	14.3260	_2017-09-18_	9/18/17 15:34
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2017-09-18_	9/18/17 16:03
CCV	2.0540	2.8300	2.8850	9.6730	2.3260	4.9750	14.2880	_2017-09-22_	9/22/17 9:26
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2017-09-22_	9/22/17 9:56
LB90688BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2017-09-22_	9/22/17 10:25
LB90688BSW	2.0130	2.8490	2.8650	9.7660	2.3730	5.2690	14.2470	_2017-09-22_	9/22/17 10:55
I5334-01	3.0440	205.0140	0.0000	0.0000	8.3550	2.3480	39.2190	_2017-09-22_	9/22/17 11:26
I5334-01DLX2	1.3710	95.7660	0.0000	0.0000	4.0320	1.3570	19.0280	_2017-09-22_	9/22/17 12:56
I5334-01DLX50	0.0000	2.9180	0.0000	0.0000	0.2250	0.0000	1.3100	_2017-09-22_	9/22/17 13:25
CCV	1.9990	2.8410	2.8870	9.7230	2.3450	4.7500	14.1090	_2017-09-22_	9/22/17 13:55
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	_2017-09-22_	9/22/17 14:25

IC-1 9-18-17
AK

CALIBRATION OF COMPONENT F-

Method: ANIONS 09-18-17.mtw
Equation: $Q = 0.880227 \cdot A + 1.34751$
RSD: 4.737 %
Correlation coefficient: 0.999084




K3 = 0 K2 = 0 K1 = 0.880227 K0 = 1.34751
Base: Area
Ref.channel: Cond
ISTD:
Formula: Linear
Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		No
2	0.06841	1.579	0.1	20	5.813	Yes	
3	0.3113	8.211	0.4	20	5.813	Yes	
4	0.6668	16.23	0.8	20	5.813	Yes	
5	1.903	43.52	2	20	5.813	Yes	
6	3.946	86.03	4	20	5.813	Yes	
7	5.194	114.7	5	20	5.813	Yes	

CALIBRATION OF COMPONENT CL-

Method: ANIONS 09-18-17.mtw
 Equation: $Q = 1.25786 \cdot A + 1.32191$
 RSD: 2.527 %
 Correlation coefficient: 0.999739




K3 = 0 K2 = 0 K1 = 1.25786 K0 = 1.32191
 Base: Area
 Ref.channel: Cond
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		No
2	0.1078	2.202	0.15	20	8.549	Yes	
3	0.4298	8.932	0.6	20	8.549	Yes	
4	0.8256	16.98	1.2	20	8.549	Yes	
5	2.204	46.62	3	20	8.549	Yes	
6	4.49	92.82	6	20	8.549	Yes	
7	5.818	119.5	7.5	20	8.549	Yes	

CALIBRATION OF COMPONENT NO2

Method: ANIONS 09-18-17.mtw
 Equation: $Q = 0.541087 \cdot A + 1.82619$
 RSD: 2.840 %
 Correlation coefficient: 0.999671




K3 = 0 K2 = 0 K1 = 0.541087 K0 = 1.82619
 Base: Area
 Ref.channel: Cond
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		No
2	0.2064	4.691	0.15	20	10.12	Yes	
3	0.8557	19.62	0.6	20	10.12	Yes	
4	1.717	39.22	1.2	20	10.12	Yes	
5	4.638	106.2	3	20	10.12	Yes	
6	9.503	214.4	6	20	10.12	Yes	
7	12.35	277.6	7.5	20	10.12	Yes	

CALIBRATION OF COMPONENT BR-

Method: ANIONS 09-18-17.mtw
 Equation: $Q = 2.97146 \cdot A - 6.10153$
 RSD: 2.170 %
 Correlation coefficient: 0.999808




K3 = 0 K2 = 0 K1 = 2.97146 K0 = -6.10153

Base: Area
 Ref.channel: Cond
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		No
2	0.1756	6.574	0.5	20	12.77	Yes	
3	0.4627	15.94	2	20	12.77	Yes	
4	0.8151	28.1	4	20	12.77	Yes	
5	1.897	68.74	10	20	12.77	Yes	
6	3.405	134.9	20	20	12.77	Yes	
7	4.147	172.1	25	20	12.77	Yes	

CALIBRATION OF COMPONENT NO3

Method: ANIONS 09-18-17.mtw
 Equation: $Q = 0.503233 \cdot A + 1.80511$
 RSD: 2.906 %
 Correlation coefficient: 0.999655




K3 = 0 K2 = 0 K1 = 0.503233 K0 = 1.80511

Base: Area
 Ref.channel: Cond
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		No
2	0.1469	4.165	0.125	20	14.38	Yes	
3	0.6036	17.75	0.5	20	14.38	Yes	
4	1.192	34.47	1	20	14.38	Yes	
5	3.187	92.04	2.5	20	14.38	Yes	
6	6.528	193.8	5	20	14.38	Yes	
7	8.504	247.4	6.25	20	14.38	Yes	

CALIBRATION OF COMPONENT HPO4

Method: ANIONS 09-18-17.mtw
 Equation: $Q = 1.82468 \cdot A + 1.29701$
 RSD: 5.131 %
 Correlation coefficient: 0.998925




K3 = 0 K2 = 0 K1 = 1.82468 K0 = 1.29701

Base: Area
 Ref.channel: Cond
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		No
2	0.0485	3.152	0.25	20	21.58	Yes	
3	0.1779	11.1	1	20	21.58	Yes	
4	0.3349	20.58	2	20	21.58	Yes	
5	0.8842	53.71	5	20	21.58	Yes	
6	1.831	104.5	10	20	21.58	Yes	
7	2.412	139.7	12.5	20	21.58	Yes	

CALIBRATION OF COMPONENT SO4

Method: ANIONS 09-18-17.mtw
 Equation: $Q = 1.69708 \cdot A + 11.5137$
 RSD: 4.420 %
 Correlation coefficient: 0.999202



K3 = 0 K2 = 0 K1 = 1.69708 K0 = 11.5137
 Base: Area
 Ref.channel: Cond
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		No
2	0.1778	8.05	0.75	20	24.1		Yes
3	0.6998	30.92	3	20	24.1		Yes
4	1.385	60.83	6	20	24.1		Yes
5	3.762	165.1	15	20	24.1		Yes
6	7.815	337	30	20	24.1		Yes
7	10.33	444.5	37.5	20	24.1		Yes

Report date: 9/18/2017 5:33:49 PM
Printed by: wet


Ident: STD1
Analysis from: 9/18/2017 11:14:33 AM
File: _2017-09-18_

Last save: 9/18/2017 1:58:08 PM

Method: ANIONS 09-18-17.mtw
Run operator: wet
Analysis number: 59025

Last save: 9/18/2017 11:40:22 AM

SAMPLE:
: AP/AK
Vial number: 2
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 9/18/2017 5:34:00 PM
Printed by: wet


Ident: STD2
Analysis from: 9/18/2017 11:43:29 AM
File: _2017-09-18_

Last save: 9/18/2017 1:58:08 PM

Method: ANIONS 09-18-17.mtw
Run operator: wet
Analysis number: 59026

Last save: 9/18/2017 11:40:22 AM

SAMPLE:
: AP/AK
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.81	0.344	0.07	7.32	1.579	5.19	0.
2	8.55	0.303	0.11	11.51	2.202	7.24	0.
3	10.12	0.335	0.21	22.07	4.691	15.42	0.
4	12.77	0.565	0.18	18.81	6.574	21.62	0.
5	14.38	0.441	0.15	15.69	4.165	13.69	0.
6	21.58	1.021	0.05	5.13	3.152	10.36	0.
7	24.10	0.654	0.18	19.03	8.050	26.47	0.
7	26.00	0.523	0.93	99.56	30.411	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 9/18/2017 5:34:08 PM
Printed by: wet


Ident: STD3
Analysis from: 9/18/2017 12:12:24 PM
File: _2017-09-18_

Last save: 9/18/2017 1:58:08 PM

Method: ANIONS 09-18-17.mtw
Run operator: wet
Analysis number: 59027

Last save: 9/18/2017 11:40:22 AM

SAMPLE:
: AP/AK
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.83	0.344	0.31	8.77	8.211	7.30	0.
2	8.55	0.306	0.43	12.13	8.932	7.94	0.
3	10.11	0.332	0.86	24.16	19.623	17.45	0.
4	12.71	0.529	0.46	13.05	15.945	14.18	0.
5	14.33	0.448	0.60	17.03	17.746	15.78	0.
6	21.55	0.980	0.18	5.01	11.097	9.87	0.
7	24.09	0.636	0.70	19.74	30.919	27.49	0.
7	26.00	0.511	3.54	99.88	112.472	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 9/18/2017 5:34:17 PM
Printed by: wet


Ident: STD4
Analysis from: 9/18/2017 12:41:20 PM
File: _2017-09-18_

Last save: 9/18/2017 1:58:10 PM

Method: ANIONS 09-18-17.mtw
Run operator: wet
Analysis number: 59028

Last save: 9/18/2017 11:40:22 AM

SAMPLE:
: AP/AK
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.94	0.332	0.67	9.60	16.230	7.50	0.
2	8.72	0.304	0.83	11.89	16.976	7.84	0.
3	10.27	0.331	1.72	24.74	39.218	18.12	0.
4	12.86	0.534	0.81	11.74	28.098	12.98	0.
5	14.46	0.440	1.19	17.18	34.466	15.93	0.
6	21.51	0.957	0.33	4.82	20.578	9.51	0.
7	24.09	0.630	1.38	19.96	60.829	28.11	0.
7	26.00	0.504	6.93	99.95	216.395	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 9/18/2017 5:34:23 PM
Printed by: wet


Ident: STD5
Analysis from: 9/18/2017 1:10:15 PM
File: _2017-09-18_

Last save: 9/18/2017 3:15:12 PM

Method: ANIONS 09-18-17.mtw
Run operator: wet
Analysis number: 59029

Last save: 9/18/2017 11:40:22 AM

SAMPLE:
: AP/AK
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.92	0.314	1.90	10.30	43.518	7.56	0.
2	8.72	0.310	2.20	11.93	46.624	8.09	0.
3	10.32	0.332	4.64	25.10	106.179	18.44	0.
4	12.90	0.569	1.90	10.27	68.738	11.93	0.
5	14.50	0.438	3.19	17.25	92.043	15.98	0.
6	21.58	0.929	0.88	4.78	53.706	9.32	0.
7	24.28	0.626	3.76	20.35	165.147	28.67	0.
7	26.00	0.503	18.47	99.98	575.955	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 9/18/2017 5:34:34 PM
Printed by: wet


Ident: STD6
Analysis from: 9/18/2017 1:39:11 PM
File: _2017-09-18_

Last save: 9/18/2017 3:15:12 PM

Method: ANIONS 09-18-17.mtw
Run operator: wet
Analysis number: 59030

Last save: 9/18/2017 11:40:22 AM

SAMPLE:
: AP/AK
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.81	0.300	3.94	10.51	86.032	7.39	0.
2	8.55	0.303	4.49	11.97	92.815	7.98	0.
3	10.10	0.329	9.50	25.33	214.422	18.43	0.
4	12.66	0.627	3.40	9.07	134.852	11.59	0.
5	14.28	0.450	6.53	17.40	193.752	16.65	0.
6	21.32	0.858	1.83	4.88	104.533	8.98	0.
7	24.00	0.615	7.81	20.83	337.024	28.97	0.
7	26.00	0.497	37.51	99.99	1163.430	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 9/18/2017 5:34:46 PM
Printed by: wet


Ident: STD7
Analysis from: 9/18/2017 2:08:06 PM
File: _2017-09-18_

Last save: 9/18/2017 3:17:50 PM

Method: ANIONS 09-18-17.mtw
Run operator: wet
Analysis number: 59031

Last save: 9/18/2017 11:40:22 AM

SAMPLE:
: AK/AP
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.86	0.308	5.19	10.65	114.718	7.57	0.
2	8.69	0.301	5.82	11.93	119.501	7.89	0.
3	10.25	0.328	12.35	25.33	277.577	18.32	0.
4	12.80	0.658	4.15	8.51	172.052	11.35	0.
5	14.38	0.440	8.50	17.44	247.356	16.32	0.
6	21.48	0.867	2.41	4.95	139.714	9.22	0.
7	24.20	0.616	10.33	21.18	444.488	29.33	0.
7	26.00	0.503	48.75	99.99	1515.406	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 9/18/2017 5:37:21 PM
Printed by: wet


Ident: ICV
Analysis from: 9/18/2017 3:34:53 PM
File: _2017-09-18_

Last save: 9/18/2017 4:00:42 PM

Method: ANIONS 09-18-17.mtw
Run operator: wet
Analysis number: 59034

Last save: 9/18/2017 3:17:20 PM

SAMPLE:
: AK/AP
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.81	0.308	1.88	10.29	43.091	7.57	0.
2	8.54	0.300	2.17	11.84	44.276	7.78	0.
3	10.08	0.326	4.60	25.13	103.110	18.11	0.
4	12.66	0.583	1.86	10.16	69.284	12.17	0.
5	14.31	0.457	3.16	17.28	95.135	16.71	0.
6	21.52	0.912	0.88	4.81	52.348	9.20	0.
7	24.15	0.620	3.75	20.47	162.044	28.46	0.
7	26.00	0.501	18.30	99.99	569.288	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 9/18/2017 5:38:00 PM
Printed by: wet


Ident: ICB
Analysis from: 9/18/2017 4:03:48 PM
File: _2017-09-18_

Last save: 9/18/2017 5:37:52 PM

Method: ANIONS 09-18-17.mtw
Run operator: wet
Analysis number: 59035

Last save: 9/18/2017 2:33:56 PM

SAMPLE:
: AK/AP
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 10/10/2017 9:45:03 AM
Printed by: wet


Ident: CCV
Analysis from: 9/22/2017 9:26:06 AM
File: _2017-09-22_

Last save: 9/22/2017 9:52:56 AM

Method: ANIONS 09-18-17.mtw
Run operator: wet
Analysis number: 59122

Last save: 9/21/2017 12:16:58 PM

SAMPLE:
: AK/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.76	0.332	1.81	10.95	45.138	8.01	0.
2	8.57	0.342	1.90	11.50	43.949	7.80	0.
3	10.19	0.379	3.98	24.07	103.256	18.31	0.
4	12.89	0.518	1.99	12.06	67.161	11.91	0.
5	14.57	0.495	2.71	16.37	88.851	15.76	0.
6	21.13	0.993	0.84	5.08	53.819	9.55	0.
7	23.86	0.718	3.30	19.97	161.604	28.66	0.
7	27.00	0.540	16.54	99.98	563.778	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 10/10/2017 9:45:27 AM
Printed by: wet


Ident: CCB
Analysis from: 9/22/2017 9:56:01 AM
File: _2017-09-22_

Last save: 10/10/2017 9:45:22 AM

Method: ANIONS 09-18-17.mtw
Run operator: wet
Analysis number: 59123

Last save: 9/21/2017 12:16:58 PM

SAMPLE:
: AK/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 10/10/2017 9:45:34 AM
Printed by: wet


Ident: LB90688BLW
Analysis from: 9/22/2017 10:25:56 AM
File: _2017-09-22_

Last save: 10/10/2017 9:45:22 AM

Method: ANIONS 09-18-17.mtw
Run operator: wet
Analysis number: 59124

Last save: 9/21/2017 12:16:58 PM

SAMPLE:
: AK/AP
Vial number: 4
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 10/10/2017 9:45:39 AM
Printed by: wet


Ident: LB90688BSW
Analysis from: 9/22/2017 10:55:51 AM
File: _2017-09-22_

Last save: 10/10/2017 9:43:42 AM

Method: ANIONS 09-18-17.mtw
Run operator: wet
Analysis number: 59125

Last save: 9/21/2017 12:16:58 PM

SAMPLE:
: AK/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.78	0.325	1.84	10.62	44.215	7.79	0.
2	8.56	0.326	2.00	11.51	44.254	7.80	0.
3	10.15	0.356	4.21	24.21	102.516	18.06	0.
4	12.79	0.525	2.00	11.53	67.784	11.94	0.
5	14.44	0.477	2.88	16.59	90.722	15.98	0.
6	21.26	0.971	0.91	5.21	57.045	10.05	0.
7	23.93	0.668	3.53	20.31	161.118	28.38	0.
7	27.00	0.521	17.37	99.99	567.654	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 10/10/2017 9:48:17 AM
 Printed by: wet


Ident: I5334-01
 Analysis from: 9/22/2017 11:26:18 AM
 File: _2017-09-22_

Last save: 10/10/2017 9:47:12 AM

Method: ANIONS 09-18-17.mtw
 Run operator: wet
 Analysis number: 59126

Last save: 9/21/2017 12:16:58 PM

SAMPLE:
 : AK/AP
 Vial number: 6
 Volume: 20.0 µL
 Dilution: 1.00
 Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.75	0.336	3.02	1.73	67.635	1.63	0.
2	9.03	0.328	150.57	86.05	3258.669	78.73	0.
3	0.00	0.000	0.00	0.00	0.000	0.00	0.
4	0.00	0.000	0.00	0.00	0.000	0.00	0.
5	14.74	0.463	10.69	6.11	328.472	7.94	0.
6	21.74	0.918	0.44	0.25	25.026	0.60	0.
7	24.31	0.663	9.89	5.65	455.406	11.00	0.
7	27.00	0.387	174.61	99.78	4135.209	99.91	0.

This report has been created by IC Net
 METROHM LTD

Report date: 10/10/2017 9:50:21 AM
Printed by: wet


Ident: I5334-01DLX2
Analysis from: 9/22/2017 12:56:03 PM
File: _2017-09-22_

Last save: 10/10/2017 9:50:18 AM

Method: ANIONS 09-18-17.mtw
Run operator: wet
Analysis number: 59129

Last save: 9/21/2017 12:16:58 PM

SAMPLE:
: AK/AP
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.97	0.323	1.39	1.62	29.623	1.53	0.
2	8.80	0.303	74.19	86.61	1521.622	78.45	0.
3	0.00	0.000	0.00	0.00	0.000	0.00	0.
4	0.00	0.000	0.00	0.00	0.000	0.00	0.
5	14.57	0.455	5.12	5.97	156.648	8.08	0.
6	21.63	0.983	0.23	0.27	14.159	0.73	0.
7	24.28	0.651	4.72	5.51	217.462	11.21	0.
7	27.00	0.388	85.65	99.99	1939.515	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 10/10/2017 9:51:03 AM
Printed by: wet


Ident: I5334-01DLX50
Analysis from: 9/22/2017 1:25:58 PM
File: _2017-09-22_

Last save: 10/10/2017 9:51:04 AM

Method: ANIONS 09-18-17.mtw
Run operator: wet
Analysis number: 59130

Last save: 9/21/2017 12:16:58 PM

SAMPLE:
: AK/AP
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	0.00	0.000	0.00	0.00	0.000	0.00	0.
2	8.79	0.325	2.04	85.25	45.344	76.40	0.
3	0.00	0.000	0.00	0.00	0.000	0.00	0.
4	0.00	0.000	0.00	0.00	0.000	0.00	0.
5	14.86	0.478	0.17	7.22	5.356	9.03	0.
6	0.00	0.000	0.00	0.00	0.000	0.00	0.
7	24.38	0.693	0.18	7.45	8.648	14.57	0.
7	27.00	0.214	2.39	99.92	59.348	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 10/10/2017 9:53:12 AM
Printed by: wet


Ident: CCV
Analysis from: 9/22/2017 1:55:53 PM
File: _2017-09-22_

Last save: 10/10/2017 9:53:04 AM

Method: ANIONS 09-18-17.mtw
Run operator: wet
Analysis number: 59131

Last save: 9/21/2017 12:16:58 PM

SAMPLE:
: AK/AP
Vial number: 2
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height uS/cm	Height %	Area uS/cm*sec	Area %	
1	5.94	0.332	1.83	10.59	43.883	7.85	0.
2	8.73	0.321	2.01	11.64	44.118	7.89	0.
3	10.32	0.349	4.25	24.60	103.336	18.48	0.
4	12.95	0.549	1.93	11.15	67.498	12.07	0.
5	14.59	0.463	2.90	16.78	89.615	16.02	0.
6	21.49	0.924	0.85	4.93	51.355	9.18	0.
7	24.18	0.650	3.51	20.30	159.485	28.52	0.
7	27.00	0.513	17.29	99.99	559.291	100.00	0.

This report has been created by IC Net
METROHM LTD

Report date: 10/10/2017 9:53:19 AM
Printed by: wet


Ident: CCB
Analysis from: 9/22/2017 2:25:48 PM
File: _2017-09-22_

Last save: 10/10/2017 9:53:04 AM

Method: ANIONS 09-18-17.mtw
Run operator: wet
Analysis number: 59132

Last save: 9/21/2017 12:16:58 PM

SAMPLE:
: AK/AP
Vial number: 3
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD