

=====

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

11/1/2017 13:51

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	95.916	0.0	0.103	
ICB1	-1.484	0.0	0.003	
CCV1	255.728	0.0	0.266	
CCB1	-1.489	0.0	0.003	
PB103797BL	-1.627	0.0	0.003	
PB103797BS	97.678	0.0	0.105	
MIDPB103797	248.620	0.0	0.259	
I6208-01	-0.923	0.0	0.004	
I6208-01DUP	-0.991	0.0	0.004	
I6208-01MS	34.092	0.0	0.040	
I6208-01MSD	35.787	0.0	0.041	
CCV2	260.010	0.0	0.271	
CCB2	-1.554	0.0	0.003	

N	13
Mean	78.443
SD	106.4643
CV%	135.72

AS

Aquakem v. 7.2AQ1

Results from time period:

Wed Nov 01 13:37:06 2017

Wed Nov 01 13:40:59 2017

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-1.6764	µg/l	11/1/2017 11:27:07	
5.OPPBCN	A	Total CN	P	4.11	µg/l	11/1/2017 11:27:08	
10PPBCN	A	Total CN	P	8.8751	µg/l	11/1/2017 11:27:09	
50PPBCN	A	Total CN	P	53.5542	µg/l	11/1/2017 11:27:10	
100PPBCN	A	Total CN	P	98.1343	µg/l	11/1/2017 11:27:11	
250PPBCN	A	Total CN	P	253.9074	µg/l	11/1/2017 11:27:12	
500PPBCN	A	Total CN	P	498.0954	µg/l	11/1/2017 11:27:13	
ICV1	S	Total CN	P	95.9156	µg/l	11/1/2017 13:37:07	
ICB1	S	Total CN	P	-1.4843	µg/l	11/1/2017 13:37:08	
CCV1	S	Total CN	P	255.7281	µg/l	11/1/2017 13:37:09	
CCB1	S	Total CN	P	-1.4887	µg/l	11/1/2017 13:37:10	
PB103797BL	S	Total CN	P	-1.6273	µg/l	11/1/2017 13:37:11	
PB103797BS	S	Total CN	P	97.6777	µg/l	11/1/2017 13:37:12	
MIDPB103797	S	Total CN	P	248.6204	µg/l	11/1/2017 13:37:13	
I6208-01	S	Total CN	P	-0.9226	µg/l	11/1/2017 13:37:14	
I6208-01DUP	S	Total CN	P	-0.9911	µg/l	11/1/2017 13:37:15	
I6208-01MS	S	Total CN	P	34.0924	µg/l	11/1/2017 13:37:16	
I6208-01MSD	S	Total CN	P	35.7874	µg/l	11/1/2017 13:40:57	
CCV2	S	Total CN	P	260.0104	µg/l	11/1/2017 13:40:58	
CCB2	S	Total CN	P	-1.5544	µg/l	11/1/2017 13:40:59	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

11/1/2017 11:32


Test Total CN

Accepted 11/1/2017 11:32

Factor 978.8
Bias 0.005

Coeff. of det. 0.999806

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-1.6764	0.0000	
2	5.0PPBCN	0.009	4.1100	5.0000	
3	10PPBCN	0.014	8.8751	10.0000	
4	50PPBCN	0.060	53.5542	50.0000	
5	100PPBCN	0.105	98.1343	100.0000	
6	250PPBCN	0.264	253.9074	250.0000	
7	500PPBCN	0.514	498.0954	500.0000	