

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

11/22/2017 16:08

Test: ^{Reactive} ~~Total~~ CN Mg/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.133	0.0	0.094	
ICB1	1.135	0.0	0.004	
CCV1	253.178	0.0	0.243	
CCB1	1.412	0.0	0.005	
PB104494BL	0.837	0.0	0.004	
I6289-08	0.285	0.0	0.004	
I6289-08DUP	0.213	0.0	0.003	
CCV2	262.277	0.0	0.252	
CCB2	0.884	0.0	0.004	

N	9
Mean	68.484
SD	111.7611
CV%	163.19

AS

Aquakem v. 7.2AQ1

Results from time period:

Wed Nov 22 15:20:28 2017

Wed Nov 22 15:23:52 2017

Sample Id	Sam/Ctr/c	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	0.4627	µg/l	11/22/2017 13:16:33	
5.OPPBCN	A	Reactive CN	P	4.8029	µg/l	11/22/2017 13:16:34	
10PPBCN	A	Reactive CN	P	9.4102	µg/l	11/22/2017 13:16:35	
50PPBCN	A	Reactive CN	P	49.6697	µg/l	11/22/2017 13:16:36	
100PPBCN	A	Reactive CN	P	97.3915	µg/l	11/22/2017 13:16:37	
250PPBCN	A	Reactive CN	P	255.3891	µg/l	11/22/2017 13:16:38	
500PPBCN	A	Reactive CN	P	497.874	µg/l	11/22/2017 13:16:39	
ICV1	S	Reactive CN	P	96.1328	µg/l	11/22/2017 15:20:29	
ICB1	S	Reactive CN	P	1.1345	µg/l	11/22/2017 15:20:30	
CCV1	S	Reactive CN	P	253.1781	µg/l	11/22/2017 15:20:33	
CCB1	S	Reactive CN	P	1.4123	µg/l	11/22/2017 15:20:34	
PB104494BL	S	Reactive CN	P	0.8368	µg/l	11/22/2017 15:20:36	
I6289-08	S	Reactive CN	P	0.2855	µg/l	11/22/2017 15:20:37	
I6289-08DUP	S	Reactive CN	P	0.2133	µg/l	11/22/2017 15:20:38	
CCV2	S	Reactive CN	P	262.2766	µg/l	11/22/2017 15:23:51	
CCB2	S	Reactive CN	P	0.8835	µg/l	11/22/2017 15:23:52	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

11/22/2017 13:46


Test Total CN

Accepted 11/22/2017 13:46

Factor 1056
Bias 0.003

Coeff. of det. 0.999800

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	0.4627	0.0000	
2	5.0PPBCN	0.008	4.8029	5.0000	
3	10PPBCN	0.012	9.4102	10.0000	
4	50PPBCN	0.050	49.6697	50.0000	
5	100PPBCN	0.095	97.3915	100.0000	
6	250PPBCN	0.245	255.3891	250.0000	
7	500PPBCN	0.475	497.8740	500.0000	