

=====

Test results

Aquakem 7.2AQ1

1391862
Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

11/30/2017 13:39

Test: Total CN My bl.

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	100.711	0.0	0.096	
ICB1	-0.036	0.0	0.004	
CCV1	250.110	0.0	0.234	
CCB1	0.341	0.0	0.004	
PB104619BL	-0.051	0.0	0.004	
PB104619BS	100.573	0.0	0.096	
LOWPB104619	9.600	0.0	0.013	
HIGHPB104619	496.009	0.0	0.460	
I6289-09	0.860	0.0	0.005	
I6289-09DUP	0.863	0.0	0.005	
I6289-09MS	33.935	0.0	0.035	
I6289-09MSD	33.978	0.0	0.035	
I6300-07	0.498	0.0	0.004	
I6639-01	0.652	0.0	0.004	
CCV2	247.052	0.0	0.231	
CCB2	-0.129	0.0	0.004	
I6641-01	-0.074	0.0	0.004	
I6641-03	0.308	0.0	0.004	
I6641-05	0.183	0.0	0.004	
CCV3	257.540	0.0	0.241	
CCB3	0.104	0.0	0.004	

N	21
Mean	73.001
SD	131.6599
CV%	180.35

As

Aquakem v. 7.2AQ1

Results from time period:

Thu Nov 30 13:19:12 2017

Thu Nov 30 13:32:33 2017

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.5055	µg/l	11/30/2017 8:50:56	
5.OPPBCN	A	Total CN	P	4.4389	µg/l	11/30/2017 8:50:57	
10PPBCN	A	Total CN	P	9.3682	µg/l	11/30/2017 8:50:58	
50PPBCN	A	Total CN	P	50.3313	µg/l	11/30/2017 8:50:59	
100PPBCN	A	Total CN	P	100.4142	µg/l	11/30/2017 8:51:00	
250PPBCN	A	Total CN	P	252.1012	µg/l	11/30/2017 8:51:01	
500PPBCN	A	Total CN	P	498.8517	µg/l	11/30/2017 8:51:02	
ICV1	S	Total CN	P	100.7106	µg/l	11/30/2017 13:19:13	
ICB1	S	Total CN	P	-0.0355	µg/l	11/30/2017 13:19:14	
CCV1	S	Total CN	P	250.1098	µg/l	11/30/2017 13:19:17	
CCB1	S	Total CN	P	0.3414	µg/l	11/30/2017 13:19:18	
PB104619BL	S	Total CN	P	-0.0512	µg/l	11/30/2017 13:19:20	
PB104619BS	S	Total CN	P	100.5726	µg/l	11/30/2017 13:19:21	
LOWPB104619	S	Total CN	P	9.6003	µg/l	11/30/2017 13:19:22	
HIGHPB104619	S	Total CN	P	496.0088	µg/l	11/30/2017 13:26:44	
I6289-09	S	Total CN	P	0.8597	µg/l	11/30/2017 13:26:45	
I6289-09DUP	S	Total CN	P	0.863	µg/l	11/30/2017 13:26:46	
I6289-09MS	S	Total CN	P	33.9346	µg/l	11/30/2017 13:26:47	
I6289-09MSD	S	Total CN	P	33.9775	µg/l	11/30/2017 13:26:48	
I6300-07	S	Total CN	P	0.4985	µg/l	11/30/2017 13:26:49	
I6639-01	S	Total CN	P	0.6521	µg/l	11/30/2017 13:26:50	
CCV2	S	Total CN	P	247.0517	µg/l	11/30/2017 13:26:51	
CCB2	S	Total CN	P	-0.1286	µg/l	11/30/2017 13:26:52	
I6641-01	S	Total CN	P	-0.0741	µg/l	11/30/2017 13:26:53	
I6641-03	S	Total CN	P	0.308	µg/l	11/30/2017 13:26:54	
I6641-05	S	Total CN	P	0.1828	µg/l	11/30/2017 13:32:27	
CCV3	S	Total CN	P	257.5405	µg/l	11/30/2017 13:32:28	
CCB3	S	Total CN	P	0.1035	µg/l	11/30/2017 13:32:29	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS.

11/30/2017 9:01


Test Total CN

Accepted 11/30/2017 9:01

Factor 1087
Bias 0.004

Coeff. of det. 0.999966

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.5055	0.0000	
2	5.0PPBCN	0.008	4.4389	5.0000	
3	10PPBCN	0.012	9.3682	10.0000	
4	50PPBCN	0.050	50.3313	50.0000	
5	100PPBCN	0.096	100.4142	100.0000	
6	250PPBCN	0.236	252.1012	250.0000	
7	500PPBCN	0.463	498.8517	500.0000	