

=====

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/7/2017 9:43

Test: ^{Reactive} ~~Total~~ CN Mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.546	0.0	0.093	
ICB1	-0.077	0.0	0.004	
CCV1	245.221	0.0	0.229	
CCB1	0.114	0.0	0.004	
PB104774BL	-0.414	0.0	0.003	
I6303-02	-0.604	0.0	0.003	
I6303-02DUP	-0.636	0.0	0.003	
I6675-02	-0.435	0.0	0.003	
I6676-02	-0.607	0.0	0.003	
I6683-02	-0.531	0.0	0.003	
I6683-04	-0.498	0.0	0.003	
I6683-06	0.236	0.0	0.004	
I6685-04	-0.305	0.0	0.004	
CCV2	246.651	0.0	0.231	
CCB2	-0.049	0.0	0.004	
I6738-03	-0.442	0.0	0.003	
I6741-02	-0.413	0.0	0.003	
PB104775BL	-0.325	0.0	0.004	
I6742-02	-0.486	0.0	0.003	
I6742-02DUP	-0.607	0.0	0.003	
I6685-02	-0.453	0.0	0.003	
CCV3	260.592	0.0	0.244	
CCB3	-0.232	0.0	0.004	

N	23
Mean	36.619
SD	87.2196
CV%	238.18

AS

Aquakem v. 7.2AQ1

Results from time period:

Thu Dec 07 08:38:33 2017

Thu Dec 07 09:40:40 2017

Sample Id	Sam/Ctr/c	Test short	nan	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive	CN	P	-0.5909	µg/l	12/7/2017 8:38:33	
5.0PPBCN	A	Reactive	CN	P	4.4002	µg/l	12/7/2017 8:38:34	
10PPBCN	A	Reactive	CN	P	9.2635	µg/l	12/7/2017 8:38:35	
50PPBCN	A	Reactive	CN	P	49.4812	µg/l	12/7/2017 8:38:36	
100PPBCN	A	Reactive	CN	P	100.7046	µg/l	12/7/2017 8:38:37	
250PPBCN	A	Reactive	CN	P	253.6195	µg/l	12/7/2017 8:38:38	
500PPBCN	A	Reactive	CN	P	498.1219	µg/l	12/7/2017 8:38:39	
ICV1	S	Reactive	CN	P	96.5459	µg/l	12/7/2017 9:10:36	
ICB1	S	Reactive	CN	P	-0.0767	µg/l	12/7/2017 9:10:38	
CCV1	S	Reactive	CN	P	245.2206	µg/l	12/7/2017 9:10:40	
CCB1	S	Reactive	CN	P	0.1143	µg/l	12/7/2017 9:10:41	
PB104774BL	S	Reactive	CN	P	-0.4142	µg/l	12/7/2017 9:10:43	
I6303-02	S	Reactive	CN	P	-0.6037	µg/l	12/7/2017 9:10:44	
I6303-02DUP	S	Reactive	CN	P	-0.6362	µg/l	12/7/2017 9:10:45	
I6675-02	S	Reactive	CN	P	-0.4346	µg/l	12/7/2017 9:18:07	
I6676-02	S	Reactive	CN	P	-0.6071	µg/l	12/7/2017 9:18:08	
I6683-02	S	Reactive	CN	P	-0.5306	µg/l	12/7/2017 9:18:09	
I6683-04	S	Reactive	CN	P	-0.4984	µg/l	12/7/2017 9:18:10	
I6683-06	S	Reactive	CN	P	0.2363	µg/l	12/7/2017 9:18:11	
I6685-04	S	Reactive	CN	P	-0.3049	µg/l	12/7/2017 9:18:13	
CCV2	S	Reactive	CN	P	246.6506	µg/l	12/7/2017 9:18:14	
CCB2	S	Reactive	CN	P	-0.0487	µg/l	12/7/2017 9:18:15	
I6738-03	S	Reactive	CN	P	-0.4416	µg/l	12/7/2017 9:18:16	
I6741-02	S	Reactive	CN	P	-0.4133	µg/l	12/7/2017 9:18:17	
PB104775BL	S	Reactive	CN	P	-0.3247	µg/l	12/7/2017 9:22:54	
I6742-02	S	Reactive	CN	P	-0.4862	µg/l	12/7/2017 9:22:55	
I6742-02DUP	S	Reactive	CN	P	-0.6067	µg/l	12/7/2017 9:22:56	
I6685-02	S	Reactive	CN	P	-0.4526	µg/l	12/7/2017 9:40:38	
CCV3	S	Reactive	CN	P	260.5917	µg/l	12/7/2017 9:40:39	
CCB3	S	Reactive	CN	P	-0.232	µg/l	12/7/2017 9:40:40	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/7/2017 8:40


Test Total CN

Accepted 12/7/2017 8:39

Factor 1087
Bias 0.004

Coeff. of det. 0.999909

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.5909	0.0000	
2	5.0PPBCN	0.008	4.4002	5.0000	
3	10PPBCN	0.012	9.2635	10.0000	
4	50PPBCN	0.049	49.4812	50.0000	
5	100PPBCN	0.096	100.7046	100.0000	
6	250PPBCN	0.237	253.6195	250.0000	
7	500PPBCN	0.462	498.1219	500.0000	