

===== 1092079
Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS12/11/2017 14:40
-----Test: Total CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.813	0.0	0.094	
ICB1	1.206	0.0	0.004	
CCV1	243.206	0.0	0.231	
CCB1	1.533	0.0	0.004	
PB104902BL	1.101	0.0	0.004	
PB104902BS	98.837	0.0	0.095	
I6813-01	2.979	0.0	0.006	
I6813-01DUP	2.924	0.0	0.005	
I6813-01MS	31.738	0.0	0.032	
I6813-01MSD	32.048	0.0	0.033	
I6814-01	31.038	0.0	0.032	
PB104903BL	1.115	0.0	0.004	
PB104903BS	98.825	0.0	0.095	
I6813-03	2.865	0.0	0.005	
CCV2	241.315	0.0	0.229	
CCB2	18.395	0.0	0.020	
I6813-03DUP	2.847	0.0	0.005	
I6813-03MS	25.762	0.0	0.027	
I6813-03MSD	26.957	0.0	0.028	
CCV3	254.839	0.0	0.242	
CCB3	1.410	0.0	0.004	

N	21
Mean	58.036
SD	85.2833
CV%	146.95

AS

Aquakem v. 7.2AQ1

Results from time period:

Mon Dec 11 14:03:05 2017

Mon Dec 11 14:37:35 2017

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	0.4779	µg/l	12/11/2017 10:52:08	
5.OPPBCN	A	Total CN	P	5.6312	µg/l	12/11/2017 10:52:09	
10PPBCN	A	Total CN	P	10.5573	µg/l	12/11/2017 10:52:10	
50PPBCN	A	Total CN	P	50.0806	µg/l	12/11/2017 10:52:11	
100PPBCN	A	Total CN	P	99.3935	µg/l	12/11/2017 10:52:12	
250PPBCN	A	Total CN	P	247.5273	µg/l	12/11/2017 10:52:13	
500PPBCN	A	Total CN	P	501.3321	µg/l	12/11/2017 10:52:14	
ICV1	S	Total CN	P	97.8134	µg/l	12/11/2017 14:03:06	
ICB1	S	Total CN	P	1.2058	µg/l	12/11/2017 14:03:07	
CCV1	S	Total CN	P	243.2063	µg/l	12/11/2017 14:03:10	
CCB1	S	Total CN	P	1.533	µg/l	12/11/2017 14:03:11	
PB104902BL	S	Total CN	P	1.1006	µg/l	12/11/2017 14:03:13	
PB104902BS	S	Total CN	P	98.8367	µg/l	12/11/2017 14:03:14	
I6813-01	S	Total CN	P	2.9788	µg/l	12/11/2017 14:03:15	
I6813-01DUP	S	Total CN	P	2.9241	µg/l	12/11/2017 14:10:36	
I6813-01MS	S	Total CN	P	31.7385	µg/l	12/11/2017 14:10:37	
I6813-01MSD	S	Total CN	P	32.0477	µg/l	12/11/2017 14:10:38	
I6814-01	S	Total CN	P	31.0385	µg/l	12/11/2017 14:10:39	
PB104903BL	S	Total CN	P	1.1149	µg/l	12/11/2017 14:10:40	
PB104903BS	S	Total CN	P	98.825	µg/l	12/11/2017 14:10:41	
I6813-03	S	Total CN	P	2.8649	µg/l	12/11/2017 14:10:42	
CCV2	S	Total CN	P	241.3154	µg/l	12/11/2017 14:10:43	
CCB2	S	Total CN	P	18.3953	µg/l	12/11/2017 14:10:44	
I6813-03DUP	S	Total CN	P	2.8472	µg/l	12/11/2017 14:10:45	
I6813-03MS	S	Total CN	P	25.7624	µg/l	12/11/2017 14:10:46	
I6813-03MSD	S	Total CN	P	26.9567	µg/l	12/11/2017 14:15:23	
CCV3	S	Total CN	P	254.8385	µg/l	12/11/2017 14:15:24	
CCB3	S	Total CN	P	1.41	µg/l	12/11/2017 14:15:25	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/11/2017 10:52


Test Total CN

Accepted 12/11/2017 10:52

Factor 1067
Bias 0.003

Coeff. of det. 0.999955

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.4779	0.0000	
2	5.0PPBCN	0.008	5.6312	5.0000	
3	10PPBCN	0.013	10.5573	10.0000	
4	50PPBCN	0.050	50.0806	50.0000	
5	100PPBCN	0.096	99.3935	100.0000	
6	250PPBCN	0.235	247.5273	250.0000	
7	500PPBCN	0.473	501.3321	500.0000	