

Chemtech Consulting Group

Analytical Review Report

Date Printed : 12/15/17

Analyst : *AK*

Data File : lb92131.csv

Approved By : *AK*

Approved Date : *12/15/17*

Worksheet # :

m9056 - A - Inorganic Anions - 18.

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1	
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units	Line 2
Chloride												
STD1	STD1				W	12/8/17						
Chloride		PASS	0.000		0.000						mg/L	
STD2	STD2				W	12/8/17						
Chloride		PASS	0.238		0.238						mg/L	
STD3	STD3				W	12/8/17						
Chloride		PASS	0.631		0.631						mg/L	
STD4	STD4				W	12/8/17						
Chloride		PASS	1.206		1.21						mg/L	
STD5	STD5				W	12/8/17						
Chloride		PASS	2.927		2.93						mg/L	
STD6	STD6				W	12/8/17						
Chloride		PASS	5.989		5.99						mg/L	
STD7	STD7				W	12/8/17						
Chloride		PASS	7.575		7.58						mg/L	
ICV1	ICV1				W	12/8/17						
Chloride		PASS	2.882		2.88	96.0	90	110			mg/L	
ICB1	ICB1				W	12/8/17						
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L	
CCV1	CCV1				W	12/13/17						
Chloride		PASS	2.950		2.95	98.0	90	110			mg/L	
CCB1	CCB1				W	12/13/17						
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L	
LB92131BLS	LB92131BLS				S	12/13/17						
Chloride		PASS	0.000		0.000			+/-0.0320			mg/Kg	
LB92131BSS	LB92131BSS				S	12/13/17						
Chloride		PASS	2.931		58.50	98.0	90.00	110.00			mg/Kg	
I6886-05	SEEP-DIRT-5			1	S	12/13/17						
Chloride		PASS	0.484		11.2						mg/Kg	
I6886-05D	SEEP-DIRT-5D			1	S	12/13/17						
Chloride		PASS	0.473		10.90				2.7	20	mg/Kg	
I6886-05S	SEEP-DIRT-5S			1	S	12/13/17						
Chloride		PASS	2.980		68.7	83.0	80	120			mg/Kg	
I6886-05SD	SEEP-DIRT-5SD			1	S	12/13/17						
Chloride		PASS	2.966		68.2	83.0	80	120	0.7	15	mg/Kg	
CCV2	CCV2				W	12/13/17						
Chloride		PASS	2.941		2.94	98.0	90	110			mg/L	
CCB2	CCB2				W	12/13/17						
Chloride		PASS	0.000		0.000			+/-0.0750			mg/L	


Chemtech Consulting Group


Analytical Review Report


Date Printed : 12/15/17

Analyst : 

Data File : lb92131.csv

Approved By : 

Approved Date : 

Worksheet # : 

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis			Line 1
Parameter			PPB		Final Conc	%Rec	LCL	UCL	RPD	Max RPD	Units
Fluoride											
STD1	STD1				W	12/8/17					
Fluoride		PASS	0.000		0.000						mg/L
STD2	STD2				W	12/8/17					
Fluoride		PASS	0.146		0.146						mg/L
STD3	STD3				W	12/8/17					
Fluoride		PASS	0.383		0.383						mg/L
STD4	STD4				W	12/8/17					
Fluoride		PASS	0.729		0.729						mg/L
STD5	STD5				W	12/8/17					
Fluoride		PASS	1.977		1.98						mg/L
STD6	STD6				W	12/8/17					
Fluoride		PASS	4.021		4.02						mg/L
STD7	STD7				W	12/8/17					
Fluoride		PASS	5.032		5.03						mg/L
ICV1	ICV1				W	12/8/17					
Fluoride		PASS	1.848		1.85	93.0	90	110			mg/L
ICB1	ICB1				W	12/8/17					
Fluoride		PASS	0.000		0.000			+/-0.0430			mg/L
CCV1	CCV1				W	12/13/17					
Fluoride		PASS	1.910		1.91	96.0	90	110			mg/L
CCB1	CCB1				W	12/13/17					
Fluoride		PASS	0.000		0.000			+/-0.0430			mg/L
LB92131BLS	LB92131BLS				S	12/13/17					
Fluoride		PASS	0.000		0.000			+/-0.0170			mg/Kg
LB92131BSS	LB92131BSS				S	12/13/17					
Fluoride		PASS	1.844		36.80	92.0	90.00	110.00			mg/Kg
I6886-05	SEEP-DIRT-5			1	S	12/13/17					
Fluoride		PASS	0.000		0.000						mg/Kg
I6886-05D	SEEP-DIRT-5D			1	S	12/13/17					
Fluoride		PASS	0.000		0.00				0	20	mg/Kg
I6886-05S	SEEP-DIRT-5S			1	S	12/13/17					
Fluoride		PASS	1.882		43.4	94.0	80	120			mg/Kg
I6886-05SD	SEEP-DIRT-5SD			1	S	12/13/17					
Fluoride		PASS	1.808		41.6	90.0	80	120	4.2	15	mg/Kg
CCV2	CCV2				W	12/13/17					
Fluoride		PASS	1.972		1.97	99.0	90	110			mg/L
CCB2	CCB2				W	12/13/17					
Fluoride		PASS	0.000		0.000			+/-0.0430			mg/L