

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/13/2017 13:24

Test: Total CN

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.884	0.0	0.092	
ICB1	0.975	0.0	0.004	
CCV1	242.485	0.0	0.225	
CCB1	0.954	0.0	0.003	
PB104993BL	0.855	0.0	0.003	
I6885-03	0.527	0.0	0.003	
I6885-03DUP	0.591	0.0	0.003	
CCV2	257.645	0.0	0.239	
CCB2	1.019	0.0	0.004	

N	9
Mean	66.882
SD	108.5771
CV%	162.34

AS

Aquakem v. 7.2AQ1

Results from time period:

Wed Dec 13 13:18:48 2017

Wed Dec 13 13:22:12 2017

Sample Id	Sam/Ctr/c	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	0.1803	µg/l	12/13/2017 11:02:10	
5.OPPBCN	A	Reactive CN	P	5.2866	µg/l	12/13/2017 11:02:11	
10PPBCN	A	Reactive CN	P	10.3898	µg/l	12/13/2017 11:02:12	
50PPBCN	A	Reactive CN	P	50.6473	µg/l	12/13/2017 11:02:13	
100PPBCN	A	Reactive CN	P	99.2472	µg/l	12/13/2017 11:02:14	
250PPBCN	A	Reactive CN	P	248.3474	µg/l	12/13/2017 11:02:15	
500PPBCN	A	Reactive CN	P	500.9015	µg/l	12/13/2017 11:02:16	
ICV1	S	Reactive CN	P	96.8842	µg/l	12/13/2017 13:18:49	
ICB1	S	Reactive CN	P	0.9749	µg/l	12/13/2017 13:18:51	
CCV1	S	Reactive CN	P	242.4847	µg/l	12/13/2017 13:18:53	
CCB1	S	Reactive CN	P	0.9544	µg/l	12/13/2017 13:18:55	
PB104993BL	S	Reactive CN	P	0.8549	µg/l	12/13/2017 13:18:56	
I6885-03	S	Reactive CN	P	0.5274	µg/l	12/13/2017 13:18:57	
I6885-03DUP	S	Reactive CN	P	0.5909	µg/l	12/13/2017 13:18:58	
CCV2	S	Reactive CN	P	257.6447	µg/l	12/13/2017 13:22:11	
CCB2	S	Reactive CN	P	1.019	µg/l	12/13/2017 13:22:12	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/13/2017 11:03


Test Total CN

Accepted 12/13/2017 11:03

Factor 1090
Bias 0.003

Coeff. of det. 0.999977

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.1803	0.0000	
2	5.0PPBCN	0.007	5.2866	5.0000	
3	10PPBCN	0.012	10.3898	10.0000	
4	50PPBCN	0.049	50.6473	50.0000	
5	100PPBCN	0.094	99.2472	100.0000	
6	250PPBCN	0.231	248.3474	250.0000	
7	500PPBCN	0.462	500.9015	500.0000	