

1092186

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/15/2017 13:01

Test: *Reactive* ~~Total~~ CN *ug/L*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.492	0.0	0.094	
ICB1	-0.008	0.0	0.004	
CCV1	245.536	0.0	0.233	
CCB1	0.186	0.0	0.004	
PB105066BL	-0.032	0.0	0.004	
I6680-12	-0.144	0.0	0.004	
I6680-12DUP	-0.116	0.0	0.004	
I6680-14	-0.321	0.0	0.003	
I6680-16	-0.402	0.0	0.003	
I6680-18	-0.331	0.0	0.003	
I6744-13	-0.334	0.0	0.003	
I6911-02	-0.158	0.0	0.004	
I6919-02	-0.072	0.0	0.004	
CCV2	250.473	0.0	0.238	
CCB2	-0.132	0.0	0.004	

N	15
Mean	39.376
SD	88.2686
CV%	224.17

AS

Aquakem v. 7.2AQ1

Results from time period:

Fri Dec 15 12:41:49 2017

Fri Dec 15 12:47:38 2017

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.0912	µg/l	12/15/2017 11:19:40	
5.OPPBCN	A	Reactive CN	P	4.8553	µg/l	12/15/2017 11:19:41	
10PPBCN	A	Reactive CN	P	9.6182	µg/l	12/15/2017 11:19:42	
50PPBCN	A	Reactive CN	P	50.6021	µg/l	12/15/2017 11:19:43	
100PPBCN	A	Reactive CN	P	99.3818	µg/l	12/15/2017 11:19:44	
250PPBCN	A	Reactive CN	P	251.1224	µg/l	12/15/2017 11:19:45	
500PPBCN	A	Reactive CN	P	499.5113	µg/l	12/15/2017 11:19:46	
ICV1	S	Reactive CN	P	96.4915	µg/l	12/15/2017 12:41:50	
ICB1	S	Reactive CN	P	-0.008	µg/l	12/15/2017 12:41:51	
CCV1	S	Reactive CN	P	245.5359	µg/l	12/15/2017 12:41:54	
CCB1	S	Reactive CN	P	0.1858	µg/l	12/15/2017 12:41:55	
PB105066BL	S	Reactive CN	P	-0.032	µg/l	12/15/2017 12:41:56	
I6680-12	S	Reactive CN	P	-0.1441	µg/l	12/15/2017 12:41:57	
I6680-12DUP	S	Reactive CN	P	-0.116	µg/l	12/15/2017 12:41:58	
I6680-14	S	Reactive CN	P	-0.3212	µg/l	12/15/2017 12:41:59	
I6680-16	S	Reactive CN	P	-0.4015	µg/l	12/15/2017 12:47:32	
I6680-18	S	Reactive CN	P	-0.3309	µg/l	12/15/2017 12:47:33	
I6744-13	S	Reactive CN	P	-0.334	µg/l	12/15/2017 12:47:34	
I6911-02	S	Reactive CN	P	-0.1579	µg/l	12/15/2017 12:47:35	
I6919-02	S	Reactive CN	P	-0.0721	µg/l	12/15/2017 12:47:36	
CCV2	S	Reactive CN	P	250.4727	µg/l	12/15/2017 12:47:37	
CCB2	S	Reactive CN	P	-0.1324	µg/l	12/15/2017 12:47:38	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/15/2017 11:22


Test Total CN

Accepted 12/15/2017 11:22

Factor 1068
Bias 0.004

Coeff. of det. 0.999988

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.0912	0.0000	
2	5.0PPBCN	0.008	4.8553	5.0000	
3	10PPBCN	0.013	9.6182	10.0000	
4	50PPBCN	0.051	50.6021	50.0000	
5	100PPBCN	0.097	99.3818	100.0000	
6	250PPBCN	0.239	251.1224	250.0000	
7	500PPBCN	0.471	499.5113	500.0000	