

LB92192

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/15/2017 14:29

Test: Total CN Mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	96.641	0.0	0.094	
ICB1	0.470	0.0	0.004	
CCV1	250.304	0.0	0.238	
CCB1	0.262	0.0	0.004	
PB105078BL	0.180	0.0	0.004	
PB105078BS	98.793	0.0	0.096	
I6676-05	1.347	0.0	0.005	
I6684-03	0.276	0.0	0.004	
I6684-03DUP	0.330	0.0	0.004	
I6684-03MS	29.699	0.0	0.031	
I6684-03MSD	29.918	0.0	0.032	
CCV2	248.775	0.0	0.237	
CCB2	0.349	0.0	0.004	

N 13
Mean 58.257
SD 91.8783
CV% 157.71

AS

Aquakem v. 7.2AQ1

Results from time period:

Fri Dec 15 14:21:41 2017

Fri Dec 15 14:27:58 2017

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Total CN	P	-0.0912	µg/l	12/15/2017 11:19:40	
5.OPPBCN	A	Total CN	P	4.8553	µg/l	12/15/2017 11:19:41	
10PPBCN	A	Total CN	P	9.6182	µg/l	12/15/2017 11:19:42	
50PPBCN	A	Total CN	P	50.6021	µg/l	12/15/2017 11:19:43	
100PPBCN	A	Total CN	P	99.3818	µg/l	12/15/2017 11:19:44	
250PPBCN	A	Total CN	P	251.1224	µg/l	12/15/2017 11:19:45	
500PPBCN	A	Total CN	P	499.5113	µg/l	12/15/2017 11:19:46	
ICV1	S	Total CN	P	96.6413	µg/l	12/15/2017 14:21:42	
ICB1	S	Total CN	P	0.4702	µg/l	12/15/2017 14:21:43	
CCV1	S	Total CN	P	250.3038	µg/l	12/15/2017 14:21:46	
CCB1	S	Total CN	P	0.2615	µg/l	12/15/2017 14:21:47	
PB105078BL	S	Total CN	P	0.1798	µg/l	12/15/2017 14:21:49	
PB105078BS	S	Total CN	P	98.7934	µg/l	12/15/2017 14:21:50	
I6676-05	S	Total CN	P	1.3467	µg/l	12/15/2017 14:21:51	
I6684-03	S	Total CN	P	0.276	µg/l	12/15/2017 14:27:51	
I6684-03DUP	S	Total CN	P	0.3301	µg/l	12/15/2017 14:27:52	
I6684-03MS	S	Total CN	P	29.699	µg/l	12/15/2017 14:27:53	
I6684-03MSD	S	Total CN	P	29.9178	µg/l	12/15/2017 14:27:54	
CCV2	S	Total CN	P	248.7748	µg/l	12/15/2017 14:27:55	
CCB2	S	Total CN	P	0.3492	µg/l	12/15/2017 14:27:56	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

12/15/2017 11:22


Test Total CN

Accepted 12/15/2017 11:22

Factor 1068
 Bias 0.004

Coeff. of det. 0.999988

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.0912	0.0000	
2	5.0PPBCN	0.008	4.8553	5.0000	
3	10PPBCN	0.013	9.6182	10.0000	
4	50PPBCN	0.051	50.6021	50.0000	
5	100PPBCN	0.097	99.3818	100.0000	
6	250PPBCN	0.239	251.1224	250.0000	
7	500PPBCN	0.471	499.5113	500.0000	