

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/15/2017 15:03

Test: *Reactive* Total CN *ug/d*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	97.132	0.0	0.095	
ICB1	0.504	0.0	0.004	
CCV1	247.298	0.0	0.235	
CCB1	0.497	0.0	0.004	
PB105077BL	0.368	0.0	0.004	
I6676-06	-0.151	0.0	0.004	
I6676-06DUP	-0.133	0.0	0.004	
I6684-04	-0.226	0.0	0.003	
CCV2	258.206	0.0	0.245	
CCB2	0.468	0.0	0.004	

N	10
Mean	60.396
SD	105.8223
CV%	175.21

AS

Aquakem v. 7.2AQ1

Results from time period:

Fri Dec 15 14:53:18 2017

Fri Dec 15 14:57:10 2017

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.0912	µg/l	12/15/2017 11:19:40	
5.OPPBCN	A	Reactive CN	P	4.8553	µg/l	12/15/2017 11:19:41	
10PPBCN	A	Reactive CN	P	9.6182	µg/l	12/15/2017 11:19:42	
50PPBCN	A	Reactive CN	P	50.6021	µg/l	12/15/2017 11:19:43	
100PPBCN	A	Reactive CN	P	99.3818	µg/l	12/15/2017 11:19:44	
250PPBCN	A	Reactive CN	P	251.1224	µg/l	12/15/2017 11:19:45	
500PPBCN	A	Reactive CN	P	499.5113	µg/l	12/15/2017 11:19:46	
ICV1	S	Reactive CN	P	97.132	µg/l	12/15/2017 14:53:19	
ICB1	S	Reactive CN	P	0.5038	µg/l	12/15/2017 14:53:20	
CCV1	S	Reactive CN	P	247.2979	µg/l	12/15/2017 14:53:23	
CCB1	S	Reactive CN	P	0.497	µg/l	12/15/2017 14:53:25	
PB105077BL	S	Reactive CN	P	0.3684	µg/l	12/15/2017 14:53:26	
I6676-06	S	Reactive CN	P	-0.1514	µg/l	12/15/2017 14:53:27	
I6676-06DUP	S	Reactive CN	P	-0.1329	µg/l	12/15/2017 14:53:28	
I6684-04	S	Reactive CN	P	-0.2257	µg/l	12/15/2017 14:57:08	
CCV2	S	Reactive CN	P	258.2065	µg/l	12/15/2017 14:57:09	
CCB2	S	Reactive CN	P	0.4676	µg/l	12/15/2017 14:57:10	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/15/2017 11:22


Test Total CN

Accepted 12/15/2017 11:22

Factor 1068
Bias 0.004

Coeff. of det. 0.999988

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.0912	0.0000	
2	5.0PPBCN	0.008	4.8553	5.0000	
3	10PPBCN	0.013	9.6182	10.0000	
4	50PPBCN	0.051	50.6021	50.0000	
5	100PPBCN	0.097	99.3818	100.0000	
6	250PPBCN	0.239	251.1224	250.0000	
7	500PPBCN	0.471	499.5113	500.0000	