

1092312

=====
Test results Aquakem 7.2AQ1 Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/21/2017 13:18

Test: *Reactive* ~~Total~~ CN *mg/l*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	94.269	0.0	0.088	
ICB1	0.861	0.0	0.004	
CCV1	238.404	0.0	0.219	
CCB1	0.846	0.0	0.004	
PB105248BL	0.818	0.0	0.004	
I6683-08	0.367	0.0	0.003	
I6683-08DUP	0.367	0.0	0.003	
I6683-10	0.250	0.0	0.003	
I6683-12	0.545	0.0	0.004	
PB105249BL	0.746	0.0	0.004	
I7007-02	0.302	0.0	0.003	
I7007-02DUP	0.288	0.0	0.003	
CCV2	256.105	0.0	0.235	
CCB2	1.022	0.0	0.004	

N 14
Mean 42.514
SD 90.3045
CV% 212.41

As

Aquakem v. 7.2AQ1

Results from time period:

Thu Dec 21 12:47:59 2017

Thu Dec 21 13:17:26 2017

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	0.0531	µg/l	12/21/2017 11:24:19	
5.OPPBCN	A	Reactive CN	P	4.9224	µg/l	12/21/2017 11:24:20	
10PPBCN	A	Reactive CN	P	9.9559	µg/l	12/21/2017 11:24:21	
50PPBCN	A	Reactive CN	P	50.1915	µg/l	12/21/2017 11:24:22	
100PPBCN	A	Reactive CN	P	99.4255	µg/l	12/21/2017 11:24:23	
250PPBCN	A	Reactive CN	P	250.7083	µg/l	12/21/2017 11:24:24	
500PPBCN	A	Reactive CN	P	499.7433	µg/l	12/21/2017 11:24:25	
ICV1	S	Reactive CN	P	94.2695	µg/l	12/21/2017 12:48:00	
ICB1	S	Reactive CN	P	0.8607	µg/l	12/21/2017 12:48:01	
CCV1	S	Reactive CN	P	238.4042	µg/l	12/21/2017 12:48:04	
CCB1	S	Reactive CN	P	0.8457	µg/l	12/21/2017 12:48:06	
PB105248BL	S	Reactive CN	P	0.8178	µg/l	12/21/2017 12:48:07	
I6683-08	S	Reactive CN	P	0.3669	µg/l	12/21/2017 12:48:08	
I6683-08DUP	S	Reactive CN	P	0.3673	µg/l	12/21/2017 12:48:09	
I6683-10	S	Reactive CN	P	0.2503	µg/l	12/21/2017 12:53:42	
I6683-12	S	Reactive CN	P	0.5447	µg/l	12/21/2017 12:53:43	
PB105249BL	S	Reactive CN	P	0.746	µg/l	12/21/2017 12:53:44	
I7007-02	S	Reactive CN	P	0.3023	µg/l	12/21/2017 13:17:23	
I7007-02DUP	S	Reactive CN	P	0.288	µg/l	12/21/2017 13:17:24	
CCV2	S	Reactive CN	P	256.1048	µg/l	12/21/2017 13:17:25	
CCB2	S	Reactive CN	P	1.0225	µg/l	12/21/2017 13:17:26	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/21/2017 11:33


Test Total CN

Accepted 12/21/2017 11:33

Factor 1105
Bias 0.003

Coeff. of det. 0.999995

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	0.0531	0.0000	
2	5.0PPBCN	0.008	4.9224	5.0000	
3	10PPBCN	0.012	9.9559	10.0000	
4	50PPBCN	0.049	50.1915	50.0000	
5	100PPBCN	0.093	99.4255	100.0000	
6	250PPBCN	0.230	250.7083	250.0000	
7	500PPBCN	0.456	499.7433	500.0000	