

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

12/22/2017 14:25

Test: ^{Reactive} ~~Total~~ CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	98.047	0.0	0.095	
ICB1	1.251	0.0	0.005	
CCV1	248.529	0.0	0.234	
CCB1	0.744	0.0	0.004	
PB105297BL	1.329	0.0	0.005	
I6677-10	0.923	0.0	0.005	
I6677-10DUP	0.196	0.0	0.004	
I6677-12	-0.259	0.0	0.004	
I7022-02	-0.416	0.0	0.003	
I7022-04	-0.275	0.0	0.004	
I7022-07	-0.173	0.0	0.004	
I7037-02	-0.099	0.0	0.004	
CCV2	250.643	0.0	0.236	
CCB2	0.587	0.0	0.004	

N	14
Mean	42.930
SD	91.3172
CV%	212.71

AS.

Aquakem v. 7.2AQ1

Results from time period:

Fri Dec 22 14:06:58 2017

Fri Dec 22 14:12:47 2017

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Reactive CN	P	0.2818	µg/l	12/22/2017 11:08:34	
5.0PPBCN	A	Reactive CN	P	5.7278	µg/l	12/22/2017 11:08:35	
10PPBCN	A	Reactive CN	P	9.7081	µg/l	12/22/2017 11:08:36	
50PPBCN	A	Reactive CN	P	49.7801	µg/l	12/22/2017 11:08:37	
100PPBCN	A	Reactive CN	P	99.8581	µg/l	12/22/2017 11:08:38	
250PPBCN	A	Reactive CN	P	249.1903	µg/l	12/22/2017 11:08:39	
500PPBCN	A	Reactive CN	P	500.4538	µg/l	12/22/2017 11:08:40	
ICV1	S	Reactive CN	P	98.0473	µg/l	12/22/2017 14:06:59	
ICB1	S	Reactive CN	P	1.2511	µg/l	12/22/2017 14:07:00	
CCV1	S	Reactive CN	P	248.5288	µg/l	12/22/2017 14:07:02	
CCB1	S	Reactive CN	P	0.7436	µg/l	12/22/2017 14:07:05	
PB105297BL	S	Reactive CN	P	1.3288	µg/l	12/22/2017 14:07:06	
I6677-10	S	Reactive CN	P	0.9226	µg/l	12/22/2017 14:07:07	
I6677-10DUP	S	Reactive CN	P	0.1963	µg/l	12/22/2017 14:07:08	
I6677-12	S	Reactive CN	P	-0.2594	µg/l	12/22/2017 14:12:41	
I7022-02	S	Reactive CN	P	-0.4162	µg/l	12/22/2017 14:12:42	
I7022-04	S	Reactive CN	P	-0.275	µg/l	12/22/2017 14:12:43	
I7022-07	S	Reactive CN	P	-0.1726	µg/l	12/22/2017 14:12:44	
I7037-02	S	Reactive CN	P	-0.0986	µg/l	12/22/2017 14:12:45	
CCV2	S	Reactive CN	P	250.6429	µg/l	12/22/2017 14:12:46	
CCB2	S	Reactive CN	P	0.5871	µg/l	12/22/2017 14:12:47	

CHEMTECH

284 Sheffield Street,
Mountainside, NJ, 07092Reviewed by : AS

12/22/2017 11:19


Test Total CN

Accepted 12/22/2017 11:18

Factor 1079
Bias 0.004

Coeff. of det. 0.999992

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	0.2818	0.0000	
2	5.0PPBCN	0.009	5.7278	5.0000	
3	10PPBCN	0.013	9.7081	10.0000	
4	50PPBCN	0.050	49.7801	50.0000	
5	100PPBCN	0.096	99.8581	100.0000	
6	250PPBCN	0.235	249.1903	250.0000	
7	500PPBCN	0.468	500.4538	500.0000	